

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092223\
 Data File : BM041985.D
 Acq On : 22 Sep 2023 13:57
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV042

Quant Time: Sep 22 15:19:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 22 13:50:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.954	152	6983	0.400	ng/ul	0.00
4) Naphthalene-d8	10.768	136	16890	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.601	164	7372	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.354	188	15095	0.400	ng/ul	0.00
17) Chrysene-d12	21.524	240	7643	0.400	ng/ul	0.00
23) Perylene-d12	23.936	264	8020	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	3906	0.410	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.352	152	9005	0.426	ng/ul	0.00
18) Fluoranthene-d10	19.376	212	12485	0.397	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.410	88	3989	0.398	ng/ul#	91
5) Naphthalene	10.818	128	23833	0.421	ng/ul	100
7) 2-Methylnaphthalene	12.423	142	14209	0.421	ng/ul	100
8) 1-Methylnaphthalene	12.643	142	14356	0.419	ng/ul	99
10) Acenaphthylene	14.328	152	20460	0.414	ng/ul	100
11) Acenaphthene	14.661	153	15956	0.421	ng/ul	99
12) Fluorene	15.651	166	17856	0.419	ng/ul	100
14) Pentachlorophenol	16.983	266	3006	0.376	ng/ul	100
15) Phenanthrene	17.396	178	28399	0.406	ng/ul	100
16) Anthracene	17.489	178	24905	0.412	ng/ul	99
19) Fluoranthene	19.403	202	32300	0.374	ng/ul	99
20) Pyrene	19.771	202	33244	0.391	ng/ul	98
21) Benzo(a)anthracene	21.507	228	23927	0.413	ng/ul	100
22) Chrysene	21.562	228	24739	0.413	ng/ul	99
24) Benzo(b)fluoranthene	23.190	252	26069	0.422	ng/ul	95
25) Benzo(k)fluoranthene	23.240	252	24907	0.415	ng/ul	95
26) Benzo(a)pyrene	23.828	252	21702	0.428	ng/ul	93
27) Indeno(1,2,3-cd)pyrene	26.431	276	31312	0.449	ng/ul#	99
28) Dibenzo(a,h)anthracene	26.448	278	22637	0.450	ng/ul	97
29) Benzo(g,h,i)perylene	27.206	276	26205	0.447	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092223\
Data File : BM041985.D
Acq On : 22 Sep 2023 13:57
Operator : MA/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SICV042

Quant Time: Sep 22 15:19:37 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 22 13:50:58 2023
Response via : Initial Calibration

