


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092418\
Data File : BM016868.D
Acq On : 24 Sep 2018 09:57
Operator : SJ/JU
Sample : DFTPP08
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP08

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM091018MA.M
Title : SVOA CALIBRATION
Last Update : Thu Sep 27 02:42:56 2018



AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2470

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.8	87360	PASS
68	69	0.00	2	1.9	1780	PASS
69	198	0.00	100	34.6	95069	PASS
70	69	0.00	2	0.4	340	PASS
127	198	10	80	44.6	122749	PASS
197	198	0.00	2	0.6	1698	PASS
198	198	100	100	100.0	274986	PASS
199	198	5	9	6.5	17992	PASS
275	198	10	60	30.1	82680	PASS
365	198	1	100	4.0	10956	PASS
441	198	0.01	100	13.4	36853	PASS
442	198	50	100	90.7	249450	PASS
443	442	15	24	19.2	47933	PASS

Scan 2470 (17.515 min): BM016868.D
DFTPP08

m/z	Abundance
40.00	1015.0
44.00	14645.0
48.90	342.0
73.10	361.0
83.90	344.0
207.10	2067.0
208.00	349.0
209.10	387.0
281.10	608.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

m/z	Abundance
37.00	797.0
38.00	1041.0
39.00	9697.0
40.10	1341.0
41.10	453.0
44.00	14393.0
45.10	355.0
49.00	1239.0
50.10	29048.0
51.00	115912.0
52.00	6060.0
55.00	634.0
56.00	3678.0
57.00	8845.0
61.00	1695.0
62.00	1913.0
63.00	4947.0
64.00	825.0
65.00	2004.0
68.00	2564.0
69.00	126720.0
70.00	500.0
73.00	1441.0
74.00	11750.0
75.00	20272.0
76.10	6029.0
77.10	134080.0
78.10	9718.0
79.00	10382.0
80.00	7292.0
81.00	9547.0
82.00	2369.0
83.10	2434.0
83.90	583.0
85.00	1834.0
86.00	3138.0
87.00	1485.0
88.00	429.0
91.00	2807.0
92.00	2621.0
93.00	17160.0
94.00	1285.0
95.10	300.0
96.00	1090.0
97.20	476.0
98.00	13198.0
99.00	12172.0
99.90	1339.0
101.00	5924.0
103.00	2248.0
104.00	3698.0
105.00	3159.0
106.10	1262.0
107.00	43744.0
108.00	6539.0
109.10	1309.0
110.00	80040.0
111.00	13137.0
112.00	1688.0
113.00	665.0
116.00	2606.0
117.00	35752.0
118.00	2898.0
119.00	545.0
120.00	418.0
122.00	3027.0
123.00	4798.0
124.00	2431.0
125.00	1672.0
126.20	1087.0
127.00	151936.0
128.00	11731.0
129.00	59600.0
130.00	5030.0
131.10	1067.0
132.10	475.0
133.10	560.0
134.00	1812.0
135.00	4891.0
135.90	1658.0
137.10	3613.0
137.90	617.0
138.90	414.0
139.90	753.0
141.00	8610.0
142.00	2690.0
143.00	1412.0
144.10	343.0
145.00	303.0
146.00	1188.0
147.00	4259.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

148.00	10180.0
149.00	2393.0
150.00	553.0
151.00	1108.0
151.90	565.0
152.90	2184.0
154.00	2112.0
155.00	4064.0
156.00	6664.0
157.00	1467.0
158.00	1422.0
159.00	1117.0
160.00	2220.0
161.00	3718.0
162.00	1096.0
163.00	415.0
163.90	548.0
165.00	3476.0
165.90	2329.0
167.00	16225.0
168.00	14189.0
169.00	1752.0
170.00	680.0
170.90	780.0
171.90	1656.0
173.00	2308.0
174.00	3326.0
175.10	6237.0
176.00	2105.0
177.00	3307.0
177.90	993.0
179.00	11370.0
180.00	9995.0
181.00	4142.0
181.90	678.0
183.10	443.0
183.90	1033.0
185.00	5464.0
186.00	43144.0
187.00	12733.0
188.00	1656.0
189.00	2807.0
189.90	561.0
190.90	1394.0
192.00	3740.0
193.00	4751.0
193.90	1128.0
196.00	9916.0
197.10	1973.0
198.00	313344.0
198.90	20528.0
200.00	1716.0
201.40	1485.0
203.00	2184.0
204.00	11922.0
205.00	19480.0
206.00	85984.0
207.00	15206.0
208.00	3701.0
209.00	1295.0
210.20	1601.0
211.00	3171.0
212.00	733.0
215.00	1122.0
216.00	1870.0
217.00	22656.0
218.10	3317.0
219.00	354.0
221.00	13299.0
221.90	2726.0
223.00	5043.0
224.00	47320.0
225.00	12967.0
226.10	1733.0
227.00	20016.0
227.90	3415.0
229.00	4377.0
229.90	415.0
231.10	1707.0
232.00	326.0
233.00	477.0
234.00	1470.0
234.90	1327.0
236.00	768.0
237.00	2900.0
238.00	590.0
239.00	638.0
239.90	896.0
241.00	1172.0
242.00	2256.0
243.00	2380.0
244.00	35976.0
245.00	4982.0
246.00	6770.0
247.00	1371.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

248.00	305.0
249.00	1549.0
250.80	470.0
251.90	375.0
253.10	725.0
254.10	1139.0
255.00	195840.0
256.00	28432.0
257.00	2555.0
258.00	12132.0
258.90	1786.0
261.00	367.0
265.00	4360.0
266.00	812.0
270.90	665.0
271.90	503.0
273.00	5220.0
274.00	14924.0
275.00	85992.0
276.00	12150.0
277.00	8126.0
278.00	1863.0
279.10	320.0
281.10	576.0
281.90	471.0
283.00	1203.0
284.00	597.0
285.00	1335.0
285.90	385.0
288.80	402.0
290.90	318.0
293.00	1912.0
294.00	627.0
295.00	542.0
296.00	24384.0
297.00	3064.0
298.00	420.0
300.90	581.0
302.00	571.0
303.00	2728.0
304.10	466.0
309.00	352.0
310.00	304.0
314.00	902.0
315.00	2963.0
316.00	1907.0
321.00	676.0
322.00	313.0
323.10	7934.0
324.00	1735.0
326.90	1405.0
328.00	790.0
332.00	568.0
332.90	683.0
334.00	5445.0
335.10	1476.0
341.10	765.0
346.00	1627.0
352.10	1975.0
353.10	1430.0
354.10	2802.0
355.10	605.0
365.00	10600.0
365.90	1436.0
370.90	646.0
372.10	3691.0
373.00	1053.0
383.00	1107.0
390.10	588.0
402.00	1682.0
403.00	2430.0
404.10	753.0
421.10	1703.0
422.10	1548.0
423.10	11900.0
424.10	5543.0
425.00	886.0
441.10	30616.0
442.10	211264.0
443.10	41664.0
444.00	3577.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

m/z	Abundance
37.00	441.0
38.10	1352.0
39.10	7682.0
40.00	1074.0
44.00	12855.0
44.90	444.0
49.10	998.0
50.00	26072.0
51.10	100744.0
52.00	5028.0
55.00	558.0
56.00	3290.0
57.00	7336.0
61.00	1484.0
62.00	1503.0
63.00	4753.0
64.00	510.0
65.00	2060.0
66.00	325.0
68.10	1934.0
69.00	107504.0
70.00	522.0
73.00	1508.0
74.00	10861.0
75.00	17840.0
76.00	6274.0
77.10	122432.0
78.00	8650.0
79.00	9528.0
80.00	6985.0
81.00	8781.0
82.00	2153.0
83.00	1868.0
83.80	465.0
85.00	1531.0
86.00	2376.0
87.00	1252.0
87.90	569.0
91.00	2209.0
92.00	2456.0
93.00	15997.0
94.00	1219.0
96.00	729.0
98.00	11327.0
99.00	9940.0
100.00	936.0
101.00	5812.0
102.00	359.0
103.00	1893.0
104.00	3085.0
105.00	3139.0
106.10	1105.0
107.00	39488.0
108.00	6174.0
109.10	1229.0
110.00	73032.0
111.00	11277.0
111.90	1479.0
112.90	555.0
116.10	2625.0
117.00	34952.0
118.00	2177.0
119.10	412.0
119.90	378.0
122.00	2951.0
123.00	4752.0
124.00	2527.0
125.10	1868.0
126.20	1118.0
127.00	144640.0
128.00	10945.0
129.00	57192.0
130.00	4135.0
131.00	1028.0
132.00	407.0
132.90	358.0
134.00	1813.0
135.00	4833.0
136.00	1641.0
137.00	3235.0
137.90	645.0
138.90	356.0
139.90	708.0
141.00	7667.0
142.00	2376.0
143.00	1392.0
144.10	398.0
144.90	419.0
146.00	1591.0
147.00	4679.0
148.00	8339.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

149.00	2571.0
149.90	472.0
151.10	1060.0
151.70	534.0
153.00	2220.0
154.00	1736.0
155.00	4628.0
156.00	6795.0
157.00	1414.0
158.00	1741.0
158.90	900.0
160.00	2683.0
161.00	3599.0
162.00	890.0
162.90	440.0
163.90	504.0
164.90	2887.0
166.10	2390.0
167.00	16408.0
168.00	11341.0
169.00	1721.0
170.20	499.0
170.90	987.0
172.00	1373.0
172.90	1881.0
173.90	2589.0
175.00	6854.0
176.10	1738.0
177.00	3369.0
178.10	1010.0
179.00	10973.0
180.00	9720.0
181.00	3878.0
182.00	754.0
183.30	324.0
184.00	778.0
185.00	5780.0
186.00	45856.0
187.00	13037.0
187.90	1645.0
189.00	3040.0
190.00	433.0
191.00	1688.0
192.10	3651.0
193.00	4277.0
194.00	1027.0
195.10	407.0
196.00	10499.0
197.10	2141.0
198.00	336256.0
199.00	22408.0
200.00	1459.0
201.50	1739.0
203.00	2820.0
204.00	12339.0
205.00	20376.0
206.00	91960.0
207.00	14354.0
208.00	3188.0
209.00	1579.0
210.10	1668.0
211.00	3625.0
212.00	557.0
215.00	1058.0
216.00	2028.0
217.00	24600.0
218.00	2882.0
219.90	394.0
221.00	14781.0
221.90	3011.0
222.90	4794.0
224.00	51904.0
225.00	14373.0
226.10	1827.0
227.00	21128.0
228.00	3357.0
229.00	4285.0
229.90	683.0
231.00	2082.0
231.90	377.0
233.00	401.0
234.00	1531.0
235.10	1462.0
236.00	741.0
237.00	2801.0
237.90	476.0
239.00	812.0
239.90	848.0
240.90	1165.0
242.00	1946.0
243.00	2790.0
244.10	39560.0
245.00	5440.0
246.00	8090.0
247.00	2020.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

248.00	536.0
249.00	1514.0
250.00	364.0
251.00	528.0
251.90	433.0
252.90	922.0
254.10	1635.0
255.00	220800.0
256.00	31752.0
257.00	2806.0
258.00	12973.0
259.00	2164.0
260.00	371.0
265.00	4930.0
266.00	572.0
271.00	687.0
272.00	552.0
273.00	6354.0
274.00	17672.0
275.00	104064.0
276.00	13034.0
277.00	8290.0
278.00	1285.0
281.10	561.0
282.00	384.0
283.00	1103.0
284.00	543.0
285.00	1336.0
289.20	300.0
289.90	371.0
290.90	300.0
292.00	440.0
293.00	2041.0
294.00	668.0
295.00	560.0
296.00	28720.0
297.00	3867.0
298.00	322.0
300.90	552.0
301.90	388.0
303.00	3156.0
304.00	1296.0
305.00	330.0
308.00	434.0
310.00	443.0
313.00	408.0
314.10	1500.0
315.00	3938.0
316.00	2069.0
321.00	887.0
322.20	514.0
323.10	10332.0
324.00	1755.0
326.90	2210.0
327.90	998.0
332.00	751.0
333.00	941.0
334.00	6477.0
335.10	1422.0
341.10	1221.0
346.00	2036.0
347.00	510.0
352.00	3070.0
353.00	1933.0
354.10	3347.0
355.00	474.0
365.00	13414.0
366.00	1677.0
371.10	833.0
372.00	5570.0
373.10	1166.0
383.00	1749.0
384.00	317.0
390.10	559.0
391.10	394.0
400.80	443.0
402.00	2174.0
403.00	2983.0
404.00	1209.0
405.00	370.0
421.00	2080.0
422.00	2709.0
423.00	17456.0
424.10	5805.0
425.10	586.0
441.10	47096.0
442.10	317184.0
443.10	61480.0
444.00	5828.0
445.00	391.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

m/z	Abundance
37.00	337.0
38.00	517.0
39.00	3213.0
40.00	828.0
44.00	13840.0
44.90	322.0
48.90	454.0
50.00	12466.0
51.00	45424.0
52.00	2728.0
55.00	300.0
56.00	1663.0
57.00	3474.0
61.00	899.0
62.00	716.0
63.00	1767.0
63.90	303.0
65.10	1215.0
68.10	844.0
69.00	50984.0
73.00	1004.0
74.00	5767.0
75.00	9112.0
76.10	2491.0
77.10	57592.0
78.00	3899.0
79.00	4197.0
80.00	3182.0
80.90	4016.0
82.00	1197.0
82.90	1066.0
84.00	300.0
85.10	851.0
86.00	1511.0
86.90	578.0
91.00	995.0
92.00	1080.0
93.00	7376.0
93.90	576.0
96.10	554.0
98.00	4970.0
99.00	5443.0
100.00	330.0
101.00	2496.0
103.00	1132.0
103.90	1592.0
105.00	1464.0
105.90	408.0
106.20	429.0
107.00	18384.0
108.00	3075.0
110.00	35256.0
111.00	5336.0
112.00	728.0
116.00	1179.0
117.00	16504.0
118.00	1399.0
120.00	321.0
122.00	1253.0
123.00	2230.0
124.00	1087.0
125.00	1208.0
126.20	610.0
127.00	71672.0
128.00	5617.0
129.00	26528.0
130.00	2194.0
130.90	396.0
133.90	744.0
135.00	2144.0
136.00	1026.0
136.90	1549.0
138.00	317.0
140.10	334.0
141.00	3648.0
142.10	1378.0
142.90	959.0
145.90	656.0
146.90	1908.0
148.00	4442.0
148.90	1077.0
149.90	347.0
150.80	379.0
151.30	463.0
152.10	365.0
153.00	1189.0
154.00	1216.0
155.00	2351.0
156.10	3363.0
157.00	544.0
157.90	701.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

159.00	405.0
160.00	1197.0
160.90	1957.0
162.00	708.0
165.00	1435.0
166.00	1282.0
167.00	8359.0
168.00	6096.0
169.00	985.0
171.90	829.0
173.00	1172.0
174.00	1972.0
175.00	2843.0
175.90	941.0
177.00	1510.0
177.90	489.0
179.00	6533.0
180.00	5458.0
181.00	2063.0
182.00	309.0
184.00	403.0
185.00	2950.0
186.00	23296.0
187.00	7449.0
188.10	691.0
189.00	1757.0
189.90	312.0
191.00	1037.0
192.00	2092.0
193.00	2798.0
193.90	569.0
195.00	300.0
196.10	4757.0
196.80	1148.0
197.10	982.0
198.00	175360.0
199.00	11040.0
200.10	761.0
201.30	754.0
203.00	1459.0
204.00	6466.0
205.00	10445.0
206.00	47752.0
207.00	8304.0
207.90	2225.0
209.00	1019.0
210.10	741.0
211.00	1999.0
214.90	669.0
216.00	1193.0
217.00	13264.0
218.00	1634.0
221.00	9383.0
221.80	1729.0
223.00	3122.0
224.00	28144.0
225.00	7804.0
226.00	999.0
227.00	12232.0
227.90	1898.0
229.00	3011.0
230.00	590.0
231.10	1060.0
232.90	386.0
234.00	754.0
234.90	927.0
236.00	907.0
237.00	1514.0
239.00	574.0
240.00	526.0
241.00	671.0
242.00	1712.0
243.10	1749.0
244.00	23344.0
245.00	3057.0
246.00	4545.0
246.90	1032.0
249.00	947.0
253.10	602.0
254.10	844.0
255.00	119528.0
256.00	16568.0
257.00	1934.0
258.00	6781.0
258.90	1062.0
260.00	309.0
265.00	2272.0
265.90	483.0
270.90	358.0
272.00	362.0
273.00	4394.0
274.00	10013.0
275.00	57984.0
276.00	7597.0
277.00	5166.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

277.90	856.0
281.00	487.0
281.90	391.0
283.00	629.0
284.90	1044.0
292.00	670.0
292.90	1122.0
295.20	385.0
296.00	17032.0
297.00	2272.0
302.00	356.0
303.00	2045.0
304.00	561.0
309.00	353.0
314.00	766.0
315.00	1920.0
315.90	841.0
321.10	623.0
321.90	310.0
323.10	5800.0
324.00	1157.0
327.00	1333.0
328.00	744.0
331.90	441.0
332.90	534.0
334.00	3490.0
335.10	770.0
341.10	773.0
346.00	1272.0
346.90	314.0
352.00	1368.0
353.00	1076.0
354.00	2195.0
365.00	8856.0
365.90	1100.0
370.90	511.0
372.10	3105.0
372.90	742.0
383.00	1001.0
389.80	388.0
402.00	1527.0
403.10	2240.0
404.00	822.0
421.10	1710.0
422.00	1974.0
423.10	12549.0
424.10	3761.0
425.10	463.0
441.10	32848.0
442.10	219904.0
443.10	40656.0
444.10	3778.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08