


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092420\
 Data File : BM027510.D
 Acq On : 24 Sep 2020 09:04
 Operator : JU/CG
 Sample : DFTPP80
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP80

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM092420MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Sep 24 20:21:51 2020



AutoFind: Scans 2491, 2492, 2493; Background Corrected with Scan 2486

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.7	55211	PASS
68	69	0.00	2	1.9	1479	PASS
69	198	0.00	100	35.5	76211	PASS
70	69	0.00	2	0.6	439	PASS
127	198	10	80	37.4	80368	PASS
197	198	0.00	2	0.5	1104	PASS
198	198	100	100	100.0	214720	PASS
199	198	5	9	7.1	15328	PASS
275	198	10	60	29.8	63904	PASS
365	198	1	100	4.7	10153	PASS
441	198	0.01	100	10.9	23404	PASS
442	198	50	100	69.2	148517	PASS
443	442	15	24	19.6	29101	PASS

m/z	Abundance
40.00	774.0
44.00	449.0
207.00	632.0
281.00	212.0

Instrument :
BNA_M
ClientSampleId :
DFTPP80

m/z	Abundance
40.00	774.0
44.00	449.0
207.00	632.0
281.00	212.0

Instrument :
BNA_M
ClientSampleId :
DFTPP80

m/z	Abundance
37.10	248.0
38.00	721.0
39.10	3544.0
40.00	502.0
44.00	236.0
49.00	533.0
50.10	14313.0
51.10	49944.0
52.10	2195.0
53.00	166.0
55.00	364.0
56.00	1575.0
57.00	3452.0
58.00	178.0
61.00	820.0
62.00	977.0
63.10	2932.0
64.10	431.0
65.10	1349.0
68.00	1129.0
69.00	72584.0
69.90	301.0
73.10	905.0
74.00	9080.0
75.00	12330.0
76.10	4044.0
77.10	77264.0
78.00	5086.0
79.00	6354.0
80.00	5003.0
81.00	6596.0
82.00	1277.0
83.10	1201.0
85.00	997.0
86.00	2120.0
87.00	645.0
87.90	411.0
91.00	1074.0
92.00	1626.0
93.00	12655.0
94.00	780.0
95.00	182.0
96.10	542.0
96.90	163.0
98.00	10259.0
99.00	6866.0
100.10	672.0
101.00	3122.0
102.10	188.0
102.90	953.0
104.00	2328.0
105.00	2229.0
106.00	598.0
107.00	27904.0
108.00	3587.0
109.00	826.0
110.00	41808.0
111.00	6444.0
112.00	963.0
112.90	349.0
116.00	1865.0
117.00	33032.0
118.00	2212.0
118.90	233.0
120.00	292.0
122.10	1564.0
123.00	2720.0
124.00	1475.0
125.00	1200.0
126.00	335.0
127.00	79624.0
128.00	6646.0
129.00	35432.0
130.00	2980.0
131.00	697.0
132.00	219.0
134.00	1019.0
135.00	3081.0
136.00	1113.0
137.00	1606.0
137.90	429.0
138.80	186.0
139.10	178.0
140.00	411.0
141.00	5078.0
142.00	1779.0
143.10	1050.0
144.00	422.0
144.90	189.0
146.00	1158.0
147.10	3210.0
148.00	7820.0
148.90	1342.0
149.90	399.0
151.10	956.0
152.10	227.0
153.00	1790.0
154.00	1657.0

Instrument :
BNA_M
ClientSampleId :
DFTPP80

155.00	3584.0
156.10	4558.0
157.00	817.0
158.00	1074.0
158.90	916.0
160.00	1530.0
161.00	2672.0
161.90	849.0
163.00	391.0
163.90	368.0
165.00	2800.0
166.00	1985.0
167.00	13412.0
168.00	7404.0
169.00	1061.0
170.00	481.0
171.00	447.0
171.90	1089.0
173.00	1508.0
174.00	2911.0
175.00	4750.0
176.00	1169.0
177.00	2417.0
177.90	876.0
179.00	10767.0
180.00	7108.0
181.00	3356.0
182.10	440.0
182.70	160.0
184.00	1069.0
185.00	5380.0
186.00	35912.0
187.00	9655.0
188.00	1089.0
189.00	3272.0
189.90	452.0
191.00	1255.0
192.00	3377.0
193.00	4512.0
194.00	667.0
196.00	6396.0
197.00	1010.0
198.00	232512.0
198.90	17264.0
200.00	1568.0
200.90	441.0
201.40	868.0
203.00	2448.0
204.00	10550.0
205.00	18424.0
206.00	69728.0
207.00	10903.0
208.00	2996.0
208.90	1220.0
210.10	1455.0
211.00	3072.0
211.90	749.0
213.00	225.0
215.00	1210.0
216.00	1600.0
217.00	21680.0
218.00	2724.0
219.00	300.0
220.00	326.0
221.00	7900.0
222.00	2399.0
223.00	4416.0
224.00	43064.0
225.00	11403.0
226.00	1095.0
227.00	21952.0
228.00	3037.0
229.00	4305.0
229.90	337.0
231.00	1255.0
232.00	256.0
233.00	367.0
233.90	1570.0
235.00	1438.0
235.90	879.0
237.00	1692.0
237.90	251.0
239.00	1038.0
240.00	504.0
240.90	895.0
242.00	2437.0
243.00	2202.0
244.00	30760.0
245.10	4499.0
246.00	7950.0
247.00	1500.0
248.00	323.0
249.00	1245.0
250.00	245.0
251.00	250.0
251.90	269.0
252.90	596.0
254.00	956.0
255.00	165824.0
256.00	24968.0
257.00	1888.0
258.00	12020.0

Instrument :
BNA_M
ClientSampleId :
DFTPP80

259.00	1852.0
259.90	364.0
261.00	346.0
263.90	291.0
264.10	292.0
265.00	4617.0
266.00	815.0
269.90	291.0
271.10	461.0
272.00	461.0
273.00	5184.0
274.00	12170.0
275.00	70888.0
276.00	9300.0
277.00	6409.0
278.00	1200.0
278.90	360.0
281.00	358.0
281.90	344.0
283.00	605.0
284.00	749.0
285.00	1034.0
285.90	168.0
288.90	298.0
289.90	188.0
291.80	223.0
292.90	1453.0
294.00	479.0
295.10	428.0
296.00	23672.0
297.00	3395.0
298.00	208.0
301.00	255.0
301.90	336.0
303.00	2762.0
304.00	855.0
308.00	380.0
308.90	215.0
310.00	332.0
313.10	255.0
314.00	1037.0
315.00	2413.0
316.00	1425.0
316.90	315.0
321.10	657.0
322.10	298.0
323.10	5861.0
323.90	1181.0
327.00	1205.0
328.00	666.0
331.90	497.0
333.00	407.0
334.00	3785.0
335.00	1240.0
341.10	896.0
341.90	152.0
346.00	1382.0
346.90	263.0
351.00	156.0
352.00	2054.0
353.00	1345.0
354.10	1662.0
355.10	381.0
359.10	161.0
365.00	9852.0
366.00	1581.0
369.90	252.0
370.80	333.0
372.00	2850.0
373.20	935.0
383.00	675.0
384.00	184.0
390.00	404.0
391.00	204.0
401.00	328.0
402.10	1113.0
403.10	1647.0
403.90	443.0
421.00	1235.0
422.00	1216.0
423.10	8269.0
424.00	1637.0
441.10	22696.0
442.10	156480.0
443.10	32976.0
444.10	3052.0

Instrument :
BNA_M
ClientSampleId :
DFTPP80

m/z	Abundance
37.00	344.0
38.10	962.0
39.00	5582.0
40.00	883.0
43.90	426.0
49.00	529.0
50.10	20592.0
51.10	66992.0
52.00	3680.0
53.00	295.0
55.00	438.0
56.00	2157.0
57.00	4567.0
57.90	266.0
58.90	170.0
61.00	1170.0
62.10	1302.0
63.10	3556.0
64.00	586.0
65.00	1168.0
66.90	160.0
68.10	1621.0
69.00	92248.0
69.90	474.0
73.00	1124.0
74.00	12220.0
75.00	15356.0
76.10	6267.0
77.10	101312.0
78.10	6425.0
79.00	8831.0
80.00	6143.0
81.00	8429.0
82.00	1998.0
83.00	1282.0
85.00	1186.0
86.00	2806.0
87.00	1116.0
87.90	291.0
89.00	242.0
91.00	1773.0
91.90	1810.0
93.00	16584.0
94.00	1081.0
95.00	238.0
95.90	437.0
97.20	293.0
98.00	13755.0
99.00	8845.0
99.90	727.0
101.00	4696.0
102.10	218.0
103.00	1230.0
104.00	2730.0
105.00	2846.0
106.00	895.0
107.00	33744.0
108.00	4708.0
108.90	1193.0
110.00	49384.0
111.00	7665.0
112.00	1209.0
112.90	472.0
115.00	178.0
116.00	2451.0
117.00	47264.0
118.00	2976.0
119.00	170.0
120.00	354.0
121.00	174.0
122.00	2589.0
123.00	4413.0
124.00	1850.0
125.00	1437.0
126.10	384.0
127.00	94488.0
128.00	7470.0
129.00	44088.0
130.00	4061.0
130.90	705.0
132.00	425.0
132.90	167.0
134.00	1271.0
135.00	3815.0
136.00	1532.0
137.00	2091.0
137.90	354.0
138.80	170.0
139.90	553.0
141.00	7549.0
142.00	2356.0
142.90	1515.0

Instrument :
BNA_M
ClientSampleId :
DFTPP80

144.10	369.0
144.90	236.0
146.00	1448.0
147.00	4810.0
148.00	10706.0
149.00	2105.0
149.90	378.0
151.10	722.0
151.90	315.0
153.00	2371.0
154.00	1745.0
155.00	3965.0
156.10	5895.0
157.00	962.0
158.00	1124.0
159.00	1177.0
160.00	2639.0
161.00	3640.0
162.00	1119.0
162.80	172.0
163.90	323.0
165.00	2980.0
166.00	1891.0
167.00	17848.0
168.00	9255.0
169.00	1355.0
169.90	435.0
171.00	723.0
172.00	1252.0
173.00	2002.0
174.00	3809.0
175.00	6171.0
176.00	1346.0
177.00	2945.0
178.00	1167.0
179.00	12874.0
180.00	8480.0
181.00	3887.0
182.10	822.0
183.00	180.0
184.00	949.0
185.00	5865.0
186.00	41104.0
187.00	12041.0
188.00	1203.0
189.00	3294.0
190.00	651.0
191.00	1456.0
192.00	4619.0
193.00	4727.0
194.00	1059.0
195.10	369.0
196.00	6881.0
196.90	1465.0
197.90	252160.0
199.00	17712.0
200.00	1648.0
201.40	878.0
203.00	2821.0
204.00	13096.0
205.00	21224.0
206.00	78416.0
207.00	10021.0
208.00	3754.0
208.90	1048.0
209.90	1452.0
211.00	3751.0
211.80	553.0
213.00	291.0
214.90	988.0
216.00	2244.0
217.00	26656.0
218.00	3000.0
218.90	408.0
219.90	353.0
221.00	7794.0
222.00	2811.0
223.00	4786.0
224.00	51144.0
225.00	12353.0
226.00	1067.0
227.00	25184.0
228.00	3083.0
229.00	4981.0
229.90	607.0
231.00	1795.0
232.00	267.0
233.10	421.0
234.00	1813.0
235.00	1937.0
236.00	1167.0
237.00	1596.0
238.00	345.0
239.00	907.0
239.90	1003.0
241.00	1468.0

Instrument :
BNA_M
ClientSampleId :
DFTPP80

242.00	2589.0
243.00	2223.0
244.00	34720.0
245.10	4641.0
246.00	8506.0
247.00	1635.0
247.90	350.0
249.00	1215.0
249.90	239.0
251.00	529.0
252.00	431.0
253.00	858.0
254.00	1065.0
255.00	187008.0
256.00	26064.0
257.00	2061.0
258.00	14088.0
258.90	2028.0
260.00	314.0
261.00	275.0
263.80	273.0
265.00	5964.0
266.00	779.0
270.10	162.0
271.00	497.0
272.00	638.0
273.00	5285.0
274.00	14041.0
275.00	74840.0
276.00	10029.0
277.00	7805.0
278.00	1238.0
278.90	255.0
281.00	286.0
282.00	229.0
283.00	756.0
284.00	765.0
285.00	1379.0
285.90	383.0
288.90	313.0
289.90	489.0
291.00	301.0
291.90	279.0
293.00	1625.0
293.90	429.0
294.90	341.0
296.00	28824.0
297.00	4859.0
298.00	186.0
301.00	215.0
302.00	496.0
303.00	2934.0
304.00	567.0
307.90	303.0
309.90	400.0
313.00	204.0
314.00	1277.0
315.00	2934.0
316.00	1453.0
317.00	261.0
321.00	759.0
322.10	355.0
323.00	6825.0
324.00	1231.0
326.00	162.0
327.00	1829.0
328.00	698.0
332.00	843.0
332.90	544.0
334.00	4546.0
335.00	1275.0
341.00	960.0
342.00	231.0
346.00	1621.0
347.00	335.0
352.00	2483.0
352.90	1517.0
354.10	1904.0
355.00	406.0
364.10	207.0
365.00	12388.0
366.00	1530.0
370.00	224.0
371.00	435.0
372.10	3286.0
373.00	641.0
383.00	920.0
383.90	325.0
389.90	323.0
391.00	346.0
392.20	185.0
402.00	1121.0
403.00	1683.0
404.00	792.0
404.80	163.0
421.00	1251.0

Instrument :
BNA_M
ClientSampleId :
DFTPP80

422.00	1021.0
423.10	8701.0
424.00	1667.0
441.10	25328.0
442.10	162688.0
443.10	33440.0
444.10	3276.0
445.00	217.0

Instrument :
BNA_M
ClientSampleId :
DFTPP80

m/z	Abundance
37.00	307.0
38.00	601.0
39.10	3790.0
40.00	801.0
43.90	547.0
49.00	540.0
50.00	15227.0
51.10	52680.0
52.10	2933.0
54.90	321.0
56.00	1530.0
57.00	3828.0
57.90	221.0
61.00	910.0
62.10	904.0
63.00	2775.0
64.10	373.0
65.00	1176.0
68.00	1446.0
69.00	72488.0
70.00	502.0
73.00	904.0
74.00	9271.0
75.00	12793.0
76.10	4584.0
77.10	80976.0
78.10	5334.0
79.00	6714.0
80.00	5196.0
81.00	6848.0
82.00	1430.0
82.90	1175.0
85.00	1325.0
86.00	2506.0
87.00	1035.0
88.00	345.0
89.00	245.0
91.00	1681.0
92.00	1874.0
93.00	14046.0
94.00	1187.0
96.00	351.0
98.00	10655.0
99.00	7908.0
99.90	430.0
101.00	3601.0
102.00	283.0
102.90	1104.0
104.00	2667.0
105.00	2169.0
106.00	776.0
107.00	29256.0
108.00	4302.0
109.00	929.0
110.00	41416.0
111.00	7386.0
112.00	1167.0
112.90	210.0
115.10	161.0
116.00	1945.0
117.00	40208.0
118.00	2417.0
119.00	253.0
120.10	307.0
121.00	270.0
121.90	2024.0
123.00	2976.0
124.00	1398.0
125.00	1107.0
126.00	294.0
127.00	83504.0
128.00	6528.0
129.00	39312.0
130.00	2954.0
131.00	607.0
131.90	272.0
133.90	1185.0
135.00	3011.0
136.00	1689.0
137.00	1451.0
138.10	476.0
138.80	218.0
140.00	604.0
141.00	5461.0
142.00	1863.0
143.00	1238.0
144.00	342.0
145.10	304.0
146.10	1142.0
147.00	3149.0
148.00	9616.0
149.00	1767.0

Instrument :
BNA_M
ClientSampleId :
DFTPP80

150.00	414.0
151.20	630.0
152.00	295.0
153.00	1712.0
153.90	1715.0
155.00	3445.0
156.00	4852.0
157.00	935.0
158.00	847.0
158.90	944.0
160.10	2267.0
161.00	2915.0
162.00	707.0
163.00	259.0
163.80	223.0
165.00	2977.0
166.10	2140.0
167.00	16696.0
168.00	8367.0
169.00	1183.0
170.00	573.0
171.00	800.0
171.90	1220.0
173.00	1924.0
174.00	3059.0
175.00	5177.0
176.00	1234.0
177.00	2589.0
178.00	816.0
179.00	11894.0
180.00	6810.0
181.00	3499.0
182.00	687.0
182.80	251.0
184.00	885.0
185.00	4762.0
186.00	38720.0
187.00	11304.0
188.00	926.0
189.00	2955.0
190.00	403.0
190.90	1543.0
192.00	3549.0
193.00	4487.0
194.00	986.0
194.90	400.0
195.90	5760.0
196.90	1066.0
197.90	237888.0
198.90	16278.0
200.00	1643.0
201.50	760.0
202.10	321.0
203.00	2745.0
204.00	13115.0
205.00	18816.0
206.00	77080.0
207.00	10491.0
208.00	3374.0
209.00	1059.0
210.00	1519.0
211.00	2886.0
211.90	397.0
213.00	384.0
214.90	1174.0
216.00	2116.0
217.00	25736.0
218.00	3378.0
219.00	238.0
220.00	529.0
221.00	7831.0
222.00	2886.0
223.00	5465.0
224.00	48160.0
225.00	12560.0
226.00	865.0
227.00	24592.0
228.00	3248.0
229.00	4391.0
230.10	372.0
231.00	1429.0
232.00	449.0
233.10	416.0
234.00	1337.0
235.00	1664.0
235.90	1050.0
237.00	1943.0
237.90	385.0
239.00	778.0
239.90	865.0
240.90	1349.0
242.00	2594.0
243.00	2540.0
244.00	34784.0
245.00	4866.0
246.00	8650.0

Instrument :
BNA_M
ClientSampleId :
DFTPP80

247.00	1759.0
247.90	359.0
249.00	1244.0
250.00	201.0
251.00	250.0
251.90	394.0
252.90	918.0
254.00	1047.0
255.00	182080.0
256.00	26504.0
257.00	1953.0
258.00	13943.0
259.00	2156.0
260.00	306.0
260.90	354.0
264.00	333.0
265.00	6241.0
265.90	769.0
266.90	194.0
270.90	529.0
272.00	627.0
273.00	5842.0
274.00	14521.0
275.00	74528.0
276.00	10460.0
277.00	8089.0
278.00	1065.0
278.90	260.0
281.00	349.0
282.00	275.0
283.00	1054.0
284.00	557.0
285.10	1124.0
286.10	394.0
288.90	428.0
289.80	241.0
292.00	551.0
293.00	2110.0
294.00	392.0
295.00	679.0
296.00	29776.0
297.00	3675.0
298.00	180.0
300.90	247.0
302.00	575.0
303.10	2676.0
304.00	713.0
307.90	527.0
309.00	187.0
309.90	383.0
313.00	178.0
314.10	1198.0
315.00	3527.0
316.10	1653.0
317.10	199.0
321.10	834.0
322.00	503.0
323.10	6488.0
324.00	1215.0
325.90	195.0
327.00	1634.0
327.90	564.0
329.00	223.0
332.00	620.0
332.90	751.0
334.00	5491.0
335.00	1154.0
341.10	851.0
342.00	321.0
346.00	1864.0
346.80	312.0
352.10	2395.0
353.00	1660.0
354.10	1918.0
355.00	501.0
365.00	12929.0
366.10	1693.0
367.00	156.0
371.00	659.0
372.00	3977.0
373.00	893.0
374.10	154.0
383.00	1082.0
384.00	226.0
390.00	556.0
390.90	333.0
392.10	232.0
401.00	176.0
402.00	1795.0
403.00	2170.0
404.00	1027.0
421.00	1654.0
422.00	1487.0
423.00	10621.0
424.10	2115.0
425.00	211.0

Instrument :
BNA_M
ClientSampleId :
DFTPP80

441.10	32184.0
442.10	201664.0
443.10	37808.0
444.10	3238.0
445.00	257.0

Instrument :
BNA_M
ClientSampleId :
DFTPP80