

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092523\
 Data File : BM042030.D
 Acq On : 25 Sep 2023 13:09
 Operator : MA/JU
 Sample : 04429-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Sep 25 14:14:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 22 13:50:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.950	152	11074	0.400	ng/ul	# 0.00
4) Naphthalene-d8	10.773	136	19981	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.597	164	11226	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.350	188	16122	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.521	240	7370	0.400	ng/ul	# 0.00
23) Perylene-d12	23.930	264	8115	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.368	96	39327	2.606	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.352	152	8546	0.342	ng/ul	0.00
18) Fluoranthene-d10	19.371	212	11625	0.383	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.406	88	15984	1.007	ng/ul#	1
5) Naphthalene	10.823	128	278509	4.161	ng/ul	99
7) 2-Methylnaphthalene	12.429	142	141554	3.548	ng/ul	98
8) 1-Methylnaphthalene	12.643	142	90551	2.235	ng/ul	82
10) Acenaphthylene	14.347	152	35630	0.474	ng/ul#	87
11) Acenaphthene	14.652	153	8019	0.139	ng/ul#	59
12) Fluorene	15.647	166	12508	0.193	ng/ul#	88
14) Pentachlorophenol	16.987	266	279029	32.722	ng/ul	100
15) Phenanthrene	17.392	178	4623	0.062	ng/ul#	1
16) Anthracene	17.485	178	2115	0.033	ng/ul#	1
26) Benzo(a)pyrene	23.775	252	4449	0.087	ng/ul#	13

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092523\
Data File : BM042030.D
Acq On : 25 Sep 2023 13:09
Operator : MA/JU
Sample : 04429-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Sep 25 14:14:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 22 13:50:58 2023
Response via : Initial Calibration

