

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092523\
 Data File : BM042027.D
 Acq On : 25 Sep 2023 11:20
 Operator : MA/JU
 Sample : 04386-15
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AJ2

Quant Time: Sep 25 12:29:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 22 13:50:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.950	152	5275	0.400	ng/ul	0.00
4) Naphthalene-d8	10.763	136	13413	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.597	164	6259	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.346	188	12390	0.400	ng/ul	0.00
17) Chrysene-d12	21.521	240	4218	0.400	ng/ul	0.00
23) Perylene-d12	23.927	264	3722	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.368	96	28814	4.009	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.346	152	5397	0.321	ng/ul	0.00
18) Fluoranthene-d10	19.371	212	8633	0.497	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.406	88	1234	0.163	ng/ul#	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092523\
Data File : BM042027.D
Acq On : 25 Sep 2023 11:20
Operator : MA/JU
Sample : 04386-15
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AJ2

Quant Time: Sep 25 12:29:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 22 13:50:58 2023
Response via : Initial Calibration

