

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092523\
 Data File : BM042054.D
 Acq On : 26 Sep 2023 04:48
 Operator : MA/JU
 Sample : PB155715BL
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK715

Quant Time: Sep 26 05:20:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 22 13:50:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	6126	0.400	ng/ul	0.00
4) Naphthalene-d8	10.757	136	15346	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.592	164	7171	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.345	188	15324	0.400	ng/ul	0.00
17) Chrysene-d12	21.518	240	6611	0.400	ng/ul	0.00
23) Perylene-d12	23.921	264	6394	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.368	96	54442	6.522	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.341	152	7051	0.367	ng/ul	-0.01
18) Fluoranthene-d10	19.366	212	10301	0.378	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092523\
Data File : BM042054.D
Acq On : 26 Sep 2023 04:48
Operator : MA/JU
Sample : PB155715BL
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK715

Quant Time: Sep 26 05:20:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 22 13:50:58 2023
Response via : Initial Calibration

