

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092523\
 Data File : BM042027.D
 Acq On : 25 Sep 2023 11:20
 Operator : MA/JU
 Sample : 04386-15
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AJ2

Quant Time: Sep 25 12:29:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 22 13:50:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.950	152	5275	0.400	ng/ul	0.00
4) Naphthalene-d8	10.763	136	13413	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.597	164	6259	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.346	188	12390	0.400	ng/ul	0.00
17) Chrysene-d12	21.521	240	4218	0.400	ng/ul	0.00
23) Perylene-d12	23.927	264	3722	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.368	96	28814	4.009	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.346	152	5397	0.321	ng/ul	0.00
18) Fluoranthene-d10	19.371	212	8633	0.497	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.406	88	1234	0.163	ng/ul#	77


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092523\
Data File : BM042027.D
Acq On : 25 Sep 2023 11:20
Operator : MA/JU
Sample : 04386-15
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AJ2


Quant Time: Sep 25 12:29:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 22 13:50:58 2023
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.163 ng/ul
 RT: 3.406 min Scan# 51
 Delta R.T. -0.004 min
 Lab File: BM042027.D
 Acq: 25 Sep 2023 11:20

Instrument :
 BNA_M
 ClientSampleId :
 C0AJ2



Tgt Ion: 88 Resp: 1234

Ion	Ratio	Lower	Upper
88	100		
43	82.1	42.9	64.3#
58	61.8	56.4	84.6

