

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092624\
 Data File : BM047660.D
 Acq On : 27 Sep 2024 01:14
 Operator : RC/JU
 Sample : P4187-12
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EY078

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 09/27/2024
 Supervised By :mohammad ahmed 09/28/2024

Quant Time: Sep 27 02:05:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 25 16:27:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.826	152	4886	0.400	ng/ul	0.00
4) Naphthalene-d8	10.618	136	15266	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.456	164	9889	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.192	188	17443m	0.400	ng/ul	0.00
17) Chrysene-d12	21.356	240	13273	0.400	ng/ul	0.00
23) Perylene-d12	23.636	264	12134m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.269	96	30877	4.919	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.207	152	7888	0.387	ng/ul	0.00
18) Fluoranthene-d10	19.211	212	17697	0.435	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.307	88	837	0.111	ng/ul#	36

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092624\
Data File : BM047660.D
Acq On : 27 Sep 2024 01:14
Operator : RC/JU
Sample : P4187-12
Misc :
ALS Vial : 23 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
EY078

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 09/27/2024
Supervised By :mohammad ahmed 09/28/2024


Quant Time: Sep 27 02:05:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 25 16:27:08 2024
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.111 ng/ul
 RT: 3.307 min Scan# 71
 Delta R.T. 0.000 min
 Lab File: BM047660.D
 Acq: 27 Sep 2024 01:14

Instrument :
 BNA_M
 ClientSampleId :
 EY078



Tgt Ion: 88	Resp: 83
Ion Ratio	Lower Upper
88 100	
43 111.5	30.9 46.3
58 37.0	42.5 63.7

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 09/27/2024
 Supervised By :mohammad ahmed 09/28/2024

