

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092721\
 Data File : BM032222.D
 Acq On : 28 Sep 2021 01:34
 Operator : CG/JU
 Sample : M3919-09
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 GB6Y8

Quant Time: Sep 28 02:31:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 27 13:37:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.661	152	7650	0.400	ng/ul	0.00
4) Naphthalene-d8	10.451	136	23378	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.307	164	11770	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.032	188	21443	0.400	ng/ul	0.00
17) Chrysene-d12	21.190	240	16358	0.400	ng/ul	0.00
23) Perylene-d12	23.353	264	16290	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.986	96	44273	5.383	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.043	152	10683	0.343	ng/ul	0.00
18) Fluoranthene-d10	19.049	212	19878	0.447	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.028	88	278	0.033	ng/ul#	42

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092721\
Data File : BM032222.D
Acq On : 28 Sep 2021 01:34
Operator : CG/JU
Sample : M3919-09
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
GB6Y8

Quant Time: Sep 28 02:31:10 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 27 13:37:22 2021
Response via : Initial Calibration

