

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092721\
 Data File : BM032255.D
 Acq On : 29 Sep 2021 00:41
 Operator : CG/JU
 Sample : M3739-16
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0B80

Quant Time: Sep 29 03:25:48 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 27 13:37:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.657	152	7073	0.400	ng/ul	0.00
4) Naphthalene-d8	10.447	136	21749	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.300	164	10325	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.026	188	18078	0.400	ng/ul	0.00
17) Chrysene-d12	21.188	240	13513	0.400	ng/ul	0.00
23) Perylene-d12	23.349	264	14146	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.986	96	33898	4.458	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.039	152	9089	0.314	ng/ul	0.00
18) Fluoranthene-d10	19.046	212	17240	0.470	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092721\
Data File : BM032255.D
Acq On : 29 Sep 2021 00:41
Operator : CG/JU
Sample : M3739-16
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0B80

Quant Time: Sep 29 03:25:48 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 27 13:37:22 2021
Response via : Initial Calibration

