

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092721\
 Data File : BM032267.D
 Acq On : 29 Sep 2021 10:13
 Operator : CG/JU
 Sample : PB139007BL
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK007

Quant Time: Sep 29 10:58:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 27 13:37:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.653	152	5768	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.442	136	16963	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.300	164	7059	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.026	188	12788	0.400	ng/ul	0.00
17) Chrysene-d12	21.188	240	9469	0.400	ng/ul	0.00
23) Perylene-d12	23.351	264	8226	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.986	96	4419	0.713	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.035	152	7721	0.342	ng/ul	-0.01
18) Fluoranthene-d10	19.046	212	10347	0.402	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092721\
Data File : BM032267.D
Acq On : 29 Sep 2021 10:13
Operator : CG/JU
Sample : PB139007BL
Misc :
ALS Vial : 72 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK007

Quant Time: Sep 29 10:58:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 27 13:37:22 2021
Response via : Initial Calibration

