

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092721\
 Data File : BM032270.D
 Acq On : 29 Sep 2021 12:22
 Operator : CG/JU
 Sample : M3739-19
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AG1

Quant Time: Sep 29 12:52:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 27 13:37:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.653	152	6687	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.447	136	20305	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.300	164	9426	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.026	188	16608	0.400	ng/ul	0.00
17) Chrysene-d12	21.188	240	12971	0.400	ng/ul	0.00
23) Perylene-d12	23.351	264	13445	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.986	96	14326	1.993	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.035	152	8548	0.316	ng/ul	-0.01
18) Fluoranthene-d10	19.046	212	15908	0.452	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092721\
Data File : BM032270.D
Acq On : 29 Sep 2021 12:22
Operator : CG/JU
Sample : M3739-19
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AG1

Quant Time: Sep 29 12:52:39 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 27 13:37:22 2021
Response via : Initial Calibration

