

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092721\
 Data File : BM032206.D
 Acq On : 27 Sep 2021 15:45
 Operator : CG/JU
 Sample : M3919-10
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 GB6Y9

Quant Time: Sep 27 16:17:55 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 27 13:37:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.660	152	7148	0.400	ng/ul	0.00
4) Naphthalene-d8	10.452	136	23620	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.308	164	13310	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.033	188	25385	0.400	ng/ul	0.00
17) Chrysene-d12	21.193	240	18209	0.400	ng/ul	0.00
23) Perylene-d12	23.361	264	16044	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.986	96	42043	5.471	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.044	152	10946	0.348	ng/ul	0.00
18) Fluoranthene-d10	19.053	212	21089	0.426	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.024	88	27408	3.467	ng/ul	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092721\
Data File : BM032206.D
Acq On : 27 Sep 2021 15:45
Operator : CG/JU
Sample : M3919-10
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
GB6Y9


Quant Time: Sep 27 16:17:55 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 27 13:37:22 2021
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 3.467 ng/ul
 RT: 3.024 min Scan# 55
 Delta R.T. -0.003 min
 Lab File: BM032206.D
 Acq: 27 Sep 2021 15:45

Instrument :
 BNA_M
 ClientSampleId :
 GB6Y9



Tgt Ion:	88	Resp:	27408
Ion Ratio	Lower	Upper	
88	100		
43	27.4	26.1	39.1
58	71.5	50.2	75.2

