

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092723\
 Data File : BM042092.D
 Acq On : 27 Sep 2023 20:12
 Operator : MA/JU
 Sample : 04450-20
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BBJ44

Quant Time: Sep 28 01:03:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 22 13:50:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.941	152	3665	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.757	136	9805	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.587	164	4834	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.341	188	10724	0.400	ng/ul	-0.01
17) Chrysene-d12	21.515	240	5681	0.400	ng/ul	0.00
23) Perylene-d12	23.921	264	6373	0.400	ng/ul	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	26271	5.260	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.341	152	5164	0.421	ng/ul	-0.01
18) Fluoranthene-d10	19.361	212	10273	0.439	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.402	88	1019	0.194	ng/ul#	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092723\
Data File : BM042092.D
Acq On : 27 Sep 2023 20:12
Operator : MA/JU
Sample : 04450-20
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BBJ44

Quant Time: Sep 28 01:03:53 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 22 13:50:58 2023
Response via : Initial Calibration

