

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092724\
 Data File : BM047682.D
 Acq On : 27 Sep 2024 20:14
 Operator : RC/JU
 Sample : SST04039
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST040023

Quant Time: Sep 28 02:06:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 28 01:59:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	150938	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	627707	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	374365	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	718161	20.000	ng/u1	0.00
79) Chrysene-d12	21.409	240	695790	20.000	ng/u1	0.00
88) Perylene-d12	24.415	264	727373	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.257	96	71164	19.014	ng/uL	0.00
4) Pyridine-d5	3.669	84	555306	46.442	ng/u1	0.00
7) Phenol-d5	6.981	99	565228	39.950	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.145	67	404119	45.887	ng/u1	0.00
11) 2-Chlorophenol-d4	7.351	132	436600	40.518	ng/u1	0.00
15) 4-Methylphenol-d8	8.522	113	430646	38.724	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	212507	43.838	ng/u1	0.00
24) 2-Nitrophenol-d4	9.692	143	212016	46.296	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	406094	43.611	ng/u1	0.00
31) 4-Chloroaniline-d4	10.739	131	594807	39.030	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	1109880	40.335	ng/u1	0.00
49) Acenaphthylene-d8	14.139	160	1356427	40.852	ng/u1	0.00
54) 4-Nitrophenol-d4	14.639	143	223234	40.946	ng/u1	0.00
60) Fluorene-d10	15.439	176	973606	40.946	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.551	200	173402	45.673	ng/u1	0.00
73) Anthracene-d10	17.286	188	1472081	42.567	ng/u1	0.00
81) Pyrene-d10	19.568	212	1642866	40.513	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.215	264	1594689	41.257	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	78414	19.413	ng/uL	99
5) Pyridine	3.693	79	582010	47.143	ng/u1	99
6) Benzaldehyde	6.951	77	350935	46.862	ng/u1	98
8) Phenol	7.004	94	610144	41.891	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.240	93	484007	42.023	ng/u1	99
12) 2-Chlorophenol	7.381	128	456358	40.891	ng/u1	98
13) 2-Methylphenol	8.257	108	424267	38.697	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.334	45	922934	54.679	ng/u1	100
16) Acetophenone	8.634	105	739995	41.349	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.628	70	418171	44.028	ng/u1	98
18) 4-Methylphenol	8.587	108	481046	40.185	ng/u1	100
19) Hexachloroethane	8.898	117	200394	41.790	ng/u1	98
22) Nitrobenzene	9.010	77	590475	48.746	ng/u1	99
23) Isophorone	9.545	82	1075136	44.507	ng/u1	99
25) 2-Nitrophenol	9.722	139	236656	44.698	ng/u1	99
26) 2,4-Dimethylphenol	9.786	107	484708	42.561	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.022	93	632778	43.289	ng/u1	97
29) 2,4-Dichlorophenol	10.257	162	414402	43.895	ng/u1	99
30) Naphthalene	10.663	128	1448212	41.876	ng/u1	100
32) 4-Chloroaniline	10.763	127	587625	39.193	ng/u1	98
33) Hexachlorobutadiene	10.957	225	260049	46.183	ng/u1	98
34) Caprolactam	11.539	113	70189	19.303	ng/u1	95
35) 4-Chloro-3-methylphenol	11.892	107	441529	41.533	ng/u1	99
36) 2-Methylnaphthalene	12.275	142	951827	40.040	ng/u1	97


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092724\
 Data File : BM047682.D
 Acq On : 27 Sep 2024 20:14
 Operator : RC/JU
 Sample : SSTD04039
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD040023

Quant Time: Sep 28 02:06:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 28 01:59:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.492	142	965538	40.200	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.639	216	487241	43.840	ng/ul	99
40) Hexachlorocyclopentadiene	12.627	237	294066	38.486	ng/ul	100
41) 2,4,6-Trichlorophenol	12.880	196	300235	44.025	ng/ul	99
42) 2,4,5-Trichlorophenol	12.951	196	325507	43.915	ng/ul	95
43) 1,1'-Biphenyl	13.280	154	1207063	40.905	ng/ul	99
44) 2-Chloronaphthalene	13.322	162	931850	41.178	ng/ul	99
45) 2-Nitroaniline	13.521	65	329249	50.011	ng/ul	95
47) Dimethylphthalate	13.904	163	1117923	39.998	ng/ul	99
48) 2,6-Dinitrotoluene	14.021	165	215208	42.196	ng/ul	93
50) Acenaphthylene	14.169	152	1516851	42.261	ng/ul	99
51) 3-Nitroaniline	14.345	138	244721	41.098	ng/ul	99
52) Acenaphthene	14.516	153	1042156	40.870	ng/ul	98
53) 2,4-Dinitrophenol	14.551	184	118613	42.265	ng/ul	89
55) 4-Nitrophenol	14.651	109	188802	41.134	ng/ul	96
56) Dibenzofuran	14.845	168	1406522	42.438	ng/ul	99
57) 2,4-Dinitrotoluene	14.804	165	322303	43.883	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.074	232	263450	46.189	ng/ul	99
59) Diethylphthalate	15.268	149	1133657	38.337	ng/ul	99
61) Fluorene	15.498	166	1195001	41.314	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.492	204	566793	42.316	ng/ul	98
63) 4-Nitroaniline	15.504	138	256816	43.585	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.568	198	193696	45.136	ng/ul#	95
67) N-Nitrosodiphenylamine	15.698	169	935831	41.918	ng/ul	100
68) 4-Bromophenyl-phenylether	16.380	248	305349	41.096	ng/ul	94
69) Hexachlorobenzene	16.498	284	352621	41.535	ng/ul	98
70) Atrazine	16.651	200	309027	39.403	ng/ul	99
71) Pentachlorophenol	16.833	266	220261	41.226	ng/ul	98
72) Phenanthrene	17.227	178	1720372	41.991	ng/ul	99
74) Anthracene	17.321	178	1798216	43.347	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.251	216	472777	45.270	ng/ul	99
76) Pentachlorobenzene	14.768	250	466183	45.197	ng/ul	99
77) Carbazole	17.586	167	1616044	42.607	ng/ul	99
78) Di-n-butylphthalate	18.156	149	1887651	38.602	ng/ul	99
80) Fluoranthene	19.239	202	1913703	40.040	ng/ul	99
82) Pyrene	19.598	202	2149704	42.199	ng/ul	99
83) Butylbenzylphthalate	20.497	149	804698	34.603	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.315	252	614920	37.380	ng/ul	99
85) Benzo(a)anthracene	21.392	228	1995318	40.145	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.321	149	1255219	35.595	ng/ul	100
87) Chrysene	21.456	228	1930177	42.093	ng/ul	99
89) Di-n-octyl phthalate	22.456	149	1964442	33.807	ng/ul	100
90) Benzo(b)fluoranthene	23.468	252	1889279	40.463	ng/ul	100
91) Benzo(k)fluoranthene	23.533	252	1960759	42.756	ng/ul	99
93) Benzo(a)pyrene	24.280	252	1823260	42.506	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.815	276	2287552	43.486	ng/ul	99
95) Dibenzo(a,h)anthracene	27.874	278	1878078	43.848	ng/ul	99
96) Benzo(g,h,i)perylene	28.879	276	1784631	44.412	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 19.413 ng/uL
RT: 3.293 min Scan# 51
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion: 88 Resp: 78414
Ion Ratio Lower Upper
88 100
43 37.1 29.1 43.7
58 83.6 67.7 101.5



#5
Pyridine
Concen: 47.143 ng/uL
RT: 3.693 min Scan# 120
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion: 79 Resp: 582010
Ion Ratio Lower Upper
79 100
52 85.7 69.4 104.2
51 41.6 32.6 49.0





#6

Benzaldehyde

Concen: 46.862 ng/ul

RT: 6.951 min Scan# 61

Delta R.T. -0.000 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Instrument :

BNA_M

ClientSampleId :

SSTD040023





Tgt Ion: 77 Resp: 350935

Ion Ratio Lower Upper

77 100

105 87.5 72.2 108.2

106 86.6 70.7 106.1



#8

Phenol

Concen: 41.891 ng/ul

RT: 7.004 min Scan# 683

Delta R.T. -0.000 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14




Tgt Ion: 94 Resp: 610144


Ion Ratio Lower Upper

94 100

65 31.5 26.3 39.5

66 39.0 30.5 45.7









#10
Bis(2-Chloroethyl)ether
Concen: 42.023 ng/ul
RT: 7.240 min Scan# 723
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion: 93 Resp: 484007
Ion Ratio Lower Upper
93 100
63 87.2 69.4 104.2
95 32.4 25.0 37.6



#12
2-Chlorophenol
Concen: 40.891 ng/ul
RT: 7.381 min Scan# 747
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:128 Resp: 456358
Ion Ratio Lower Upper
128 100
64 55.0 42.7 64.1
130 33.5 26.0 39.0









#13
2-Methylphenol
Concen: 38.697 ng/ul
RT: 8.257 min Scan# 896
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:108 Resp: 424267
Ion Ratio Lower Upper
108 100
107 86.9 69.3 103.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 54.679 ng/ul
RT: 8.334 min Scan# 909
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion: 45 Resp: 922934
Ion Ratio Lower Upper
45 100
77 11.1 9.0 13.6
79 8.2 6.5 9.7





#16

Acetophenone

Concen: 41.349 ng/ul

RT: 8.634 min Scan# 90

Delta R.T. -0.000 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Instrument :

BNA_M

ClientSampleId :

SSTD040023





Tgt Ion:105 Resp: 739995

Ion Ratio Lower Upper

105 100

77 85.9 66.6 100.0

51 34.8 27.0 40.4



#17

N-Nitroso-di-n-propylamine

Concen: 44.028 ng/ul

RT: 8.628 min Scan# 959

Delta R.T. 0.006 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Tgt Ion: 70 Resp: 418171




Ion Ratio Lower Upper


70 100

42 71.4 55.3 82.9

101 8.0 6.8 10.2

130 17.3 13.7 20.5









#18
4-Methylphenol
Concen: 40.185 ng/ul
RT: 8.587 min Scan# 91
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:108 Resp: 481046
Ion Ratio Lower Upper
108 100
107 110.2 88.4 132.6



#19
Hexachloroethane
Concen: 41.790 ng/ul
RT: 8.898 min Scan# 1005
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:117 Resp: 200394
Ion Ratio Lower Upper
117 100
201 91.5 71.0 106.6
199 56.1 44.2 66.4









#22
Nitrobenzene
Concen: 48.746 ng/ul
RT: 9.010 min Scan# 1024
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion: 77 Resp: 590475
Ion Ratio Lower Upper
77 100
123 39.7 31.7 47.5
65 13.3 11.1 16.7



#23
Isophorone
Concen: 44.507 ng/ul
RT: 9.545 min Scan# 1115
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion: 82 Resp: 1075136
Ion Ratio Lower Upper
82 100
95 6.5 4.9 7.3
138 15.5 12.0 18.0









#25
2-Nitrophenol
Concen: 44.698 ng/ul
RT: 9.722 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:139 Resp: 236656
Ion Ratio Lower Upper
139 100
65 61.0 49.6 74.4
109 29.8 23.4 35.2



#26
2,4-Dimethylphenol
Concen: 42.561 ng/ul
RT: 9.786 min Scan# 1156
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:107 Resp: 484708
Ion Ratio Lower Upper
107 100
121 47.3 37.8 56.8
122 82.8 65.4 98.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 43.289 ng/ul
 RT: 10.022 min Scan# 1196
 Delta R.T. -0.000 min
 Lab File: BM047682.D
 Acq: 27 Sep 2024 20:14

Instrument :
 BNA_M
 ClientSampleId :
 SSTD040023

Tgt Ion: 93 Resp: 632778




Ion	Ratio	Lower	Upper
93	100		
95	31.8	27.1	40.7
123	10.9	8.2	12.4




#29
 2,4-Dichlorophenol
 Concen: 43.895 ng/ul
 RT: 10.257 min Scan# 1236
 Delta R.T. -0.000 min
 Lab File: BM047682.D
 Acq: 27 Sep 2024 20:14

Tgt Ion:162 Resp: 414402

Ion	Ratio	Lower	Upper
162	100		
164	64.8	52.2	78.2
98	39.4	30.6	46.0









#30
Naphthalene
Concen: 41.876 ng/ul
RT: 10.663 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:128 Resp: 1448212
Ion Ratio Lower Upper
128 100
129 10.5 8.6 12.8
127 12.9 10.4 15.6



#32
4-Chloroaniline
Concen: 39.193 ng/ul
RT: 10.763 min Scan# 1322
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:127 Resp: 587625
Ion Ratio Lower Upper
127 100
129 31.4 25.9 38.9





#33

Hexachlorobutadiene

Concen: 46.183 ng/ul

RT: 10.957 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Instrument :

BNA_M

ClientSampleId :

SSTD040023





Tgt Ion:225 Resp: 260049

Ion Ratio Lower Upper

225 100

223 63.5 50.3 75.5

227 63.6 49.1 73.7



#34

Caprolactam

Concen: 19.303 ng/ul

RT: 11.539 min Scan# 1454

Delta R.T. 0.006 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14




Tgt Ion:113 Resp: 70189


Ion Ratio Lower Upper

113 100

55 218.2 167.9 251.9

56 156.9 131.3 196.9









#35
4-Chloro-3-methylphenol
Concen: 41.533 ng/ul
RT: 11.892 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:107 Resp: 441529
Ion Ratio Lower Upper
107 100
144 25.0 18.7 28.1
142 78.0 62.7 94.1



#36
2-Methylnaphthalene
Concen: 40.040 ng/ul
RT: 12.275 min Scan# 1579
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:142 Resp: 951827
Ion Ratio Lower Upper
142 100
141 87.8 72.5 108.7









#37
1-Methylnaphthalene
Concen: 40.200 ng/ul
RT: 12.492 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:142 Resp: 965538
Ion Ratio Lower Upper
142 100
141 91.6 73.4 110.2
116 3.9 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 43.840 ng/ul
RT: 12.639 min Scan# 1641
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:216 Resp: 487241
Ion Ratio Lower Upper
216 100
214 76.5 61.0 91.4
179 20.4 17.7 26.5
108 20.8 17.0 25.4









#40
Hexachlorocyclopentadiene
Concen: 38.486 ng/ul
RT: 12.627 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:237 Resp: 294066
Ion Ratio Lower Upper
237 100
235 62.4 49.8 74.6
272 11.9 9.7 14.5



#41
2,4,6-Trichlorophenol
Concen: 44.025 ng/ul
RT: 12.880 min Scan# 1682
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:196 Resp: 300235
Ion Ratio Lower Upper
196 100
198 97.5 77.1 115.7
200 30.5 24.5 36.7





#42

2,4,5-Trichlorophenol

Concen: 43.915 ng/ul

RT: 12.951 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM047682.D


Acq: 27 Sep 2024 20:14

Instrument :

BNA_M

ClientSampleId :

SSTD040023






Tgt Ion:196 Resp: 325507

Ion Ratio Lower Upper

196 100

198 93.5 79.5 119.3

200 30.8 25.8 38.8



#43

1,1'-Biphenyl

Concen: 40.905 ng/ul

RT: 13.280 min Scan# 1750

Delta R.T. -0.000 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14




Tgt Ion:154 Resp: 1207063


Ion Ratio Lower Upper

154 100

153 39.4 31.1 46.7

76 15.1 12.1 18.1









#44
2-Chloronaphthalene
Concen: 41.178 ng/ul
RT: 13.322 min Scan# 1757
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:162 Resp: 931850
Ion Ratio Lower Upper
162 100
164 33.2 25.3 37.9
127 41.0 32.9 49.3



#45
2-Nitroaniline
Concen: 50.011 ng/ul
RT: 13.521 min Scan# 1791
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion: 65 Resp: 329249
Ion Ratio Lower Upper
65 100
92 61.9 49.1 73.7
138 92.2 67.6 101.4









#47
Dimethylphthalate
Concen: 39.998 ng/ul
RT: 13.904 min Scan# 1856
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:163 Resp: 1117923
Ion Ratio Lower Upper
163 100
194 3.8 3.0 4.6
164 10.3 7.8 11.6



#48
2,6-Dinitrotoluene
Concen: 42.196 ng/ul
RT: 14.021 min Scan# 1876
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:165 Resp: 215208
Ion Ratio Lower Upper
165 100
89 66.5 58.5 87.7
121 23.7 20.2 30.2









#50
Acenaphthylene
Concen: 42.261 ng/ul
RT: 14.169 min Scan# 1901
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:152 Resp: 1516851
Ion Ratio Lower Upper
152 100
151 19.8 16.2 24.2
153 12.9 10.7 16.1



#51
3-Nitroaniline
Concen: 41.098 ng/ul
RT: 14.345 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:138 Resp: 244721
Ion Ratio Lower Upper
138 100
108 10.3 9.0 13.6
92 126.3 101.5 152.3









#52
Acenaphthene
Concen: 40.870 ng/ul
RT: 14.516 min Scan# 1959
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:153 Resp: 1042156
Ion Ratio Lower Upper
153 100
152 49.5 39.7 59.5
154 88.5 72.6 108.8



#53
2,4-Dinitrophenol
Concen: 42.265 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:184 Resp: 118613
Ion Ratio Lower Upper
184 100
63 81.5 77.5 116.3
154 66.3 55.8 83.6









#55
4-Nitrophenol
Concen: 41.134 ng/ul
RT: 14.651 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:109 Resp: 188802
Ion Ratio Lower Upper
109 100
139 124.7 105.7 158.5
65 151.2 123.1 184.7



#56
Dibenzofuran
Concen: 42.438 ng/ul
RT: 14.845 min Scan# 2016
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:168 Resp: 1406522
Ion Ratio Lower Upper
168 100
139 39.4 31.0 46.6









#57
2,4-Dinitrotoluene
Concen: 43.883 ng/ul
RT: 14.804 min Scan# 2009
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:165 Resp: 322303
Ion Ratio Lower Upper
165 100
63 56.2 45.8 68.6
182 2.2 1.7 2.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 46.189 ng/ul
RT: 15.074 min Scan# 2055
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:232 Resp: 263450
Ion Ratio Lower Upper
232 100
131 51.4 40.2 60.4
130 2.8 2.6 4.0
166 31.6 25.3 37.9





#59

Diethylphthalate

Concen: 38.337 ng/ul

RT: 15.268 min Scan# 2088

Delta R.T. -0.000 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Instrument :

BNA_M

ClientSampleId :

SSTD040023





Tgt Ion:149 Resp: 1133657

Ion Ratio Lower Upper

149 100

177 21.1 16.9 25.3

150 12.6 9.2 13.8



#61

Fluorene

Concen: 41.314 ng/ul

RT: 15.498 min Scan# 2127

Delta R.T. 0.006 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14




Tgt Ion:166 Resp: 1195001


Ion Ratio Lower Upper

166 100

165 97.6 77.8 116.6

167 13.6 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 42.316 ng/ul
RT: 15.492 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:204 Resp: 566793
Ion Ratio Lower Upper
204 100
206 32.8 27.0 40.4
141 60.2 49.2 73.8



#63
4-Nitroaniline
Concen: 43.585 ng/ul
RT: 15.504 min Scan# 2128
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:138 Resp: 256816
Ion Ratio Lower Upper
138 100
92 54.8 44.9 67.3
108 79.8 62.4 93.6









#66
4,6-Dinitro-2-methylphenol
Concen: 45.136 ng/ul
RT: 15.568 min Scan# 2139
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:198 Resp: 193696
Ion Ratio Lower Upper
198 100
182 4.4 5.1 7.7#
77 29.8 26.2 39.2



#67
N-Nitrosodiphenylamine
Concen: 41.918 ng/ul
RT: 15.698 min Scan# 2161
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:169 Resp: 935831
Ion Ratio Lower Upper
169 100
168 67.4 53.5 80.3
167 37.4 30.0 45.0









#68
4-Bromophenyl-phenylether
Concen: 41.096 ng/ul
RT: 16.380 min Scan# 2111
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:248 Resp: 305349
Ion Ratio Lower Upper
248 100
250 100.6 76.5 114.7
141 82.4 60.8 91.2



#69
Hexachlorobenzene
Concen: 41.535 ng/ul
RT: 16.498 min Scan# 2297
Delta R.T. -0.000 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:284 Resp: 352621
Ion Ratio Lower Upper
284 100
142 40.8 32.4 48.6
249 30.8 26.3 39.5





#70

Atrazine

Concen: 39.403 ng/ul

RT: 16.651 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Instrument :

BNA_M

ClientSampleId :

SSTD040023





Tgt Ion:200 Resp: 309027

Ion Ratio Lower Upper

200 100

173 26.0 20.2 30.2

215 48.9 39.4 59.2



#71

Pentachlorophenol

Concen: 41.226 ng/ul

RT: 16.833 min Scan# 2354

Delta R.T. -0.006 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14




Tgt Ion:266 Resp: 220261


Ion Ratio Lower Upper

266 100

264 60.7 50.2 75.2

268 64.3 50.9 76.3





#72

Phenanthrene

Concen: 41.991 ng/ul

RT: 17.227 min Scan# 2421

Delta R.T. -0.000 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Instrument :

BNA_M

ClientSampleId :

SSTD040023





Tgt Ion:178 Resp: 1720372

Ion Ratio Lower Upper

178 100

179 15.5 12.0 18.0

176 19.5 15.4 23.0



#74

Anthracene

Concen: 43.347 ng/ul

RT: 17.321 min Scan# 2437

Delta R.T. 0.006 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14




Tgt Ion:178 Resp: 1798216


Ion Ratio Lower Upper

178 100

179 15.2 12.4 18.6

176 18.8 14.7 22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 45.270 ng/uL

RT: 13.251 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Instrument :

BNA_M

ClientSampleId :

SSTD040023





Tgt Ion:216 Resp: 472777

Ion Ratio Lower Upper

216 100

214 78.1 63.0 94.4

218 48.2 37.6 56.4



#76

Pentachlorobenzene

Concen: 45.197 ng/uL

RT: 14.768 min Scan# 2003

Delta R.T. -0.000 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14




Tgt Ion:250 Resp: 466183


Ion Ratio Lower Upper

250 100

248 61.8 50.2 75.2

252 62.6 49.4 74.0









#77
Carbazole
Concen: 42.607 ng/ul
RT: 17.586 min Scan# 2481
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:167 Resp: 1616044
Ion Ratio Lower Upper
167 100
166 21.2 17.4 26.0
139 13.0 10.9 16.3



#78
Di-n-butylphthalate
Concen: 38.602 ng/ul
RT: 18.156 min Scan# 2579
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:149 Resp: 1887651
Ion Ratio Lower Upper
149 100
150 9.0 7.4 11.2
104 5.5 4.6 7.0





#80

Fluoranthene

Concen: 40.040 ng/ul

RT: 19.239 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Instrument :

BNA_M

ClientSampleId :

SSTD040023





Tgt Ion:202 Resp: 1913703

Ion Ratio Lower Upper

202 100

101 11.0 9.0 13.4

100 8.7 6.8 10.2



#82

Pyrene

Concen: 42.199 ng/ul

RT: 19.598 min Scan# 2824

Delta R.T. -0.000 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14




Tgt Ion:202 Resp: 2149704


Ion Ratio Lower Upper

202 100

101 13.6 10.9 16.3

100 11.4 9.4 14.2









#83
 Butylbenzylphthalate
 Concen: 34.603 ng/ul
 RT: 20.497 min Scan# 2976
 Delta R.T. 0.006 min
 Lab File: BM047682.D
 Acq: 27 Sep 2024 20:14




Instrument :
 BNA_M
 ClientSampleId :
 SSTD040023


Tgt Ion:149 Resp: 804698
 Ion Ratio Lower Upper
 149 100
 91 77.5 66.9 100.3
 206 19.2 14.4 21.6



#84
 3,3'-Dichlorobenzidine
 Concen: 37.380 ng/ul
 RT: 21.315 min Scan# 3116
 Delta R.T. -0.000 min
 Lab File: BM047682.D
 Acq: 27 Sep 2024 20:14

Tgt Ion:252 Resp: 614920
 Ion Ratio Lower Upper
 252 100
 254 65.4 51.5 77.3
 126 11.7 9.4 14.0





#85

Benzo(a)anthracene

Concen: 40.145 ng/ul

RT: 21.392 min Scan# 3129

Delta R.T. -0.000 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Instrument :

BNA_M

ClientSampleId :

SSTD040023





Tgt Ion:228 Resp: 1995318

Ion Ratio Lower Upper

228 100

229 19.7 15.6 23.4

226 27.6 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 35.595 ng/ul

RT: 21.321 min Scan# 3117

Delta R.T. -0.000 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14




Tgt Ion:149 Resp: 1255219


Ion Ratio Lower Upper

149 100

167 26.9 21.3 31.9

279 3.5 2.7 4.1





#87

Chrysene

Concen: 42.093 ng/ul

RT: 21.456 min Scan# 3139

Delta R.T. 0.006 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Instrument :

BNA_M

ClientSampleId :

SSTD040023





Tgt Ion:228 Resp: 1930177

Ion Ratio Lower Upper

228 100

226 30.3 23.6 35.4

229 19.4 15.3 22.9



#89

Di-n-octyl phthalate

Concen: 33.807 ng/ul




RT: 22.456 min Scan# 3310


Delta R.T. -0.000 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Tgt Ion:149 Resp: 1964442









#90
Benzo(b)fluoranthene
Concen: 40.463 ng/ul
RT: 23.468 min Scan# 3481
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14




Instrument :
BNA_M
ClientSampleId :
SSTD040023


Tgt Ion:252 Resp: 1889279
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.2
125 10.7 9.0 13.4



#91
Benzo(k)fluoranthene
Concen: 42.756 ng/ul
RT: 23.533 min Scan# 3493
Delta R.T. 0.006 min
Lab File: BM047682.D
Acq: 27 Sep 2024 20:14

Tgt Ion:252 Resp: 1960759
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 10.5 8.9 13.3





#93

Benzo(a)pyrene

Concen: 42.506 ng/ul

RT: 24.280 min Scan# 3

Delta R.T. 0.006 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Instrument :

BNA_M

ClientSampleId :

SSTD040023





Tgt Ion:252 Resp: 1823260

Ion Ratio Lower Upper

252 100

253 22.1 17.2 25.8

125 11.9 10.1 15.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 43.486 ng/ul

RT: 27.815 min Scan# 4221

Delta R.T. 0.006 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14




Tgt Ion:276 Resp: 2287552


Ion Ratio Lower Upper

276 100

138 21.4 16.9 25.3

277 24.2 18.9 28.3





#95

Dibenzo(a,h)anthracene

Concen: 43.848 ng/ul

RT: 27.874 min Scan# 4402

Delta R.T. 0.012 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Instrument :

BNA_M

ClientSampleId :

SSTD040023





Tgt Ion:278 Resp: 1878078

Ion Ratio Lower Upper

278 100

139 16.9 13.4 20.2

279 24.0 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 44.412 ng/ul

RT: 28.879 min Scan# 4402

Delta R.T. 0.018 min

Lab File: BM047682.D

Acq: 27 Sep 2024 20:14

Tgt Ion:276 Resp: 1784631

Ion Ratio Lower Upper

276 100

138 23.2 18.7 28.1

277 22.4 18.3 27.5

