

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092724\
 Data File : BM047685.D
 Acq On : 27 Sep 2024 22:13
 Operator : RC/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV026

Quant Time: Sep 28 02:34:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 28 02:30:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	191736	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	783036	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	470574	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	934534	20.000	ng/u1	0.00
79) Chrysene-d12	21.409	240	872408	20.000	ng/u1	0.00
88) Perylene-d12	24.415	264	955923	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.257	96	43551	7.340	ng/uL	0.00
4) Pyridine-d5	3.675	84	337656	19.437	ng/u1	0.00
7) Phenol-d5	6.981	99	344410	18.996	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.145	67	261359	20.315	ng/u1	0.00
11) 2-Chlorophenol-d4	7.345	132	281546	21.344	ng/u1	0.00
15) 4-Methylphenol-d8	8.522	113	267765	19.864	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	132820	21.304	ng/u1	0.00
24) 2-Nitrophenol-d4	9.692	143	133113	22.280	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	262078	21.959	ng/u1	0.00
31) 4-Chloroaniline-d4	10.739	131	388310	21.045	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	727101	21.312	ng/u1	0.00
49) Acenaphthylene-d8	14.139	160	871269	21.539	ng/u1	0.00
54) 4-Nitrophenol-d4	14.633	143	141277	20.798	ng/u1	0.00
60) Fluorene-d10	15.439	176	617666	20.901	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.551	200	104759	19.156	ng/u1	0.00
73) Anthracene-d10	17.280	188	941330	20.485	ng/u1	0.00
81) Pyrene-d10	19.568	212	1061540	21.439	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.209	264	1061690	21.406	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	46750	7.214	ng/uL	99
5) Pyridine	3.693	79	311155	17.013	ng/u1	98
6) Benzaldehyde	6.951	77	177525	18.005	ng/u1	99
8) Phenol	7.004	94	361925	18.761	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.240	93	295270	19.209	ng/u1	95
12) 2-Chlorophenol	7.381	128	287002	20.537	ng/u1	98
13) 2-Methylphenol	8.257	108	261101	19.410	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.339	45	522795	18.066	ng/u1	99
16) Acetophenone	8.634	105	454058	19.686	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.622	70	245377	20.085	ng/u1	98
18) 4-Methylphenol	8.581	108	291469	19.281	ng/u1	99
19) Hexachloroethane	8.898	117	124489	19.819	ng/u1	90
22) Nitrobenzene	9.010	77	344406	19.446	ng/u1	97
23) Isophorone	9.539	82	633416	19.993	ng/u1	96
25) 2-Nitrophenol	9.722	139	149044	21.806	ng/u1	98
26) 2,4-Dimethylphenol	9.781	107	264294	18.046	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.022	93	384619	19.849	ng/u1	98
29) 2,4-Dichlorophenol	10.257	162	261886	21.250	ng/u1	99
30) Naphthalene	10.663	128	907338	20.145	ng/u1	99
32) 4-Chloroaniline	10.763	127	377269	20.765	ng/u1	99
33) Hexachlorobutadiene	10.951	225	169234	20.783	ng/u1	98
34) Caprolactam	11.533	113	80485	20.420	ng/u1	86
35) 4-Chloro-3-methylphenol	11.886	107	269803	21.149	ng/u1	96
36) 2-Methylnaphthalene	12.269	142	614647	21.477	ng/u1	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092724\
 Data File : BM047685.D
 Acq On : 27 Sep 2024 22:13
 Operator : RC/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV026

Quant Time: Sep 28 02:34:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 28 02:30:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.492	142	572242	19.654	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.639	216	311924	20.363	ng/ul	98
40) Hexachlorocyclopentadiene	12.627	237	176969	19.101	ng/ul	98
41) 2,4,6-Trichlorophenol	12.880	196	194174	21.908	ng/ul	97
42) 2,4,5-Trichlorophenol	12.945	196	208009	21.462	ng/ul	95
43) 1,1'-Biphenyl	13.280	154	780535	20.717	ng/ul	98
44) 2-Chloronaphthalene	13.322	162	591210	20.289	ng/ul	98
45) 2-Nitroaniline	13.521	65	187224	20.240	ng/ul	90
47) Dimethylphthalate	13.904	163	715169	20.699	ng/ul	100
48) 2,6-Dinitrotoluene	14.016	165	145206	23.930	ng/ul	91
50) Acenaphthylene	14.169	152	953782	20.908	ng/ul	99
51) 3-Nitroaniline	14.345	138	166449	22.664	ng/ul	95
52) Acenaphthene	14.516	153	646417	20.045	ng/ul	97
53) 2,4-Dinitrophenol	14.551	184	69266	18.473	ng/ul#	83
55) 4-Nitrophenol	14.645	109	108269	18.984	ng/ul	98
56) Dibenzofuran	14.845	168	893598	20.586	ng/ul	98
57) 2,4-Dinitrotoluene	14.804	165	200184	21.965	ng/ul	90
58) 2,3,4,6-Tetrachlorophenol	15.068	232	179362	23.477	ng/ul#	98
59) Diethylphthalate	15.268	149	708142	20.484	ng/ul	98
61) Fluorene	15.492	166	718988	19.853	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.486	204	347635	20.315	ng/ul	99
63) 4-Nitroaniline	15.504	138	172898	24.241	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.563	198	118461	19.549	ng/ul	94
67) N-Nitrosodiphenylamine	15.698	169	595471	20.448	ng/ul	100
68) 4-Bromophenyl-phenylether	16.380	248	201680	20.936	ng/ul	98
69) Hexachlorobenzene	16.498	284	237444	21.181	ng/ul	95
70) Atrazine	16.645	200	200005	20.632	ng/ul	96
71) Pentachlorophenol	16.833	266	146891	20.607	ng/ul	98
72) Phenanthrene	17.227	178	1077575	19.662	ng/ul	99
74) Anthracene	17.315	178	1106739	19.800	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.245	216	287586	18.586	ng/ul	99
76) Pentachlorobenzene	14.768	250	293607	19.440	ng/ul	96
77) Carbazole	17.580	167	1010305	19.941	ng/ul	100
78) Di-n-butylphthalate	18.151	149	1149067	20.371	ng/ul	99
80) Fluoranthene	19.233	202	1215911	20.851	ng/ul	100
82) Pyrene	19.598	202	1327204	20.449	ng/ul	98
83) Butylbenzylphthalate	20.492	149	482020	21.007	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.315	252	418019	22.676	ng/ul	99
85) Benzo(a)anthracene	21.392	228	1253746	20.525	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.321	149	718034	20.843	ng/ul	98
87) Chrysene	21.450	228	1210353	20.325	ng/ul	98
89) Di-n-octyl phthalate	22.456	149	1144884	18.568	ng/ul	100
90) Benzo(b)fluoranthene	23.462	252	1205917	20.229	ng/ul	99
91) Benzo(k)fluoranthene	23.527	252	1248609	20.484	ng/ul	99
93) Benzo(a)pyrene	24.274	252	1181199	20.863	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	27.803	276	1475890	20.342	ng/ul	99
95) Dibenzo(a,h)anthracene	27.868	278	1220860	20.532	ng/ul	99
96) Benzo(g,h,i)perylene	28.862	276	1132381	19.920	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :

BNA M

ClientSampleId :

SICV026


Quant Time: Sep 28 02:34:09 2024


Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M

Quant Title : SVOA CALIBRATION

QLast Update : Sat Sep 28 02:30:36 2024

Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.214 ng/uL
RT: 3.293 min Scan# 51
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion: 88 Resp: 46750
Ion Ratio Lower Upper
88 100
43 37.1 29.1 43.7
58 83.3 67.7 101.5



#5
Pyridine
Concen: 17.013 ng/uL
RT: 3.693 min Scan# 120
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion: 79 Resp: 311155
Ion Ratio Lower Upper
79 100
52 84.8 69.4 104.2
51 40.4 32.6 49.0









#6
Benzaldehyde
Concen: 18.005 ng/ul
RT: 6.951 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion: 77 Resp: 177525
Ion Ratio Lower Upper
77 100
105 91.1 72.2 108.2
106 88.1 70.7 106.1



#8
Phenol
Concen: 18.761 ng/ul
RT: 7.004 min Scan# 683
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion: 94 Resp: 361925
Ion Ratio Lower Upper
94 100
65 30.2 26.3 39.5
66 38.0 30.5 45.7









#10
Bis(2-Chloroethyl)ether
Concen: 19.209 ng/ul
RT: 7.240 min Scan# 723
Delta R.T. 0.006 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion: 93 Resp: 295270
Ion Ratio Lower Upper
93 100
63 80.9 69.4 104.2
95 32.2 25.0 37.6



#12
2-Chlorophenol
Concen: 20.537 ng/ul
RT: 7.381 min Scan# 747
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:128 Resp: 287002
Ion Ratio Lower Upper
128 100
64 51.2 42.7 64.1
130 32.5 26.0 39.0









#13
2-Methylphenol
Concen: 19.410 ng/ul
RT: 8.257 min Scan# 896
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:108 Resp: 261101
Ion Ratio Lower Upper
108 100
107 90.3 69.3 103.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.066 ng/ul
RT: 8.339 min Scan# 910
Delta R.T. 0.006 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion: 45 Resp: 522795
Ion Ratio Lower Upper
45 100
77 11.9 9.0 13.6
79 8.2 6.5 9.7









#16
Acetophenone
Concen: 19.686 ng/ul
RT: 8.634 min Scan# 960
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:105 Resp: 454058
Ion Ratio Lower Upper
105 100
77 82.2 66.6 100.0
51 32.3 27.0 40.4



#17
N-Nitroso-di-n-propylamine
Concen: 20.085 ng/ul
RT: 8.622 min Scan# 958
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion: 70 Resp: 245377
Ion Ratio Lower Upper
70 100
42 67.9 55.3 82.9
101 8.5 6.8 10.2
130 18.5 13.7 20.5









#18
4-Methylphenol
Concen: 19.281 ng/ul
RT: 8.581 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:108 Resp: 291469
Ion Ratio Lower Upper
108 100
107 109.2 88.4 132.6



#19
Hexachloroethane
Concen: 19.819 ng/ul
RT: 8.898 min Scan# 1005
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:117 Resp: 124489
Ion Ratio Lower Upper
117 100
201 98.6 71.0 106.6
199 61.3 44.2 66.4









#22
Nitrobenzene
Concen: 19.446 ng/ul
RT: 9.010 min Scan# 1102
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion: 77 Resp: 344406
Ion Ratio Lower Upper
77 100
123 42.1 31.7 47.5
65 13.4 11.1 16.7



#23
Isophorone
Concen: 19.993 ng/ul
RT: 9.539 min Scan# 1114
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion: 82 Resp: 633416
Ion Ratio Lower Upper
82 100
95 6.6 4.9 7.3
138 17.1 12.0 18.0









#25
2-Nitrophenol
Concen: 21.806 ng/ul
RT: 9.722 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:139 Resp: 149044
Ion Ratio Lower Upper
139 100
65 59.5 49.6 74.4
109 29.5 23.4 35.2



#26
2,4-Dimethylphenol
Concen: 18.046 ng/ul
RT: 9.781 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:107 Resp: 264294
Ion Ratio Lower Upper
107 100
121 49.4 37.8 56.8
122 84.5 65.4 98.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.849 ng/ul
 RT: 10.022 min Scan# 1196
 Delta R.T. -0.000 min
 Lab File: BM047685.D
 Acq: 27 Sep 2024 22:13




Instrument :
 BNA_M
 ClientSampleId :
 SICV026


Tgt Ion: 93 Resp: 384619
 Ion Ratio Lower Upper
 93 100
 95 32.7 27.1 40.7
 123 10.9 8.2 12.4



#29
 2,4-Dichlorophenol
 Concen: 21.250 ng/ul
 RT: 10.257 min Scan# 1236
 Delta R.T. -0.000 min
 Lab File: BM047685.D
 Acq: 27 Sep 2024 22:13

Tgt Ion:162 Resp: 261886
 Ion Ratio Lower Upper
 162 100
 164 65.4 52.2 78.2
 98 37.5 30.6 46.0









#30
Naphthalene
Concen: 20.145 ng/ul
RT: 10.663 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:128 Resp: 907338
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 12.8 10.4 15.6



#32
4-Chloroaniline
Concen: 20.765 ng/ul
RT: 10.763 min Scan# 1322
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:127 Resp: 377269
Ion Ratio Lower Upper
127 100
129 32.0 25.9 38.9









#33
Hexachlorobutadiene
Concen: 20.783 ng/ul
RT: 10.951 min Scan# 1354
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:225 Resp: 169234
Ion Ratio Lower Upper
225 100
223 61.6 50.3 75.5
227 59.8 49.1 73.7



#34
Caprolactam
Concen: 20.420 ng/ul
RT: 11.533 min Scan# 1453
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:113 Resp: 80485
Ion Ratio Lower Upper
113 100
55 189.4 167.9 251.9
56 144.6 131.3 196.9









#35
4-Chloro-3-methylphenol
Concen: 21.149 ng/ul
RT: 11.886 min Scan# 1513
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:107 Resp: 269803
Ion Ratio Lower Upper
107 100
144 25.8 18.7 28.1
142 81.6 62.7 94.1



#36
2-Methylnaphthalene
Concen: 21.477 ng/ul
RT: 12.269 min Scan# 1578
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:142 Resp: 614647
Ion Ratio Lower Upper
142 100
141 86.4 72.5 108.7





#37

1-Methylnaphthalene

Concen: 19.654 ng/ul

RT: 12.492 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Instrument :

BNA_M

ClientSampleId :

SICV026





Tgt Ion:142 Resp: 572242

Ion Ratio Lower Upper

142 100

141 90.0 73.4 110.2

116 4.0 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.363 ng/ul

RT: 12.639 min Scan# 1641

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Tgt Ion:216 Resp: 311924




Ion Ratio Lower Upper


216 100

214 77.4 61.0 91.4

179 20.4 17.7 26.5

108 18.9 17.0 25.4









#40
Hexachlorocyclopentadiene
Concen: 19.101 ng/ul
RT: 12.627 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:237 Resp: 176969
Ion Ratio Lower Upper
237 100
235 60.4 49.8 74.6
272 13.4 9.7 14.5



#41
2,4,6-Trichlorophenol
Concen: 21.908 ng/ul
RT: 12.880 min Scan# 1682
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:196 Resp: 194174
Ion Ratio Lower Upper
196 100
198 92.6 77.1 115.7
200 31.4 24.5 36.7





#42

2,4,5-Trichlorophenol

Concen: 21.462 ng/ul

RT: 12.945 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Instrument :

BNA_M

ClientSampleId :

SICV026





Tgt Ion:196 Resp: 208009

Ion Ratio Lower Upper

196 100

198 94.0 79.5 119.3

200 29.9 25.8 38.8



#43

1,1'-Biphenyl

Concen: 20.717 ng/ul

RT: 13.280 min Scan# 1750

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13




Tgt Ion:154 Resp: 780535


Ion Ratio Lower Upper

154 100

153 38.7 31.1 46.7

76 13.1 12.1 18.1









#44
2-Chloronaphthalene
Concen: 20.289 ng/ul
RT: 13.322 min Scan# 1757
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:162 Resp: 591210
Ion Ratio Lower Upper
162 100
164 32.4 25.3 37.9
127 39.4 32.9 49.3



#45
2-Nitroaniline
Concen: 20.240 ng/ul
RT: 13.521 min Scan# 1791
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion: 65 Resp: 187224
Ion Ratio Lower Upper
65 100
92 64.8 49.1 73.7
138 97.5 67.6 101.4









#47
Dimethylphthalate
Concen: 20.699 ng/ul
RT: 13.904 min Scan# 1856
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:163 Resp: 715169
Ion Ratio Lower Upper
163 100
194 3.9 3.0 4.6
164 9.9 7.8 11.6



#48
2,6-Dinitrotoluene
Concen: 23.930 ng/ul
RT: 14.016 min Scan# 1875
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:165 Resp: 145206
Ion Ratio Lower Upper
165 100
89 65.4 58.5 87.7
121 21.8 20.2 30.2









#50
Acenaphthylene
Concen: 20.908 ng/ul
RT: 14.169 min Scan# 1901
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:152 Resp: 953782
Ion Ratio Lower Upper
152 100
151 19.4 16.2 24.2
153 13.1 10.7 16.1



#51
3-Nitroaniline
Concen: 22.664 ng/ul
RT: 14.345 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:138 Resp: 166449
Ion Ratio Lower Upper
138 100
108 10.4 9.0 13.6
92 120.4 101.5 152.3









#52
Acenaphthene
Concen: 20.045 ng/ul
RT: 14.516 min Scan# 1959
Delta R.T. 0.006 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:153 Resp: 646417
Ion Ratio Lower Upper
153 100
152 48.7 39.7 59.5
154 86.9 72.6 108.8



#53
2,4-Dinitrophenol
Concen: 18.473 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. 0.006 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:184 Resp: 69266
Ion Ratio Lower Upper
184 100
63 73.8 77.5 116.3#
154 62.3 55.8 83.6









#55
4-Nitrophenol
Concen: 18.984 ng/ul
RT: 14.645 min Scan# 1983
Delta R.T. -0.006 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:109 Resp: 108269
Ion Ratio Lower Upper
109 100
139 132.1 105.7 158.5
65 149.4 123.1 184.7



#56
Dibenzofuran
Concen: 20.586 ng/ul
RT: 14.845 min Scan# 2016
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:168 Resp: 893598
Ion Ratio Lower Upper
168 100
139 37.8 31.0 46.6





#57

2,4-Dinitrotoluene

Concen: 21.965 ng/ul

RT: 14.804 min Scan# 2009

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Instrument :

BNA_M

ClientSampleId :

SICV026





Tgt Ion:165 Resp: 200184

Ion Ratio Lower Upper

165 100

63 49.2 45.8 68.6

182 2.5 1.7 2.5



#58

2,3,4,6-Tetrachlorophenol

Concen: 23.477 ng/ul

RT: 15.068 min Scan# 2054

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Tgt Ion:232 Resp: 179362




Ion Ratio Lower Upper


232 100

131 48.9 40.2 60.4

130 2.5 2.6 4.0#

166 31.2 25.3 37.9





#59

Diethylphthalate

Concen: 20.484 ng/ul

RT: 15.268 min Scan# 2088

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Instrument :

BNA_M

ClientSampleId :

SICV026





Tgt Ion:149 Resp: 708142

Ion Ratio Lower Upper

149 100

177 22.1 16.9 25.3

150 12.3 9.2 13.8



#61

Fluorene

Concen: 19.853 ng/ul

RT: 15.492 min Scan# 2126

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13




Tgt Ion:166 Resp: 718988


Ion Ratio Lower Upper

166 100

165 93.5 77.8 116.6

167 13.3 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 20.315 ng/ul
RT: 15.486 min Scan# 2125
Delta R.T. -0.006 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:204 Resp: 347635
Ion Ratio Lower Upper
204 100
206 32.3 27.0 40.4
141 61.0 49.2 73.8



#63
4-Nitroaniline
Concen: 24.241 ng/ul
RT: 15.504 min Scan# 2128
Delta R.T. 0.006 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:138 Resp: 172898
Ion Ratio Lower Upper
138 100
92 51.2 44.9 67.3
108 74.5 62.4 93.6









#66
4,6-Dinitro-2-methylphenol
Concen: 19.549 ng/ul
RT: 15.563 min Scan# 2138
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:198 Resp: 118461
Ion Ratio Lower Upper
198 100
182 5.6 5.1 7.7
77 29.1 26.2 39.2



#67
N-Nitrosodiphenylamine
Concen: 20.448 ng/ul
RT: 15.698 min Scan# 2161
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:169 Resp: 595471
Ion Ratio Lower Upper
169 100
168 67.2 53.5 80.3
167 37.6 30.0 45.0









#68
4-Bromophenyl-phenylether
Concen: 20.936 ng/ul
RT: 16.380 min Scan# 211
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:248 Resp: 201680
Ion Ratio Lower Upper
248 100
250 94.9 76.5 114.7
141 72.8 60.8 91.2



#69
Hexachlorobenzene
Concen: 21.181 ng/ul
RT: 16.498 min Scan# 2297
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:284 Resp: 237444
Ion Ratio Lower Upper
284 100
142 36.3 32.4 48.6
249 30.8 26.3 39.5





#70

Atrazine

Concen: 20.632 ng/ul

RT: 16.645 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Instrument :

BNA_M

ClientSampleId :

SICV026





Tgt Ion:200 Resp: 200005

Ion Ratio Lower Upper

200 100

173 28.7 20.2 30.2

215 47.8 39.4 59.2



#71

Pentachlorophenol

Concen: 20.607 ng/ul

RT: 16.833 min Scan# 2354

Delta R.T. -0.006 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13




Tgt Ion:266 Resp: 146891


Ion Ratio Lower Upper

266 100

264 64.0 50.2 75.2

268 65.7 50.9 76.3





#72

Phenanthrene

Concen: 19.662 ng/ul

RT: 17.227 min Scan# 2421

Delta R.T. -0.000 min

Lab File: BM047685.D


Acq: 27 Sep 2024 22:13

Instrument :

BNA_M

ClientSampleId :

SICV026






Tgt Ion:178 Resp: 1077575

Ion Ratio Lower Upper

178 100

179 15.4 12.0 18.0

176 19.0 15.4 23.0



#74

Anthracene

Concen: 19.800 ng/ul

RT: 17.315 min Scan# 2436

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13




Tgt Ion:178 Resp: 1106739


Ion Ratio Lower Upper

178 100

179 15.4 12.4 18.6

176 19.0 14.7 22.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.586 ng/uL
RT: 13.245 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:216 Resp: 287586
Ion Ratio Lower Upper
216 100
214 78.6 63.0 94.4
218 48.5 37.6 56.4



#76
Pentachlorobenzene
Concen: 19.440 ng/uL
RT: 14.768 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:250 Resp: 293607
Ion Ratio Lower Upper
250 100
248 64.0 50.2 75.2
252 66.2 49.4 74.0





#77

Carbazole

Concen: 19.941 ng/ul

RT: 17.580 min Scan# 2481

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Instrument :

BNA_M

ClientSampleId :

SICV026





Tgt Ion:167 Resp: 1010305

Ion Ratio Lower Upper

167 100

166 21.5 17.4 26.0

139 13.5 10.9 16.3



#78

Di-n-butylphthalate

Concen: 20.371 ng/ul

RT: 18.151 min Scan# 2578

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13




Tgt Ion:149 Resp: 1149067


Ion Ratio Lower Upper

149 100

150 9.3 7.4 11.2

104 5.3 4.6 7.0





#80

Fluoranthene

Concen: 20.851 ng/ul

RT: 19.233 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Instrument :

BNA_M

ClientSampleId :

SICV026





Tgt Ion:202 Resp: 1215911

Ion Ratio Lower Upper

202 100

101 11.3 9.0 13.4

100 8.5 6.8 10.2



#82

Pyrene

Concen: 20.449 ng/ul

RT: 19.598 min Scan# 2824

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13




Tgt Ion:202 Resp: 1327204


Ion Ratio Lower Upper

202 100

101 13.0 10.9 16.3

100 10.6 9.4 14.2





#83

Butylbenzylphthalate

Concen: 21.007 ng/ul

RT: 20.492 min Scan# 2976

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Instrument :

BNA_M

ClientSampleId :

SICV026





Tgt Ion:149 Resp: 482020

Ion Ratio Lower Upper

149 100

91 76.6 66.9 100.3

206 19.3 14.4 21.6



#84

3,3'-Dichlorobenzidine

Concen: 22.676 ng/ul

RT: 21.315 min Scan# 3116

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13




Tgt Ion:252 Resp: 418019


Ion Ratio Lower Upper

252 100

254 64.9 51.5 77.3

126 10.7 9.4 14.0





#85

Benzo(a)anthracene

Concen: 20.525 ng/ul

RT: 21.392 min Scan# 3129

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Instrument :

BNA_M

ClientSampleId :

SICV026





Tgt Ion:228 Resp: 1253746

Ion Ratio Lower Upper

228 100

229 19.7 15.6 23.4

226 27.4 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.843 ng/ul

RT: 21.321 min Scan# 3117

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13




Tgt Ion:149 Resp: 718034


Ion Ratio Lower Upper

149 100

167 27.9 21.3 31.9

279 3.9 2.7 4.1









#87
Chrysene
Concen: 20.325 ng/ul
RT: 21.450 min Scan# 3139
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13




Instrument :
BNA_M
ClientSampleId :
SICV026


Tgt Ion:228 Resp: 1210353
Ion Ratio Lower Upper
228 100
226 30.7 23.6 35.4
229 19.4 15.3 22.9



#89
Di-n-octyl phthalate
Concen: 18.568 ng/ul
RT: 22.456 min Scan# 3310
Delta R.T. -0.000 min
Lab File: BM047685.D
Acq: 27 Sep 2024 22:13

Tgt Ion:149 Resp: 1144884





#90

Benzo(b)fluoranthene

Concen: 20.229 ng/ul

RT: 23.462 min Scan# 3481

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Instrument :

BNA_M

ClientSampleId :

SICV026





Tgt Ion:252 Resp: 1205917

Ion Ratio Lower Upper

252 100

253 22.3 17.4 26.2

125 10.2 9.0 13.4



#91

Benzo(k)fluoranthene

Concen: 20.484 ng/ul

RT: 23.527 min Scan# 3492

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13




Tgt Ion:252 Resp: 1248609


Ion Ratio Lower Upper

252 100

253 22.0 17.5 26.3

125 9.9 8.9 13.3





#93

Benzo(a)pyrene

Concen: 20.863 ng/ul

RT: 24.274 min Scan# 3

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Instrument :

BNA_M

ClientSampleId :

SICV026





Tgt Ion:252 Resp: 1181199

Ion Ratio Lower Upper

252 100

253 21.5 17.2 25.8

125 10.8 10.1 15.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.342 ng/ul

RT: 27.803 min Scan# 4219

Delta R.T. -0.006 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13




Tgt Ion:276 Resp: 1475890


Ion Ratio Lower Upper

276 100

138 20.7 16.9 25.3

277 23.8 18.9 28.3





#95

Dibenzo(a,h)anthracene

Concen: 20.532 ng/ul

RT: 27.868 min Scan# 41

Delta R.T. 0.006 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Instrument :

BNA_M

ClientSampleId :

SICV026





Tgt Ion:278 Resp: 1220860

Ion Ratio Lower Upper

278 100

139 16.3 13.4 20.2

279 23.8 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 19.920 ng/ul

RT: 28.862 min Scan# 4399

Delta R.T. -0.000 min

Lab File: BM047685.D

Acq: 27 Sep 2024 22:13

Tgt Ion:276 Resp: 1132381

Ion Ratio Lower Upper

276 100

138 23.5 18.7 28.1

277 22.9 18.3 27.5

