

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092823\
 Data File : BM042126.D
 Acq On : 28 Sep 2023 22:03
 Operator : MA/JU
 Sample : 04480-12
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BBJY8

Quant Time: Sep 29 04:41:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 29 04:39:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.937	152	3653	0.400	ng/ul	0.00
4) Naphthalene-d8	10.752	136	9921	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.583	164	4989	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.337	188	12246	0.400	ng/ul	0.00
17) Chrysene-d12	21.509	240	7264	0.400	ng/ul	0.00
23) Perylene-d12	23.912	264	7983	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.360	96	25530	5.129	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.335	152	6729	0.542	ng/ul	0.00
18) Fluoranthene-d10	19.357	212	13311	0.445	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.397	88	2683	0.512	ng/ul#	91
5) Naphthalene	10.801	128	30741	0.925	ng/ul	99
7) 2-Methylnaphthalene	12.407	142	3942	0.199	ng/ul	100
8) 1-Methylnaphthalene	12.627	142	12877	0.640	ng/ul	99
11) Acenaphthene	14.647	153	585	0.023	ng/ul#	92
12) Fluorene	15.633	166	802	0.028	ng/ul#	87
15) Phenanthrene	17.379	178	1965	0.035	ng/ul#	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092823\
Data File : BM042126.D
Acq On : 28 Sep 2023 22:03
Operator : MA/JU
Sample : 04480-12
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BBJY8

Quant Time: Sep 29 04:41:48 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 29 04:39:33 2023
Response via : Initial Calibration

