

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092823\
 Data File : BM042135.D
 Acq On : 29 Sep 2023 03:28
 Operator : MA/JU
 Sample : PB155824BS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS824

Quant Time: Sep 29 04:43:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 29 04:39:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.937	152	3228	0.400	ng/ul	0.00
4) Naphthalene-d8	10.752	136	8090	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.583	164	3944	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.337	188	8659	0.400	ng/ul	0.00
17) Chrysene-d12	21.510	240	5566	0.400	ng/ul	0.00
23) Perylene-d12	23.912	264	5618	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	4426	1.006	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.335	152	5324	0.526	ng/ul	0.00
18) Fluoranthene-d10	19.362	212	10022	0.437	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.397	88	11494	2.484	ng/ul#	86
5) Naphthalene	10.801	128	12185	0.450	ng/ul	99
7) 2-Methylnaphthalene	12.407	142	7583	0.469	ng/ul	98
8) 1-Methylnaphthalene	12.627	142	7327	0.447	ng/ul	99
10) Acenaphthylene	14.310	152	12199	0.462	ng/ul	100
11) Acenaphthene	14.648	153	8845	0.437	ng/ul	98
12) Fluorene	15.638	166	10121	0.443	ng/ul	100
14) Pentachlorophenol	16.970	266	3487	0.761	ng/ul	99
15) Phenanthrene	17.379	178	16855	0.420	ng/ul	99
16) Anthracene	17.472	178	15361	0.443	ng/ul	99
19) Fluoranthene	19.389	202	20474	0.325	ng/ul	99
20) Pyrene	19.757	202	22070	0.357	ng/ul	99
21) Benzo(a)anthracene	21.495	228	16363	0.388	ng/ul	99
22) Chrysene	21.547	228	17158	0.394	ng/ul	99
24) Benzo(b)fluoranthene	23.170	252	17398	0.402	ng/ul	99
25) Benzo(k)fluoranthene	23.219	252	17349	0.413	ng/ul	99
26) Benzo(a)pyrene	23.807	252	15321	0.431	ng/ul	98
27) Indeno(1,2,3-cd)pyrene	26.401	276	19068	0.391	ng/ul#	97
28) Dibenzo(a,h)anthracene	26.421	278	14025	0.398	ng/ul	98
29) Benzo(g,h,i)perylene	27.176	276	16155	0.393	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092823\
Data File : BM042135.D
Acq On : 29 Sep 2023 03:28
Operator : MA/JU
Sample : PB155824BS
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS824

Quant Time: Sep 29 04:43:21 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 29 04:39:33 2023
Response via : Initial Calibration

