

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092824\
 Data File : BM047687.D
 Acq On : 28 Sep 2024 12:04
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020147

Quant Time: Sep 29 00:38:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 28 02:30:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	169777	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	700379	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	410094	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	792268	20.000	ng/u1	0.00
79) Chrysene-d12	21.409	240	748204	20.000	ng/u1	0.00
88) Perylene-d12	24.415	264	801232	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.257	96	43397	8.260	ng/uL	0.00
4) Pyridine-d5	3.675	84	311475	20.249	ng/u1	0.00
7) Phenol-d5	6.975	99	313321	19.517	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	228268	20.038	ng/u1	0.00
11) 2-Chlorophenol-d4	7.345	132	241903	20.710	ng/u1	0.00
15) 4-Methylphenol-d8	8.516	113	231656	19.408	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	112116	20.105	ng/u1	0.00
24) 2-Nitrophenol-d4	9.692	143	107428	20.103	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	214542	20.097	ng/u1	0.00
31) 4-Chloroaniline-d4	10.739	131	320222	19.403	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	597607	20.100	ng/u1	0.00
49) Acenaphthylene-d8	14.139	160	711579	20.185	ng/u1	0.00
54) 4-Nitrophenol-d4	14.633	143	112049	18.928	ng/u1	0.00
60) Fluorene-d10	15.439	176	509344	19.778	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.545	200	82425	17.778	ng/u1	0.00
73) Anthracene-d10	17.280	188	775956	19.918	ng/u1	0.00
81) Pyrene-d10	19.568	212	839217	19.763	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.209	264	825411	19.855	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	46657	8.131	ng/uL	97
5) Pyridine	3.693	79	332004	20.501	ng/u1	99
6) Benzaldehyde	6.951	77	214650	24.586	ng/u1	99
8) Phenol	7.004	94	336779	19.715	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.234	93	274274	20.150	ng/u1	99
12) 2-Chlorophenol	7.381	128	256611	20.738	ng/u1	98
13) 2-Methylphenol	8.251	108	230655	19.365	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.334	45	517407	20.193	ng/u1	99
16) Acetophenone	8.634	105	403348	19.749	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.622	70	221374	20.464	ng/u1	97
18) 4-Methylphenol	8.581	108	263110	19.656	ng/u1	95
19) Hexachloroethane	8.898	117	112286	20.188	ng/u1	96
22) Nitrobenzene	9.010	77	321797	20.314	ng/u1	98
23) Isophorone	9.539	82	574059	20.258	ng/u1	99
25) 2-Nitrophenol	9.722	139	124937	20.436	ng/u1	99
26) 2,4-Dimethylphenol	9.781	107	267464	20.418	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.016	93	351395	20.274	ng/u1	97
29) 2,4-Dichlorophenol	10.257	162	219681	19.929	ng/u1	99
30) Naphthalene	10.657	128	799064	19.835	ng/u1	100
32) 4-Chloroaniline	10.763	127	316558	19.480	ng/u1	100
33) Hexachlorobutadiene	10.951	225	145153	19.930	ng/u1	97
34) Caprolactam	11.527	113	64735	18.363	ng/u1	97
35) 4-Chloro-3-methylphenol	11.886	107	233568	20.469	ng/u1	98
36) 2-Methylnaphthalene	12.269	142	513171	20.048	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092824\
 Data File : BM047687.D
 Acq On : 28 Sep 2024 12:04
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD020147


Quant Time: Sep 29 00:38:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 28 02:30:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.486	142	517597	19.875	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.639	216	261587	19.595	ng/ul	99
40) Hexachlorocyclopentadiene	12.627	237	155033	19.201	ng/ul	100
41) 2,4,6-Trichlorophenol	12.880	196	157979	20.453	ng/ul	97
42) 2,4,5-Trichlorophenol	12.945	196	172927	20.474	ng/ul	98
43) 1,1'-Biphenyl	13.280	154	655714	19.970	ng/ul	99
44) 2-Chloronaphthalene	13.321	162	502205	19.776	ng/ul	99
45) 2-Nitroaniline	13.521	65	164912	20.457	ng/ul	97
47) Dimethylphthalate	13.904	163	602408	20.006	ng/ul	99
48) 2,6-Dinitrotoluene	14.016	165	107567	20.341	ng/ul	97
50) Acenaphthylene	14.168	152	800034	20.124	ng/ul	100
51) 3-Nitroaniline	14.345	138	126504	19.766	ng/ul	97
52) Acenaphthene	14.510	153	559712	19.916	ng/ul	97
53) 2,4-Dinitrophenol	14.545	184	52674	16.120	ng/ul	97
55) 4-Nitrophenol	14.645	109	97121	19.541	ng/ul	97
56) Dibenzofuran	14.845	168	751674	19.870	ng/ul	99
57) 2,4-Dinitrotoluene	14.804	165	161693	20.358	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.068	232	137601	20.667	ng/ul	98
59) Diethylphthalate	15.268	149	606470	20.130	ng/ul	99
61) Fluorene	15.492	166	629165	19.935	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.486	204	296484	19.881	ng/ul	98
63) 4-Nitroaniline	15.504	138	132726	21.353	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.563	198	95576	18.604	ng/ul#	97
67) N-Nitrosodiphenylamine	15.698	169	497986	20.171	ng/ul	99
68) 4-Bromophenyl-phenylether	16.380	248	162491	19.897	ng/ul	98
69) Hexachlorobenzene	16.498	284	187273	19.705	ng/ul	98
70) Atrazine	16.645	200	163949	19.949	ng/ul	99
71) Pentachlorophenol	16.833	266	110205	18.236	ng/ul	100
72) Phenanthrene	17.227	178	919776	19.797	ng/ul	100
74) Anthracene	17.315	178	954773	20.148	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.245	216	260917	19.890	ng/uL	99
76) Pentachlorobenzene	14.768	250	248284	19.391	ng/uL	98
77) Carbazole	17.580	167	842102	19.605	ng/ul	100
78) Di-n-butylphthalate	18.151	149	977542	20.442	ng/ul	99
80) Fluoranthene	19.239	202	1003760	20.070	ng/ul	100
82) Pyrene	19.598	202	1110001	19.941	ng/ul	98
83) Butylbenzylphthalate	20.492	149	401949	20.425	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.315	252	309007	19.545	ng/ul	99
85) Benzo(a)anthracene	21.392	228	1042147	19.893	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.321	149	603107	20.413	ng/ul	100
87) Chrysene	21.450	228	1007016	19.718	ng/ul	99
89) Di-n-octyl phthalate	22.456	149	969565	18.761	ng/ul	100
90) Benzo(b)fluoranthene	23.462	252	1004268	20.099	ng/ul	100
91) Benzo(k)fluoranthene	23.527	252	1007694	19.723	ng/ul	99
93) Benzo(a)pyrene	24.274	252	946037	19.936	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.803	276	1190957	19.584	ng/ul	100
95) Dibenzo(a,h)anthracene	27.862	278	969548	19.454	ng/ul	99
96) Benzo(g,h,i)perylene	28.868	276	927871	19.474	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 29 00:38:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 28 02:30:36 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.131 ng/uL
RT: 3.293 min Scan# 51
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion: 88 Resp: 46657
Ion Ratio Lower Upper
88 100
43 38.0 29.1 43.7
58 82.4 67.7 101.5



#5
Pyridine
Concen: 20.501 ng/uL
RT: 3.693 min Scan# 120
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion: 79 Resp: 332004
Ion Ratio Lower Upper
79 100
52 86.4 69.4 104.2
51 41.9 32.6 49.0









#6
Benzaldehyde
Concen: 24.586 ng/ul
RT: 6.951 min Scan# 674
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion: 77 Resp: 214650
Ion Ratio Lower Upper
77 100
105 90.8 72.2 108.2
106 88.0 70.7 106.1



#8
Phenol
Concen: 19.715 ng/ul
RT: 7.004 min Scan# 683
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion: 94 Resp: 336779
Ion Ratio Lower Upper
94 100
65 31.6 26.3 39.5
66 38.0 30.5 45.7









#10
Bis(2-Chloroethyl)ether
Concen: 20.150 ng/ul
RT: 7.234 min Scan# 722
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion: 93 Resp: 274274
Ion Ratio Lower Upper
93 100
63 86.3 69.4 104.2
95 32.2 25.0 37.6



#12
2-Chlorophenol
Concen: 20.738 ng/ul
RT: 7.381 min Scan# 747
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion: 128 Resp: 256611
Ion Ratio Lower Upper
128 100
64 55.1 42.7 64.1
130 33.0 26.0 39.0





#13

2-Methylphenol

Concen: 19.365 ng/ul

RT: 8.251 min Scan# 895

Delta R.T. -0.006 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :





SSTD020147

Tgt Ion:108 Resp: 230655

Ion Ratio Lower Upper

108 100

107 86.9 69.3 103.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.193 ng/ul

RT: 8.334 min Scan# 909

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04




Tgt Ion: 45 Resp: 517407


Ion Ratio Lower Upper

45 100

77 10.9 9.0 13.6

79 8.0 6.5 9.7





#16

Acetophenone

Concen: 19.749 ng/ul

RT: 8.634 min Scan# 960

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :

SSTD020147





Tgt Ion:105 Resp: 403348

Ion Ratio Lower Upper

105 100

77 85.5 66.6 100.0

51 33.6 27.0 40.4



#17

N-Nitroso-di-n-propylamine

Concen: 20.464 ng/ul

RT: 8.622 min Scan# 958

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Tgt Ion: 70 Resp: 221374




Ion Ratio Lower Upper


70 100

42 71.4 55.3 82.9

101 8.8 6.8 10.2

130 17.9 13.7 20.5









#18
4-Methylphenol
Concen: 19.656 ng/ul
RT: 8.581 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:108 Resp: 263110
Ion Ratio Lower Upper
108 100
107 105.4 88.4 132.6



#19
Hexachloroethane
Concen: 20.188 ng/ul
RT: 8.898 min Scan# 1005
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:117 Resp: 112286
Ion Ratio Lower Upper
117 100
201 91.1 71.0 106.6
199 59.4 44.2 66.4









#22
Nitrobenzene
Concen: 20.314 ng/ul
RT: 9.010 min Scan# 1102
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion: 77 Resp: 321797
Ion Ratio Lower Upper
77 100
123 38.4 31.7 47.5
65 13.8 11.1 16.7



#23
Isophorone
Concen: 20.258 ng/ul
RT: 9.539 min Scan# 1114
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion: 82 Resp: 574059
Ion Ratio Lower Upper
82 100
95 6.5 4.9 7.3
138 15.5 12.0 18.0









#25
2-Nitrophenol
Concen: 20.436 ng/ul
RT: 9.722 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:139 Resp: 124937
Ion Ratio Lower Upper
139 100
65 61.5 49.6 74.4
109 30.1 23.4 35.2



#26
2,4-Dimethylphenol
Concen: 20.418 ng/ul
RT: 9.781 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:107 Resp: 267464
Ion Ratio Lower Upper
107 100
121 49.0 37.8 56.8
122 82.9 65.4 98.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.274 ng/ul
 RT: 10.016 min Scan# 1195
 Delta R.T. -0.006 min
 Lab File: BM047687.D
 Acq: 28 Sep 2024 12:04




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020147


Tgt Ion: 93 Resp: 351395
 Ion Ratio Lower Upper
 93 100
 95 31.7 27.1 40.7
 123 10.6 8.2 12.4



#29
 2,4-Dichlorophenol
 Concen: 19.929 ng/ul
 RT: 10.257 min Scan# 1236
 Delta R.T. -0.000 min
 Lab File: BM047687.D
 Acq: 28 Sep 2024 12:04

Tgt Ion: 162 Resp: 219681
 Ion Ratio Lower Upper
 162 100
 164 65.6 52.2 78.2
 98 39.8 30.6 46.0









#30
Naphthalene
Concen: 19.835 ng/ul
RT: 10.657 min Scan# 1304
Delta R.T. -0.006 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:128 Resp: 799064
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 12.9 10.4 15.6



#32
4-Chloroaniline
Concen: 19.480 ng/ul
RT: 10.763 min Scan# 1322
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:127 Resp: 316558
Ion Ratio Lower Upper
127 100
129 32.4 25.9 38.9









#33
Hexachlorobutadiene
Concen: 19.930 ng/ul
RT: 10.951 min Scan# 1354
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:225 Resp: 145153
Ion Ratio Lower Upper
225 100
223 60.8 50.3 75.5
227 63.3 49.1 73.7



#34
Caprolactam
Concen: 18.363 ng/ul
RT: 11.527 min Scan# 1452
Delta R.T. -0.006 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:113 Resp: 64735
Ion Ratio Lower Upper
113 100
55 213.7 167.9 251.9
56 160.3 131.3 196.9









#35
4-Chloro-3-methylphenol
Concen: 20.469 ng/ul
RT: 11.886 min Scan# 1513
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:107 Resp: 233568
Ion Ratio Lower Upper
107 100
144 25.9 18.7 28.1
142 79.5 62.7 94.1



#36
2-Methylnaphthalene
Concen: 20.048 ng/ul
RT: 12.269 min Scan# 1578
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:142 Resp: 513171
Ion Ratio Lower Upper
142 100
141 88.1 72.5 108.7





#37

1-Methylnaphthalene

Concen: 19.875 ng/ul

RT: 12.486 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :

SSTD020147





Tgt Ion:142 Resp: 517597

Ion Ratio Lower Upper

142 100

141 92.4 73.4 110.2

116 4.2 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.595 ng/ul

RT: 12.639 min Scan# 1641

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Tgt Ion:216 Resp: 261587




Ion Ratio Lower Upper


216 100

214 77.2 61.0 91.4

179 20.5 17.7 26.5

108 20.9 17.0 25.4





#40

Hexachlorocyclopentadiene

Concen: 19.201 ng/ul

RT: 12.627 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :

SSTD020147





Tgt Ion:237 Resp: 155033

Ion Ratio Lower Upper

237 100

235 62.0 49.8 74.6

272 11.8 9.7 14.5



#41

2,4,6-Trichlorophenol

Concen: 20.453 ng/ul

RT: 12.880 min Scan# 1682

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04




Tgt Ion:196 Resp: 157979


Ion Ratio Lower Upper

196 100

198 98.5 77.1 115.7

200 33.4 24.5 36.7









#42
2,4,5-Trichlorophenol
Concen: 20.474 ng/ul
RT: 12.945 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:196 Resp: 172927
Ion Ratio Lower Upper
196 100
198 97.5 79.5 119.3
200 30.5 25.8 38.8



#43
1,1'-Biphenyl
Concen: 19.970 ng/ul
RT: 13.280 min Scan# 1750
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:154 Resp: 655714
Ion Ratio Lower Upper
154 100
153 39.3 31.1 46.7
76 15.0 12.1 18.1









#44
2-Chloronaphthalene
Concen: 19.776 ng/ul
RT: 13.321 min Scan# 1757
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:162 Resp: 502205
Ion Ratio Lower Upper
162 100
164 32.4 25.3 37.9
127 40.4 32.9 49.3



#45
2-Nitroaniline
Concen: 20.457 ng/ul
RT: 13.521 min Scan# 1791
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion: 65 Resp: 164912
Ion Ratio Lower Upper
65 100
92 62.3 49.1 73.7
138 88.0 67.6 101.4





#47

Dimethylphthalate

Concen: 20.006 ng/ul

RT: 13.904 min Scan# 1856

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :

SSTD020147





Tgt Ion:163 Resp: 602408

Ion Ratio Lower Upper

163 100

194 3.9 3.0 4.6

164 10.2 7.8 11.6



#48

2,6-Dinitrotoluene

Concen: 20.341 ng/ul

RT: 14.016 min Scan# 1875

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04




Tgt Ion:165 Resp: 107567


Ion Ratio Lower Upper

165 100

89 69.3 58.5 87.7

121 25.0 20.2 30.2









#50
Acenaphthylene
Concen: 20.124 ng/ul
RT: 14.168 min Scan# 1901
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:152 Resp: 800034
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2
153 13.1 10.7 16.1



#51
3-Nitroaniline
Concen: 19.766 ng/ul
RT: 14.345 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:138 Resp: 126504
Ion Ratio Lower Upper
138 100
108 10.3 9.0 13.6
92 123.6 101.5 152.3





#52

Acenaphthene

Concen: 19.916 ng/ul

RT: 14.510 min Scan# 1959

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :

SSTD020147





Tgt Ion:153 Resp: 559712

Ion Ratio Lower Upper

153 100

152 47.6 39.7 59.5

154 87.9 72.6 108.8



#53

2,4-Dinitrophenol

Concen: 16.120 ng/ul

RT: 14.545 min Scan# 1965

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04




Tgt Ion:184 Resp: 52674


Ion Ratio Lower Upper

184 100

63 93.8 77.5 116.3

154 67.2 55.8 83.6









#55
4-Nitrophenol
Concen: 19.541 ng/ul
RT: 14.645 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:109 Resp: 97121
Ion Ratio Lower Upper
109 100
139 126.4 105.7 158.5
65 156.8 123.1 184.7



#56
Dibenzofuran
Concen: 19.870 ng/ul
RT: 14.845 min Scan# 2016
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:168 Resp: 751674
Ion Ratio Lower Upper
168 100
139 39.6 31.0 46.6





#57

2,4-Dinitrotoluene

Concen: 20.358 ng/ul

RT: 14.804 min Scan# 2009

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :

SSTD020147





Tgt Ion:165 Resp: 161693

Ion Ratio Lower Upper

165 100

63 55.5 45.8 68.6

182 2.4 1.7 2.5



#58

2,3,4,6-Tetrachlorophenol

Concen: 20.667 ng/ul

RT: 15.068 min Scan# 2054

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Tgt Ion:232 Resp: 137601




Ion Ratio Lower Upper


232 100

131 51.6 40.2 60.4

130 2.8 2.6 4.0

166 32.2 25.3 37.9









#59
Diethylphthalate
Concen: 20.130 ng/ul
RT: 15.268 min Scan# 2088
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:149 Resp: 606470
Ion Ratio Lower Upper
149 100
177 21.2 16.9 25.3
150 12.3 9.2 13.8



#61
Fluorene
Concen: 19.935 ng/ul
RT: 15.492 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:166 Resp: 629165
Ion Ratio Lower Upper
166 100
165 97.4 77.8 116.6
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 19.881 ng/ul
RT: 15.486 min Scan# 2125
Delta R.T. -0.006 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:204 Resp: 296484
Ion Ratio Lower Upper
204 100
206 32.5 27.0 40.4
141 63.1 49.2 73.8



#63
4-Nitroaniline
Concen: 21.353 ng/ul
RT: 15.504 min Scan# 2128
Delta R.T. 0.006 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:138 Resp: 132726
Ion Ratio Lower Upper
138 100
92 54.6 44.9 67.3
108 78.2 62.4 93.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.604 ng/ul
 RT: 15.563 min Scan# 2138
 Delta R.T. -0.000 min
 Lab File: BM047687.D
 Acq: 28 Sep 2024 12:04




Instrument :
 BNA_M
 ClientSampleId :
 SST020147


Tgt Ion:198 Resp: 95576
 Ion Ratio Lower Upper
 198 100
 182 5.1 5.1 7.7#
 77 30.9 26.2 39.2



#67
 N-Nitrosodiphenylamine
 Concen: 20.171 ng/ul
 RT: 15.698 min Scan# 2161
 Delta R.T. -0.000 min
 Lab File: BM047687.D
 Acq: 28 Sep 2024 12:04

Tgt Ion:169 Resp: 497986
 Ion Ratio Lower Upper
 169 100
 168 66.2 53.5 80.3
 167 37.0 30.0 45.0









#68
4-Bromophenyl-phenylether
Concen: 19.897 ng/ul
RT: 16.380 min Scan# 2277
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:248 Resp: 162491
Ion Ratio Lower Upper
248 100
250 97.2 76.5 114.7
141 78.6 60.8 91.2



#69
Hexachlorobenzene
Concen: 19.705 ng/ul
RT: 16.498 min Scan# 2297
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:284 Resp: 187273
Ion Ratio Lower Upper
284 100
142 39.5 32.4 48.6
249 31.1 26.3 39.5





#70

Atrazine

Concen: 19.949 ng/ul

RT: 16.645 min Scan# 2322

Delta R.T. -0.006 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :

SSTD020147





Tgt Ion:200 Resp: 163949

Ion Ratio Lower Upper

200 100

173 25.9 20.2 30.2

215 48.6 39.4 59.2



#71

Pentachlorophenol

Concen: 18.236 ng/ul

RT: 16.833 min Scan# 2354

Delta R.T. -0.006 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04




Tgt Ion:266 Resp: 110205


Ion Ratio Lower Upper

266 100

264 62.7 50.2 75.2

268 63.8 50.9 76.3





#72

Phenanthrene

Concen: 19.797 ng/ul

RT: 17.227 min Scan# 2421

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :

SSTD020147





Tgt Ion:178 Resp: 919776

Ion Ratio Lower Upper

178 100

179 15.0 12.0 18.0

176 19.0 15.4 23.0



#74

Anthracene

Concen: 20.148 ng/ul

RT: 17.315 min Scan# 2436

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04




Tgt Ion:178 Resp: 954773


Ion Ratio Lower Upper

178 100

179 15.5 12.4 18.6

176 18.8 14.7 22.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.890 ng/uL
RT: 13.245 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:216 Resp: 260917
Ion Ratio Lower Upper
216 100
214 77.1 63.0 94.4
218 47.3 37.6 56.4



#76
Pentachlorobenzene
Concen: 19.391 ng/uL
RT: 14.768 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:250 Resp: 248284
Ion Ratio Lower Upper
250 100
248 61.3 50.2 75.2
252 63.5 49.4 74.0









#77
 Carbazole
 Concen: 19.605 ng/ul
 RT: 17.580 min Scan# 2481
 Delta R.T. -0.000 min
 Lab File: BM047687.D
 Acq: 28 Sep 2024 12:04




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020147


Tgt Ion:167 Resp: 842102
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.4 26.0
 139 13.3 10.9 16.3



#78
 Di-n-butylphthalate
 Concen: 20.442 ng/ul
 RT: 18.151 min Scan# 2578
 Delta R.T. -0.000 min
 Lab File: BM047687.D
 Acq: 28 Sep 2024 12:04

Tgt Ion:149 Resp: 977542
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.4 11.2
 104 5.6 4.6 7.0









#80
Fluoranthene
Concen: 20.070 ng/ul
RT: 19.239 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:202 Resp: 1003760
Ion Ratio Lower Upper
202 100
101 11.0 9.0 13.4
100 8.4 6.8 10.2



#82
Pyrene
Concen: 19.941 ng/ul
RT: 19.598 min Scan# 2824
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:202 Resp: 1110001
Ion Ratio Lower Upper
202 100
101 12.9 10.9 16.3
100 11.1 9.4 14.2





#83

Butylbenzylphthalate

Concen: 20.425 ng/ul

RT: 20.492 min Scan# 2976

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :

SSTD020147





Tgt Ion:149 Resp: 401949

Ion Ratio Lower Upper

149 100

91 83.5 66.9 100.3

206 17.4 14.4 21.6



#84

3,3'-Dichlorobenzidine

Concen: 19.545 ng/ul

RT: 21.315 min Scan# 3116

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04




Tgt Ion:252 Resp: 309007


Ion Ratio Lower Upper

252 100

254 63.9 51.5 77.3

126 11.5 9.4 14.0









#85
Benzo(a)anthracene
Concen: 19.893 ng/ul
RT: 21.392 min Scan# 3129
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04




Instrument :
BNA_M
ClientSampleId :
SSTD020147


Tgt Ion:228 Resp: 1042147
Ion Ratio Lower Upper
228 100
229 18.9 15.6 23.4
226 27.7 21.9 32.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.413 ng/ul
RT: 21.321 min Scan# 3117
Delta R.T. -0.000 min
Lab File: BM047687.D
Acq: 28 Sep 2024 12:04

Tgt Ion:149 Resp: 603107
Ion Ratio Lower Upper
149 100
167 26.8 21.3 31.9
279 3.5 2.7 4.1





#87

Chrysene

Concen: 19.718 ng/ul

RT: 21.450 min Scan# 3139

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :

SSTD020147





Tgt Ion:228 Resp: 1007016

Ion Ratio Lower Upper

228 100

226 30.0 23.6 35.4

229 19.7 15.3 22.9



#89

Di-n-octyl phthalate

Concen: 18.761 ng/ul




RT: 22.456 min Scan# 3310


Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Tgt Ion:149 Resp: 969565





#90

Benzo(b)fluoranthene

Concen: 20.099 ng/ul

RT: 23.462 min Scan# 3481

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :

SSTD020147





Tgt Ion:252 Resp: 1004268

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.2

125 11.5 9.0 13.4



#91

Benzo(k)fluoranthene

Concen: 19.723 ng/ul

RT: 23.527 min Scan# 3492

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04




Tgt Ion:252 Resp: 1007694


Ion Ratio Lower Upper

252 100

253 21.5 17.5 26.3

125 10.5 8.9 13.3





#93

Benzo(a)pyrene

Concen: 19.936 ng/ul

RT: 24.274 min Scan# 3619

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :

SSTD020147





Tgt Ion:252 Resp: 946037

Ion Ratio Lower Upper

252 100

253 21.2 17.2 25.8

125 12.0 10.1 15.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.584 ng/ul

RT: 27.803 min Scan# 4219

Delta R.T. -0.006 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04




Tgt Ion:276 Resp: 1190957


Ion Ratio Lower Upper

276 100

138 21.0 16.9 25.3

277 23.5 18.9 28.3





#95

Dibenzo(a,h)anthracene

Concen: 19.454 ng/ul

RT: 27.862 min Scan# 4229

Delta R.T. -0.000 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Instrument :

BNA_M

ClientSampleId :

SSTD020147





Tgt Ion:278 Resp: 969548

Ion Ratio Lower Upper

278 100

139 16.5 13.4 20.2

279 23.5 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 19.474 ng/ul

RT: 28.868 min Scan# 4400

Delta R.T. 0.006 min

Lab File: BM047687.D

Acq: 28 Sep 2024 12:04

Tgt Ion:276 Resp: 927871

Ion Ratio Lower Upper

276 100

138 22.6 18.7 28.1

277 23.5 18.3 27.5

