

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092824\
 Data File : BM047702.D
 Acq On : 28 Sep 2024 21:57
 Operator : RC/JU
 Sample : P3927-04DL 5X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DDC82DL

Quant Time: Sep 29 00:42:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 28 02:30:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	219364	20.000	ng/ul	0.00
20) Naphthalene-d8	10.610	136	883801	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.445	164	544263	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.186	188	1114103	20.000	ng/ul	0.00
79) Chrysene-d12	21.409	240	1016356	20.000	ng/ul	0.00
88) Perylene-d12	24.415	264	1175542	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.263	96	3641	0.536	ng/uL	0.00
4) Pyridine-d5	3.681	84	28292	1.423	ng/ul	0.00
7) Phenol-d5	6.975	99	54166	2.611	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	41229	2.801	ng/ul	0.00
11) 2-Chlorophenol-d4	7.345	132	42843	2.839	ng/ul	0.00
15) 4-Methylphenol-d8	8.516	113	41939	2.719	ng/ul	0.00
21) Nitrobenzene-d5	8.969	128	17691	2.514	ng/ul	0.00
24) 2-Nitrophenol-d4	9.686	143	16936	2.511	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	37949	2.817	ng/ul	0.00
31) 4-Chloroaniline-d4	10.739	131	48328	2.321	ng/ul	0.00
46) Dimethylphthalate-d6	13.851	166	130600	3.310	ng/ul	0.00
49) Acenaphthylene-d8	14.139	160	140237	2.997	ng/ul	0.00
54) 4-Nitrophenol-d4	14.633	143	11252	1.432	ng/ul	0.00
60) Fluorene-d10	15.433	176	112996	3.306	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.545	200	10161	1.559	ng/ul	0.00
73) Anthracene-d10	17.280	188	170496	3.112	ng/ul	0.00
81) Pyrene-d10	19.568	212	196226	3.402	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.203	264	186654	3.060	ng/ul	0.00
Target Compounds						
58) 2,3,4,6-Tetrachlorophenol	15.068	232	29130	3.297	ng/ul	98
71) Pentachlorophenol	16.833	266	338465	39.829	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092824\
Data File : BM047702.D
Acq On : 28 Sep 2024 21:57
Operator : RC/JU
Sample : P3927-04DL 5X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DDC82DL

Quant Time: Sep 29 00:42:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 28 02:30:36 2024
Response via : Initial Calibration









#58
2,3,4,6-Tetrachlorophenol
Concen: 3.297 ng/ul
RT: 15.068 min Scan# 2054
Delta R.T. -0.000 min
Lab File: BM047702.D
Acq: 28 Sep 2024 21:57

Instrument :
BNA_M
ClientSampleId :
DDC82DL

Tgt Ion:	232	Resp:	29130
Ion Ratio	Lower	Upper	
232	100		
131	48.6	40.2	60.4
130	2.8	2.6	4.0
166	31.9	25.3	37.9



#71
Pentachlorophenol
Concen: 39.829 ng/ul
RT: 16.833 min Scan# 2354
Delta R.T. -0.006 min
Lab File: BM047702.D
Acq: 28 Sep 2024 21:57

Tgt Ion:	266	Resp:	338465
Ion Ratio	Lower	Upper	
266	100		
264	63.8	50.2	75.2
268	64.1	50.9	76.3

