

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092922\
 Data File : BM036797.D
 Acq On : 29 Sep 2022 12:24
 Operator : CG/JU
 Sample : SST01079
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST010027

Quant Time: Sep 30 01:12:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 30 01:10:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.981	152	282318	20.000	ng/ul	0.00
20) Naphthalene-d8	10.792	136	1285188	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.610	164	822957	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.345	188	1687848	20.000	ng/ul	0.00
79) Chrysene-d12	21.515	240	1708605	20.000	ng/ul	0.00
88) Perylene-d12	23.933	264	1740957	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	31709	4.210	ng/uL	0.00
4) Pyridine-d5	3.799	84	211948	9.650	ng/ul	0.00
7) Phenol-d5	7.139	99	242490	9.611	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.310	67	166771	10.885	ng/ul	0.00
11) 2-Chlorophenol-d4	7.510	132	177416	9.768	ng/ul	0.00
15) 4-Methylphenol-d8	8.692	113	188404	10.106	ng/ul	0.00
21) Nitrobenzene-d5	9.145	128	88675	10.194	ng/ul	0.00
24) 2-Nitrophenol-d4	9.875	143	81757	10.493	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.410	165	167383	9.270	ng/ul	0.00
31) 4-Chloroaniline-d4	10.927	131	295624	9.847	ng/ul	0.00
46) Dimethylphthalate-d6	14.021	166	552408	9.953	ng/ul	0.00
49) Acenaphthylene-d8	14.304	160	612999	9.311	ng/ul	0.00
54) 4-Nitrophenol-d4	14.798	143	102576	9.528	ng/ul	0.00
60) Fluorene-d10	15.598	176	485194	9.755	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.710	200	64354	9.637	ng/ul	0.00
73) Anthracene-d10	17.445	188	732601	9.493	ng/ul	0.00
81) Pyrene-d10	19.727	212	815847	9.706	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.780	264	810137	9.525	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.404	88	32262	3.977	ng/uL	96
5) Pyridine	3.816	79	212458	9.628	ng/ul	95
6) Benzaldehyde	7.122	77	134055	10.973	ng/ul	99
8) Phenol	7.169	94	262025	9.677	ng/ul	100
10) Bis(2-Chloroethyl)ether	7.404	93	220422	10.462	ng/ul	99
12) 2-Chlorophenol	7.545	128	197306	10.042	ng/ul	97
13) 2-Methylphenol	8.428	108	191033	9.598	ng/ul	95
14) 2,2'-oxybis(1-Chloropr...	8.516	45	303213	12.143	ng/ul	99
16) Acetophenone	8.810	105	333989	10.329	ng/ul	95
17) N-Nitroso-di-n-propyla...	8.792	70	174681	11.196	ng/ul	98
18) 4-Methylphenol	8.751	108	215858	10.030	ng/ul	97
19) Hexachloroethane	9.069	117	82062	10.311	ng/ul	98
22) Nitrobenzene	9.192	77	250013	10.825	ng/ul	98
23) Isophorone	9.722	82	470276	9.986	ng/ul	100
25) 2-Nitrophenol	9.904	139	98631	10.249	ng/ul	97
26) 2,4-Dimethylphenol	9.963	107	236628	9.969	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.204	93	303932	10.828	ng/ul	99
29) 2,4-Dichlorophenol	10.433	162	180710	9.519	ng/ul	99
30) Naphthalene	10.845	128	710818	10.153	ng/ul	100
32) 4-Chloroaniline	10.951	127	310953	10.032	ng/ul	99
33) Hexachlorobutadiene	11.127	225	106561	9.178	ng/ul	96
34) Caprolactam	11.733	113	56315	9.145	ng/ul	97
35) 4-Chloro-3-methylphenol	12.069	107	211798	10.107	ng/ul	99
36) 2-Methylnaphthalene	12.445	142	466005	9.990	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092922\
 Data File : BM036797.D
 Acq On : 29 Sep 2022 12:24
 Operator : CG/JU
 Sample : SSTD01079
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010027

Quant Time: Sep 30 01:12:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 30 01:10:23 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.663	142	469221	10.006	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.810	216	215970	9.272	ng/ul	98
40) Hexachlorocyclopentadiene	12.792	237	110868	9.133	ng/ul	95
41) 2,4,6-Trichlorophenol	13.045	196	129855	9.105	ng/ul	97
42) 2,4,5-Trichlorophenol	13.116	196	144940	9.479	ng/ul	99
43) 1,1'-Biphenyl	13.451	154	613195	10.108	ng/ul	100
44) 2-Chloronaphthalene	13.492	162	470538	9.885	ng/ul	99
45) 2-Nitroaniline	13.692	65	133572	11.400	ng/ul	97
47) Dimethylphthalate	14.068	163	589101	10.006	ng/ul	99
48) 2,6-Dinitrotoluene	14.186	165	92619	9.676	ng/ul	88
50) Acenaphthylene	14.333	152	735100	9.766	ng/ul	100
51) 3-Nitroaniline	14.515	138	108428	9.445	ng/ul	100
52) Acenaphthene	14.674	153	530588	10.074	ng/ul	100
53) 2,4-Dinitrophenol	14.715	184	44106	10.384	ng/ul	99
55) 4-Nitrophenol	14.810	109	89915	9.771	ng/ul	98
56) Dibenzofuran	15.004	168	716679	10.133	ng/ul	99
57) 2,4-Dinitrotoluene	14.968	165	142376	10.231	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.227	232	119455	9.332	ng/ul	97
59) Diethylphthalate	15.427	149	595845	10.110	ng/ul	99
61) Fluorene	15.651	166	601565	9.986	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.645	204	274882	9.541	ng/ul	98
63) 4-Nitroaniline	15.668	138	105957	9.499	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.727	198	73820	9.796	ng/ul#	97
67) N-Nitrosodiphenylamine	15.857	169	493365	9.925	ng/ul	99
68) 4-Bromophenyl-phenylether	16.539	248	146935	8.832	ng/ul	99
69) Hexachlorobenzene	16.651	284	171090	8.695	ng/ul	98
70) Atrazine	16.809	200	156515	8.922	ng/ul	99
71) Pentachlorophenol	16.992	266	87328	7.472	ng/ul	95
72) Phenanthrene	17.386	178	938212	10.092	ng/ul	100
74) Anthracene	17.480	178	931574	9.858	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.416	216	215974	9.356	ng/ul	99
76) Pentachlorobenzene	14.921	250	230036	9.213	ng/ul	98
77) Carbazole	17.745	167	863432	9.869	ng/ul	99
78) Di-n-butylphthalate	18.309	149	954864	9.629	ng/ul	99
80) Fluoranthene	19.398	202	1001679	9.725	ng/ul	99
82) Pyrene	19.756	202	1100758	10.096	ng/ul	99
83) Butylbenzylphthalate	20.650	149	405256	9.809	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.433	252	351789	10.537	ng/ul	98
85) Benzo(a)anthracene	21.497	228	1073348	10.026	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.427	149	613738	10.169	ng/ul	99
87) Chrysene	21.550	228	1025144	10.089	ng/ul	99
89) Di-n-octyl phthalate	22.362	149	1023906	9.755	ng/ul	100
90) Benzo(b)fluoranthene	23.191	252	1095289	9.921	ng/ul	100
91) Benzo(k)fluoranthene	23.244	252	1049067	9.510	ng/ul	99
93) Benzo(a)pyrene	23.827	252	948362	9.772	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.444	276	1150025	10.147	ng/ul	99
95) Dibenzo(a,h)anthracene	26.468	278	1042928	10.147	ng/ul	100
96) Benzo(g,h,i)perylene	27.221	276	1034456	10.528	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092922\
 Data File : BM036797.D
 Acq On : 29 Sep 2022 12:24
 Operator : CG/JU
 Sample : SST01079
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST010027

Quant Time: Sep 30 01:12:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 30 01:10:23 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.977 ng/uL
RT: 3.404 min Scan# 54
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion: 88 Resp: 32262
Ion Ratio Lower Upper
88 100
43 29.1 21.5 32.3
58 72.6 55.5 83.3



#5
Pyridine
Concen: 9.628 ng/uL
RT: 3.816 min Scan# 124
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion: 79 Resp: 212458
Ion Ratio Lower Upper
79 100
52 76.8 57.6 86.4
51 36.2 26.9 40.3









#6
Benzaldehyde
Concen: 10.973 ng/ul
RT: 7.122 min Scan# 686
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion: 77 Resp: 134055
Ion Ratio Lower Upper
77 100
105 95.6 77.2 115.8
106 92.4 74.7 112.1



#8
Phenol
Concen: 9.677 ng/ul
RT: 7.169 min Scan# 694
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion: 94 Resp: 262025
Ion Ratio Lower Upper
94 100
65 28.5 23.0 34.4
66 36.2 29.1 43.7









#10
Bis(2-Chloroethyl)ether
Concen: 10.462 ng/ul
RT: 7.404 min Scan# 734
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion: 93 Resp: 220422
Ion Ratio Lower Upper
93 100
63 78.9 62.7 94.1
95 32.8 25.4 38.2



#12
2-Chlorophenol
Concen: 10.042 ng/ul
RT: 7.545 min Scan# 758
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:128 Resp: 197306
Ion Ratio Lower Upper
128 100
64 46.7 39.1 58.7
130 33.3 25.6 38.4





#13

2-Methylphenol

Concen: 9.598 ng/ul

RT: 8.428 min Scan# 908

Delta R.T. 0.000 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24

Instrument :

BNA_M

ClientSampleId :





SSTD010027

Tgt Ion:108 Resp: 191033

Ion Ratio Lower Upper

108 100

107 91.0 69.2 103.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 12.143 ng/ul

RT: 8.516 min Scan# 923

Delta R.T. -0.006 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24




Tgt Ion: 45 Resp: 303213


Ion Ratio Lower Upper

45 100

77 15.4 12.4 18.6

79 12.1 9.3 13.9









#16
Acetophenone
Concen: 10.329 ng/ul
RT: 8.810 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:105 Resp: 333989
Ion Ratio Lower Upper
105 100
77 82.6 62.7 94.1
51 28.4 21.2 31.8



#17
N-Nitroso-di-n-propylamine
Concen: 11.196 ng/ul
RT: 8.792 min Scan# 970
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion: 70 Resp: 174681
Ion Ratio Lower Upper
70 100
42 53.2 44.2 66.2
101 9.1 7.3 10.9
130 17.9 15.0 22.6









#18
4-Methylphenol
Concen: 10.030 ng/ul
RT: 8.751 min Scan# 963
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:108 Resp: 215858
Ion Ratio Lower Upper
108 100
107 113.9 88.5 132.7



#19
Hexachloroethane
Concen: 10.311 ng/ul
RT: 9.069 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:117 Resp: 82062
Ion Ratio Lower Upper
117 100
201 88.1 68.3 102.5
199 55.9 43.8 65.6









#22
Nitrobenzene
Concen: 10.825 ng/ul
RT: 9.192 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion: 77 Resp: 250013
Ion Ratio Lower Upper
77 100
123 41.2 34.3 51.5
65 12.9 10.9 16.3



#23
Isophorone
Concen: 9.986 ng/ul
RT: 9.722 min Scan# 1128
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion: 82 Resp: 470276
Ion Ratio Lower Upper
82 100
95 6.4 5.0 7.4
138 16.0 12.9 19.3









#25
2-Nitrophenol
Concen: 10.249 ng/ul
RT: 9.904 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:139 Resp: 98631
Ion Ratio Lower Upper
139 100
65 56.4 46.0 69.0
109 32.7 24.2 36.2



#26
2,4-Dimethylphenol
Concen: 9.969 ng/ul
RT: 9.963 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:107 Resp: 236628
Ion Ratio Lower Upper
107 100
121 48.7 41.4 62.0
122 87.0 71.1 106.7





#27

Bis(2-Chloroethoxy)methane

Concen: 10.828 ng/ul

RT: 10.204 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24

Instrument :

BNA_M

ClientSampleId :

SSTD010027





Tgt Ion: 93 Resp: 303932

Ion Ratio Lower Upper

93 100

95 31.7 25.6 38.4

123 11.2 8.6 12.8



#29

2,4-Dichlorophenol

Concen: 9.519 ng/ul

RT: 10.433 min Scan# 1249

Delta R.T. -0.006 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24




Tgt Ion:162 Resp: 180710


Ion Ratio Lower Upper

162 100

164 66.1 51.8 77.6

98 39.2 30.9 46.3









#30
Naphthalene
Concen: 10.153 ng/ul
RT: 10.845 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:128 Resp: 710818
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 12.9 10.4 15.6



#32
4-Chloroaniline
Concen: 10.032 ng/ul
RT: 10.951 min Scan# 1337
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:127 Resp: 310953
Ion Ratio Lower Upper
127 100
129 31.5 25.7 38.5









#33
Hexachlorobutadiene
Concen: 9.178 ng/ul
RT: 11.127 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:225 Resp: 106561
Ion Ratio Lower Upper
225 100
223 58.1 50.0 75.0
227 61.1 50.4 75.6



#34
Caprolactam
Concen: 9.145 ng/ul
RT: 11.733 min Scan# 1470
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:113 Resp: 56315
Ion Ratio Lower Upper
113 100
55 172.6 138.6 208.0
56 126.0 105.9 158.9









#35
4-Chloro-3-methylphenol
Concen: 10.107 ng/ul
RT: 12.069 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:107 Resp: 211798
Ion Ratio Lower Upper
107 100
144 24.2 20.2 30.4
142 75.5 61.3 91.9



#36
2-Methylnaphthalene
Concen: 9.990 ng/ul
RT: 12.445 min Scan# 1591
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:142 Resp: 466005
Ion Ratio Lower Upper
142 100
141 89.1 71.4 107.2









#37
1-Methylnaphthalene
Concen: 10.006 ng/ul
RT: 12.663 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:142 Resp: 469221
Ion Ratio Lower Upper
142 100
141 89.9 73.4 110.2
116 4.1 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 9.272 ng/ul
RT: 12.810 min Scan# 1653
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:216 Resp: 215970
Ion Ratio Lower Upper
216 100
214 76.9 63.4 95.0
179 21.8 17.5 26.3
108 22.0 17.3 25.9









#40
Hexachlorocyclopentadiene
Concen: 9.133 ng/ul
RT: 12.792 min Scan# 1650
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:237 Resp: 110868
Ion Ratio Lower Upper
237 100
235 66.8 49.8 74.6
272 13.2 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 9.105 ng/ul
RT: 13.045 min Scan# 1693
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:196 Resp: 129855
Ion Ratio Lower Upper
196 100
198 94.1 77.0 115.4
200 28.6 24.6 37.0





#42

2,4,5-Trichlorophenol

Concen: 9.479 ng/ul

RT: 13.116 min Scan# 11

Delta R.T. 0.000 min

Lab File: BM036797.D


Acq: 29 Sep 2022 12:24

Instrument :

BNA_M

ClientSampleId :

SSTD010027






Tgt Ion:196 Resp: 144940

Ion Ratio Lower Upper

196 100

198 94.6 76.4 114.6

200 30.6 24.2 36.2



#43

1,1'-Biphenyl

Concen: 10.108 ng/ul

RT: 13.451 min Scan# 1762

Delta R.T. 0.000 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24




Tgt Ion:154 Resp: 613195


Ion Ratio Lower Upper

154 100

153 40.1 31.9 47.9

76 14.2 11.4 17.2









#44
2-Chloronaphthalene
Concen: 9.885 ng/ul
RT: 13.492 min Scan# 1769
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:162 Resp: 470538
Ion Ratio Lower Upper
162 100
164 32.0 25.7 38.5
127 40.1 32.5 48.7



#45
2-Nitroaniline
Concen: 11.400 ng/ul
RT: 13.692 min Scan# 1803
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion: 65 Resp: 133572
Ion Ratio Lower Upper
65 100
92 67.7 53.5 80.3
138 98.1 81.4 122.0









#47
Dimethylphthalate
Concen: 10.006 ng/ul
RT: 14.068 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:163 Resp: 589101
Ion Ratio Lower Upper
163 100
194 4.1 3.0 4.6
164 9.8 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 9.676 ng/ul
RT: 14.186 min Scan# 1887
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:165 Resp: 92619
Ion Ratio Lower Upper
165 100
89 77.8 53.0 79.4
121 26.1 19.0 28.4





#50

Acenaphthylene

Concen: 9.766 ng/ul

RT: 14.333 min Scan# 1912

Delta R.T. 0.000 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24

Instrument :

BNA_M

ClientSampleId :

SSTD010027





Tgt Ion:152 Resp: 735100

Ion Ratio Lower Upper

152 100

151 19.9 16.0 24.0

153 13.0 10.4 15.6



#51

3-Nitroaniline

Concen: 9.445 ng/ul

RT: 14.515 min Scan# 1943

Delta R.T. 0.000 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24




Tgt Ion:138 Resp: 108428


Ion Ratio Lower Upper

138 100

108 11.9 8.6 12.8

92 122.3 97.7 146.5









#52
Acenaphthene
Concen: 10.074 ng/ul
RT: 14.674 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:153 Resp: 530588
Ion Ratio Lower Upper
153 100
152 48.0 38.6 58.0
154 88.3 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 10.384 ng/ul
RT: 14.715 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:184 Resp: 44106
Ion Ratio Lower Upper
184 100
63 87.0 69.2 103.8
154 65.6 51.9 77.9









#55
4-Nitrophenol
Concen: 9.771 ng/ul
RT: 14.810 min Scan# 1993
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:109 Resp: 89915
Ion Ratio Lower Upper
109 100
139 130.7 104.7 157.1
65 146.5 113.8 170.6



#56
Dibenzofuran
Concen: 10.133 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:168 Resp: 716679
Ion Ratio Lower Upper
168 100
139 39.0 31.6 47.4









#57
2,4-Dinitrotoluene
Concen: 10.231 ng/ul
RT: 14.968 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:165 Resp: 142376
Ion Ratio Lower Upper
165 100
63 52.9 42.2 63.4
182 3.0 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 9.332 ng/ul
RT: 15.227 min Scan# 2064
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:232 Resp: 119455
Ion Ratio Lower Upper
232 100
131 55.5 46.2 69.2
130 2.6 2.2 3.4
166 32.6 27.8 41.8









#59
Diethylphthalate
Concen: 10.110 ng/ul
RT: 15.427 min Scan# 2098
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:149 Resp: 595845
Ion Ratio Lower Upper
149 100
177 21.5 16.9 25.3
150 11.7 10.0 15.0



#61
Fluorene
Concen: 9.986 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:166 Resp: 601565
Ion Ratio Lower Upper
166 100
165 97.5 77.7 116.5
167 13.3 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 9.541 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:204 Resp: 274882
Ion Ratio Lower Upper
204 100
206 33.4 25.8 38.6
141 67.0 52.4 78.6



#63
4-Nitroaniline
Concen: 9.499 ng/ul
RT: 15.668 min Scan# 2139
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:138 Resp: 105957
Ion Ratio Lower Upper
138 100
92 57.9 44.3 66.5
108 85.7 64.2 96.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 9.796 ng/ul
 RT: 15.727 min Scan# 2149
 Delta R.T. 0.000 min
 Lab File: BM036797.D
 Acq: 29 Sep 2022 12:24




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010027


Tgt Ion:198 Resp: 73820
 Ion Ratio Lower Upper
 198 100
 182 8.0 4.6 7.0#
 77 33.3 25.7 38.5



#67
 N-Nitrosodiphenylamine
 Concen: 9.925 ng/ul
 RT: 15.857 min Scan# 2171
 Delta R.T. -0.006 min
 Lab File: BM036797.D
 Acq: 29 Sep 2022 12:24

Tgt Ion:169 Resp: 493365
 Ion Ratio Lower Upper
 169 100
 168 68.8 54.3 81.5
 167 37.4 29.7 44.5









#68
4-Bromophenyl-phenylether
Concen: 8.832 ng/ul
RT: 16.539 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:248 Resp: 146935
Ion Ratio Lower Upper
248 100
250 96.4 77.1 115.7
141 86.0 69.9 104.9



#69
Hexachlorobenzene
Concen: 8.695 ng/ul
RT: 16.651 min Scan# 2306
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:284 Resp: 171090
Ion Ratio Lower Upper
284 100
142 42.2 32.5 48.7
249 30.2 23.8 35.8









#70
Atrazine
Concen: 8.922 ng/ul
RT: 16.809 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:200 Resp: 156515
Ion Ratio Lower Upper
200 100
173 27.0 22.5 33.7
215 48.0 38.5 57.7



#71
Pentachlorophenol
Concen: 7.472 ng/ul
RT: 16.992 min Scan# 2364
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:266 Resp: 87328
Ion Ratio Lower Upper
266 100
264 66.5 48.7 73.1
268 64.6 50.5 75.7









#72
Phenanthrene
Concen: 10.092 ng/ul
RT: 17.386 min Scan# 2432
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:178 Resp: 938212
Ion Ratio Lower Upper
178 100
179 15.4 12.1 18.1
176 19.6 15.5 23.3



#74
Anthracene
Concen: 9.858 ng/ul
RT: 17.480 min Scan# 2447
Delta R.T. 0.000 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:178 Resp: 931574
Ion Ratio Lower Upper
178 100
179 15.0 12.6 18.8
176 18.8 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 9.356 ng/uL

RT: 13.416 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24

Instrument :

BNA_M

ClientSampleId :

SSTD010027





Tgt Ion:216 Resp: 215974

Ion Ratio Lower Upper

216 100

214 76.9 62.6 93.8

218 47.9 38.2 57.4



#76

Pentachlorobenzene

Concen: 9.213 ng/uL

RT: 14.921 min Scan# 2012

Delta R.T. 0.000 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24




Tgt Ion:250 Resp: 230036


Ion Ratio Lower Upper

250 100

248 62.7 51.8 77.6

252 63.4 52.1 78.1





#77

Carbazole

Concen: 9.869 ng/ul

RT: 17.745 min Scan# 24

Delta R.T. -0.006 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24

Instrument :

BNA_M

ClientSampleId :

SSTD010027





Tgt Ion:167 Resp: 863432

Ion Ratio Lower Upper

167 100

166 21.3 17.4 26.0

139 13.8 10.8 16.2



#78

Di-n-butylphthalate

Concen: 9.629 ng/ul

RT: 18.309 min Scan# 2588

Delta R.T. -0.006 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24




Tgt Ion:149 Resp: 954864


Ion Ratio Lower Upper

149 100

150 8.9 7.3 10.9

104 5.9 4.5 6.7





#80

Fluoranthene

Concen: 9.725 ng/ul

RT: 19.398 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24

Instrument :

BNA_M

ClientSampleId :

SSTD010027





Tgt Ion:202 Resp: 1001679

Ion Ratio Lower Upper

202 100

101 12.7 10.5 15.7

100 9.6 8.2 12.2



#82

Pyrene

Concen: 10.096 ng/ul

RT: 19.756 min Scan# 2834

Delta R.T. 0.000 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24




Tgt Ion:202 Resp: 1100758


Ion Ratio Lower Upper

202 100

101 15.7 12.6 19.0

100 12.2 10.5 15.7









#83
 Butylbenzylphthalate
 Concen: 9.809 ng/ul
 RT: 20.650 min Scan# 2986
 Delta R.T. -0.006 min
 Lab File: BM036797.D
 Acq: 29 Sep 2022 12:24




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010027


Tgt Ion:149 Resp: 405256
 Ion Ratio Lower Upper
 149 100
 91 83.0 61.8 92.8
 206 17.7 15.0 22.6



#84
 3,3'-Dichlorobenzidine
 Concen: 10.537 ng/ul
 RT: 21.433 min Scan# 3119
 Delta R.T. 0.000 min
 Lab File: BM036797.D
 Acq: 29 Sep 2022 12:24

Tgt Ion:252 Resp: 351789
 Ion Ratio Lower Upper
 252 100
 254 63.2 51.8 77.8
 126 13.9 11.5 17.3





#85

Benzo(a)anthracene

Concen: 10.026 ng/ul

RT: 21.497 min Scan# 3131

Delta R.T. -0.006 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24

Instrument :

BNA_M

ClientSampleId :

SSTD010027





Tgt Ion:228 Resp: 1073348

Ion Ratio Lower Upper

228 100

229 19.1 15.8 23.8

226 27.4 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 10.169 ng/ul

RT: 21.427 min Scan# 3118

Delta R.T. 0.000 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24




Tgt Ion:149 Resp: 613738


Ion Ratio Lower Upper

149 100

167 26.4 21.4 32.2

279 3.3 2.5 3.7





#87

Chrysene

Concen: 10.089 ng/ul

RT: 21.550 min Scan# 3140

Delta R.T. -0.006 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24

Instrument :

BNA_M

ClientSampleId :

SSTD010027





Tgt Ion:228 Resp: 1025144

Ion Ratio Lower Upper

228 100

226 29.9 24.3 36.5

229 19.4 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 9.755 ng/ul




RT: 22.362 min Scan# 3277


Delta R.T. 0.000 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24

Tgt Ion:149 Resp: 1023906









#90
Benzo(b)fluoranthene
Concen: 9.921 ng/ul
RT: 23.191 min Scan# 3419
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24




Instrument :
BNA_M
ClientSampleId :
SSTD010027


Tgt Ion:252 Resp: 1095289
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 12.7 9.9 14.9



#91
Benzo(k)fluoranthene
Concen: 9.510 ng/ul
RT: 23.244 min Scan# 3427
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:252 Resp: 1049067
Ion Ratio Lower Upper
252 100
253 21.9 17.9 26.9
125 12.3 9.7 14.5





#93

Benzo(a)pyrene

Concen: 9.772 ng/ul

RT: 23.827 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24

Instrument :

BNA_M

ClientSampleId :

SSTD010027





Tgt Ion:252 Resp: 948362

Ion Ratio Lower Upper

252 100

253 21.8 17.3 25.9

125 14.1 10.5 15.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 10.147 ng/ul

RT: 26.444 min Scan# 3971

Delta R.T. -0.012 min

Lab File: BM036797.D

Acq: 29 Sep 2022 12:24




Tgt Ion:276 Resp: 1150025


Ion Ratio Lower Upper

276 100

138 29.6 24.0 36.0

277 24.2 19.7 29.5









#95
Dibenzo(a,h)anthracene
Concen: 10.147 ng/ul
RT: 26.468 min Scan# 3976
Delta R.T. -0.006 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Instrument :
BNA_M
ClientSampleId :
SSTD010027

Tgt Ion:278 Resp: 1042928
Ion Ratio Lower Upper
278 100
139 20.3 16.1 24.1
279 24.2 19.4 29.2



#96
Benzo(g,h,i)perylene
Concen: 10.528 ng/ul
RT: 27.221 min Scan# 4103
Delta R.T. -0.012 min
Lab File: BM036797.D
Acq: 29 Sep 2022 12:24

Tgt Ion:276 Resp: 1034456
Ion Ratio Lower Upper
276 100
138 26.7 21.0 31.4
277 23.5 19.1 28.7

