

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092922\
 Data File : BM036798.D
 Acq On : 29 Sep 2022 13:00
 Operator : CG/JU
 Sample : SST02080
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST020028

Quant Time: Sep 30 01:12:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 30 01:10:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.986	152	284080	20.000	ng/ul	0.00
20) Naphthalene-d8	10.792	136	1318069	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.610	164	836303	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.345	188	1715651	20.000	ng/ul	0.00
79) Chrysene-d12	21.515	240	1799776	20.000	ng/ul	0.00
88) Perylene-d12	23.938	264	1842901	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	66058	8.717	ng/uL	0.00
4) Pyridine-d5	3.798	84	501240	22.680	ng/ul	0.00
7) Phenol-d5	7.139	99	578054	22.770	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.310	67	383613	24.882	ng/ul	0.00
11) 2-Chlorophenol-d4	7.510	132	422179	23.099	ng/ul	0.00
15) 4-Methylphenol-d8	8.692	113	460238	24.535	ng/ul	0.00
21) Nitrobenzene-d5	9.151	128	222598	24.951	ng/ul	0.00
24) 2-Nitrophenol-d4	9.875	143	214510	26.844	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.410	165	423230	22.854	ng/ul	0.00
31) 4-Chloroaniline-d4	10.927	131	697088	22.640	ng/ul	0.00
46) Dimethylphthalate-d6	14.021	166	1306624	23.165	ng/ul	0.00
49) Acenaphthylene-d8	14.304	160	1497227	22.379	ng/ul	0.00
54) 4-Nitrophenol-d4	14.804	143	268907	24.580	ng/ul	0.00
60) Fluorene-d10	15.598	176	1130690	22.369	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.715	200	182998	26.960	ng/ul	0.00
73) Anthracene-d10	17.445	188	1767196	22.528	ng/ul	0.00
81) Pyrene-d10	19.733	212	1976169	22.319	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.780	264	2043476	22.696	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.404	88	73596	9.016	ng/uL	100
5) Pyridine	3.816	79	507632	22.862	ng/ul	100
6) Benzaldehyde	7.122	77	332108	27.015	ng/ul	100
8) Phenol	7.169	94	629423	23.101	ng/ul	100
10) Bis(2-Chloroethyl)ether	7.404	93	510701	24.090	ng/ul	100
12) 2-Chlorophenol	7.545	128	467992	23.670	ng/ul	100
13) 2-Methylphenol	8.428	108	469898	23.463	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.522	45	688798	27.414	ng/ul	100
16) Acetophenone	8.816	105	806794	24.796	ng/ul	100
17) N-Nitroso-di-n-propyla...	8.798	70	427751	27.246	ng/ul	100
18) 4-Methylphenol	8.757	108	531728	24.553	ng/ul	100
19) Hexachloroethane	9.069	117	186835	23.330	ng/ul	100
22) Nitrobenzene	9.192	77	598753	25.278	ng/ul	100
23) Isophorone	9.722	82	1173898	24.306	ng/ul	100
25) 2-Nitrophenol	9.904	139	254777	25.813	ng/ul	100
26) 2,4-Dimethylphenol	9.963	107	572636	23.523	ng/ul	100
27) Bis(2-Chloroethoxy)met...	10.204	93	708771	24.622	ng/ul	100
29) 2,4-Dichlorophenol	10.439	162	441439	22.674	ng/ul	100
30) Naphthalene	10.845	128	1633586	22.751	ng/ul	100
32) 4-Chloroaniline	10.957	127	727214	22.876	ng/ul	100
33) Hexachlorobutadiene	11.127	225	242993	20.407	ng/ul	100
34) Caprolactam	11.739	113	150159	23.777	ng/ul	100
35) 4-Chloro-3-methylphenol	12.069	107	524434	24.402	ng/ul	100
36) 2-Methylnaphthalene	12.445	142	1091985	22.826	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092922\
 Data File : BM036798.D
 Acq On : 29 Sep 2022 13:00
 Operator : CG/JU
 Sample : SSTD02080
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020028

Quant Time: Sep 30 01:12:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 30 01:10:23 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.663	142	1108267	23.044	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.810	216	507562	21.443	ng/ul	100
40) Hexachlorocyclopentadiene	12.792	237	265917	21.556	ng/ul	100
41) 2,4,6-Trichlorophenol	13.051	196	328204	22.645	ng/ul	100
42) 2,4,5-Trichlorophenol	13.116	196	361249	23.248	ng/ul	100
43) 1,1'-Biphenyl	13.451	154	1420192	23.037	ng/ul	100
44) 2-Chloronaphthalene	13.492	162	1101771	22.775	ng/ul	100
45) 2-Nitroaniline	13.698	65	350880	29.468	ng/ul	100
47) Dimethylphthalate	14.068	163	1389098	23.217	ng/ul	100
48) 2,6-Dinitrotoluene	14.192	165	254990	26.213	ng/ul	100
50) Acenaphthylene	14.333	152	1747976	22.852	ng/ul	100
51) 3-Nitroaniline	14.515	138	296979	25.455	ng/ul	100
52) Acenaphthene	14.674	153	1226194	22.910	ng/ul	100
53) 2,4-Dinitrophenol	14.715	184	128940	29.873	ng/ul	100
55) 4-Nitrophenol	14.815	109	229374	24.527	ng/ul	100
56) Dibenzofuran	15.004	168	1638902	22.801	ng/ul	100
57) 2,4-Dinitrotoluene	14.968	165	377929	26.724	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.227	232	295006	22.680	ng/ul	100
59) Diethylphthalate	15.427	149	1397591	23.334	ng/ul	100
61) Fluorene	15.651	166	1402736	22.914	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.645	204	637494	21.775	ng/ul	100
63) 4-Nitroaniline	15.674	138	297537	26.249	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.727	198	201221	26.270	ng/ul	100
67) N-Nitrosodiphenylamine	15.862	169	1157211	22.901	ng/ul	100
68) 4-Bromophenyl-phenylether	16.539	248	351198	20.768	ng/ul	100
69) Hexachlorobenzene	16.651	284	403998	20.199	ng/ul	100
70) Atrazine	16.809	200	388132	21.767	ng/ul	100
71) Pentachlorophenol	16.998	266	238988	20.116	ng/ul	100
72) Phenanthrene	17.392	178	2187964	23.153	ng/ul	100
74) Anthracene	17.480	178	2189674	22.796	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.415	216	503455	21.455	ng/ul	100
76) Pentachlorobenzene	14.921	250	527383	20.780	ng/ul	100
77) Carbazole	17.750	167	2066028	23.232	ng/ul	100
78) Di-n-butylphthalate	18.315	149	2384823	23.660	ng/ul	100
80) Fluoranthene	19.397	202	2442715	22.514	ng/ul	100
82) Pyrene	19.756	202	2630495	22.904	ng/ul	100
83) Butylbenzylphthalate	20.656	149	1074505	24.690	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.433	252	916855	26.072	ng/ul	100
85) Benzo(a)anthracene	21.503	228	2568789	22.780	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.427	149	1677370	26.385	ng/ul	100
87) Chrysene	21.556	228	2429221	22.695	ng/ul	100
89) Di-n-octyl phthalate	22.362	149	2812355	25.313	ng/ul	100
90) Benzo(b)fluoranthene	23.197	252	2603877	22.281	ng/ul	100
91) Benzo(k)fluoranthene	23.250	252	2639146	22.600	ng/ul	100
93) Benzo(a)pyrene	23.832	252	2363383	23.004	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.456	276	2914753	24.295	ng/ul	100
95) Dibenzo(a,h)anthracene	26.473	278	2627531	24.149	ng/ul	100
96) Benzo(g,h,i)perylene	27.232	276	2580486	24.811	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092922\
 Data File : BM036798.D
 Acq On : 29 Sep 2022 13:00
 Operator : CG/JU
 Sample : SSTD02080
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020028

Quant Time: Sep 30 01:12:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 30 01:10:23 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.016 ng/uL
RT: 3.404 min Scan# 54
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion	Ratio	Lower	Upper
88	100		
43	26.9	21.5	32.3
58	69.4	55.5	83.3



#5
Pyridine
Concen: 22.862 ng/uL
RT: 3.816 min Scan# 124
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion	Ratio	Lower	Upper
79	100		
52	72.0	57.6	86.4
51	33.6	26.9	40.3









#6
Benzaldehyde
Concen: 27.015 ng/ul
RT: 7.122 min Scan# 686
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion: 77 Resp: 332108
Ion Ratio Lower Upper
77 100
105 96.5 77.2 115.8
106 93.4 74.7 112.1



#8
Phenol
Concen: 23.101 ng/ul
RT: 7.169 min Scan# 694
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion: 94 Resp: 629423
Ion Ratio Lower Upper
94 100
65 28.7 23.0 34.4
66 36.4 29.1 43.7









#10
Bis(2-Chloroethyl)ether
Concen: 24.090 ng/ul
RT: 7.404 min Scan# 734
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion: 93 Resp: 510701
Ion Ratio Lower Upper
93 100
63 78.4 62.7 94.1
95 31.8 25.4 38.2



#12
2-Chlorophenol
Concen: 23.670 ng/ul
RT: 7.545 min Scan# 758
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:128 Resp: 467992
Ion Ratio Lower Upper
128 100
64 48.9 39.1 58.7
130 32.0 25.6 38.4





#13

2-Methylphenol

Concen: 23.463 ng/ul

RT: 8.428 min Scan# 908

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00

Instrument :

BNA_M

ClientSampleId :





SSTD020028

Tgt Ion:108 Resp: 469898

Ion Ratio Lower Upper

108 100

107 86.5 69.2 103.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 27.414 ng/ul

RT: 8.522 min Scan# 924

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00




Tgt Ion: 45 Resp: 688798


Ion Ratio Lower Upper

45 100

77 15.5 12.4 18.6

79 11.6 9.3 13.9









#16
Acetophenone
Concen: 24.796 ng/ul
RT: 8.816 min Scan# 974
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:105 Resp: 806794
Ion Ratio Lower Upper
105 100
77 78.4 62.7 94.1
51 26.5 21.2 31.8



#17
N-Nitroso-di-n-propylamine
Concen: 27.246 ng/ul
RT: 8.798 min Scan# 971
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion: 70 Resp: 427751
Ion Ratio Lower Upper
70 100
42 55.2 44.2 66.2
101 9.1 7.3 10.9
130 18.8 15.0 22.6









#18
4-Methylphenol
Concen: 24.553 ng/ul
RT: 8.757 min Scan# 964
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:108 Resp: 531728
Ion Ratio Lower Upper
108 100
107 110.6 88.5 132.7



#19
Hexachloroethane
Concen: 23.330 ng/ul
RT: 9.069 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:117 Resp: 186835
Ion Ratio Lower Upper
117 100
201 85.4 68.3 102.5
199 54.7 43.8 65.6









#22
Nitrobenzene
Concen: 25.278 ng/ul
RT: 9.192 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion: 77 Resp: 598753
Ion Ratio Lower Upper
77 100
123 42.9 34.3 51.5
65 13.6 10.9 16.3



#23
Isophorone
Concen: 24.306 ng/ul
RT: 9.722 min Scan# 1128
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion: 82 Resp: 1173898
Ion Ratio Lower Upper
82 100
95 6.2 5.0 7.4
138 16.1 12.9 19.3









#25
2-Nitrophenol
Concen: 25.813 ng/ul
RT: 9.904 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:139 Resp: 254777
Ion Ratio Lower Upper
139 100
65 57.5 46.0 69.0
109 30.2 24.2 36.2



#26
2,4-Dimethylphenol
Concen: 23.523 ng/ul
RT: 9.963 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:107 Resp: 572636
Ion Ratio Lower Upper
107 100
121 51.7 41.4 62.0
122 88.9 71.1 106.7





#27

Bis(2-Chloroethoxy)methane

Concen: 24.622 ng/ul

RT: 10.204 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00

Instrument :

BNA_M

ClientSampleId :

SSTD020028





Tgt Ion: 93 Resp: 708771

Ion Ratio Lower Upper

93 100

95 32.0 25.6 38.4

123 10.7 8.6 12.8



#29

2,4-Dichlorophenol

Concen: 22.674 ng/ul

RT: 10.439 min Scan# 1250

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00




Tgt Ion:162 Resp: 441439


Ion Ratio Lower Upper

162 100

164 64.7 51.8 77.6

98 38.6 30.9 46.3









#30
Naphthalene
Concen: 22.751 ng/ul
RT: 10.845 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:128 Resp: 1633586
Ion Ratio Lower Upper
128 100
129 10.7 8.6 12.8
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 22.876 ng/ul
RT: 10.957 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:127 Resp: 727214
Ion Ratio Lower Upper
127 100
129 32.1 25.7 38.5









#33
Hexachlorobutadiene
Concen: 20.407 ng/ul
RT: 11.127 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:225 Resp: 242993
Ion Ratio Lower Upper
225 100
223 62.5 50.0 75.0
227 63.0 50.4 75.6



#34
Caprolactam
Concen: 23.777 ng/ul
RT: 11.739 min Scan# 1471
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:113 Resp: 150159
Ion Ratio Lower Upper
113 100
55 173.3 138.6 208.0
56 132.4 105.9 158.9









#35
4-Chloro-3-methylphenol
Concen: 24.402 ng/ul
RT: 12.069 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:107 Resp: 524434
Ion Ratio Lower Upper
107 100
144 25.3 20.2 30.4
142 76.6 61.3 91.9



#36
2-Methylnaphthalene
Concen: 22.826 ng/ul
RT: 12.445 min Scan# 1591
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:142 Resp: 1091985
Ion Ratio Lower Upper
142 100
141 89.3 71.4 107.2









#37
1-Methylnaphthalene
Concen: 23.044 ng/ul
RT: 12.663 min Scan# 1628
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:142 Resp: 1108267
Ion Ratio Lower Upper
142 100
141 91.8 73.4 110.2
116 3.9 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.443 ng/ul
RT: 12.810 min Scan# 1653
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:216 Resp: 507562
Ion Ratio Lower Upper
216 100
214 79.2 63.4 95.0
179 21.9 17.5 26.3
108 21.6 17.3 25.9









#40
Hexachlorocyclopentadiene
Concen: 21.556 ng/ul
RT: 12.792 min Scan# 1650
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:237 Resp: 265917
Ion Ratio Lower Upper
237 100
235 62.2 49.8 74.6
272 12.9 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 22.645 ng/ul
RT: 13.051 min Scan# 1694
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:196 Resp: 328204
Ion Ratio Lower Upper
196 100
198 96.2 77.0 115.4
200 30.8 24.6 37.0





#42

2,4,5-Trichlorophenol

Concen: 23.248 ng/ul

RT: 13.116 min Scan# 11

Delta R.T. 0.000 min

Lab File: BM036798.D


Acq: 29 Sep 2022 13:00

Instrument :

BNA_M

ClientSampleId :

SSTD020028






Tgt Ion:196 Resp: 361249

Ion Ratio Lower Upper

196 100

198 95.5 76.4 114.6

200 30.2 24.2 36.2



#43

1,1'-Biphenyl

Concen: 23.037 ng/ul

RT: 13.451 min Scan# 1762

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00




Tgt Ion:154 Resp: 1420192


Ion Ratio Lower Upper

154 100

153 39.9 31.9 47.9

76 14.3 11.4 17.2









#44
2-Chloronaphthalene
Concen: 22.775 ng/ul
RT: 13.492 min Scan# 1769
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:162 Resp: 1101771
Ion Ratio Lower Upper
162 100
164 32.1 25.7 38.5
127 40.6 32.5 48.7



#45
2-Nitroaniline
Concen: 29.468 ng/ul
RT: 13.698 min Scan# 1804
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion: 65 Resp: 350880
Ion Ratio Lower Upper
65 100
92 66.9 53.5 80.3
138 101.7 81.4 122.0









#47
Dimethylphthalate
Concen: 23.217 ng/ul
RT: 14.068 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:163 Resp: 1389098
Ion Ratio Lower Upper
163 100
194 3.8 3.0 4.6
164 10.3 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 26.213 ng/ul
RT: 14.192 min Scan# 1888
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:165 Resp: 254990
Ion Ratio Lower Upper
165 100
89 66.2 53.0 79.4
121 23.7 19.0 28.4









#50
Acenaphthylene
Concen: 22.852 ng/ul
RT: 14.333 min Scan# 1912
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:152 Resp: 1747976
Ion Ratio Lower Upper
152 100
151 20.0 16.0 24.0
153 13.0 10.4 15.6



#51
3-Nitroaniline
Concen: 25.455 ng/ul
RT: 14.515 min Scan# 1943
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:138 Resp: 296979
Ion Ratio Lower Upper
138 100
108 10.7 8.6 12.8
92 122.1 97.7 146.5









#52
Acenaphthene
Concen: 22.910 ng/ul
RT: 14.674 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:153 Resp: 1226194
Ion Ratio Lower Upper
153 100
152 48.3 38.6 58.0
154 88.2 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 29.873 ng/ul
RT: 14.715 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:184 Resp: 128940
Ion Ratio Lower Upper
184 100
63 86.5 69.2 103.8
154 64.9 51.9 77.9









#55
4-Nitrophenol
Concen: 24.527 ng/ul
RT: 14.815 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:109 Resp: 229374
Ion Ratio Lower Upper
109 100
139 130.9 104.7 157.1
65 142.2 113.8 170.6



#56
Dibenzofuran
Concen: 22.801 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:168 Resp: 1638902
Ion Ratio Lower Upper
168 100
139 39.5 31.6 47.4









#57
2,4-Dinitrotoluene
Concen: 26.724 ng/ul
RT: 14.968 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:165 Resp: 377929
Ion Ratio Lower Upper
165 100
63 52.8 42.2 63.4
182 2.8 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.680 ng/ul
RT: 15.227 min Scan# 2064
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:232 Resp: 295006
Ion Ratio Lower Upper
232 100
131 57.7 46.2 69.2
130 2.8 2.2 3.4
166 34.8 27.8 41.8









#59
Diethylphthalate
Concen: 23.334 ng/ul
RT: 15.427 min Scan# 2098
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:149 Resp: 1397591
Ion Ratio Lower Upper
149 100
177 21.1 16.9 25.3
150 12.5 10.0 15.0



#61
Fluorene
Concen: 22.914 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:166 Resp: 1402736
Ion Ratio Lower Upper
166 100
165 97.1 77.7 116.5
167 13.3 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 21.775 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:204 Resp: 637494
Ion Ratio Lower Upper
204 100
206 32.2 25.8 38.6
141 65.5 52.4 78.6



#63
4-Nitroaniline
Concen: 26.249 ng/ul
RT: 15.674 min Scan# 2140
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:138 Resp: 297537
Ion Ratio Lower Upper
138 100
92 55.4 44.3 66.5
108 80.3 64.2 96.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 26.270 ng/ul
 RT: 15.727 min Scan# 2149
 Delta R.T. 0.000 min
 Lab File: BM036798.D
 Acq: 29 Sep 2022 13:00




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020028


Tgt Ion:198 Resp: 201221
 Ion Ratio Lower Upper
 198 100
 182 5.8 4.6 7.0
 77 32.1 25.7 38.5



#67
 N-Nitrosodiphenylamine
 Concen: 22.901 ng/ul
 RT: 15.862 min Scan# 2172
 Delta R.T. 0.000 min
 Lab File: BM036798.D
 Acq: 29 Sep 2022 13:00

Tgt Ion:169 Resp: 1157211
 Ion Ratio Lower Upper
 169 100
 168 67.9 54.3 81.5
 167 37.1 29.7 44.5









#68
4-Bromophenyl-phenylether
Concen: 20.768 ng/ul
RT: 16.539 min Scan# 211
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:248 Resp: 351198
Ion Ratio Lower Upper
248 100
250 96.4 77.1 115.7
141 87.4 69.9 104.9



#69
Hexachlorobenzene
Concen: 20.199 ng/ul
RT: 16.651 min Scan# 2306
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:284 Resp: 403998
Ion Ratio Lower Upper
284 100
142 40.6 32.5 48.7
249 29.8 23.8 35.8





#70

Atrazine

Concen: 21.767 ng/ul

RT: 16.809 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00

Instrument :

BNA_M

ClientSampleId :

SSTD020028





Tgt Ion:200 Resp: 388132

Ion Ratio Lower Upper

200 100

173 28.1 22.5 33.7

215 48.1 38.5 57.7



#71

Pentachlorophenol

Concen: 20.116 ng/ul

RT: 16.998 min Scan# 2365

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00




Tgt Ion:266 Resp: 238988


Ion Ratio Lower Upper

266 100

264 60.9 48.7 73.1

268 63.1 50.5 75.7









#72
Phenanthrene
Concen: 23.153 ng/ul
RT: 17.392 min Scan# 2432
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:178 Resp: 2187964
Ion Ratio Lower Upper
178 100
179 15.1 12.1 18.1
176 19.4 15.5 23.3



#74
Anthracene
Concen: 22.796 ng/ul
RT: 17.480 min Scan# 2447
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:178 Resp: 2189674
Ion Ratio Lower Upper
178 100
179 15.7 12.6 18.8
176 18.9 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.455 ng/uL

RT: 13.415 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00

Instrument :

BNA_M

ClientSampleId :

SSTD020028





Tgt Ion:216 Resp: 503455

Ion Ratio Lower Upper

216 100

214 78.2 62.6 93.8

218 47.8 38.2 57.4



#76

Pentachlorobenzene

Concen: 20.780 ng/uL

RT: 14.921 min Scan# 2012

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00




Tgt Ion:250 Resp: 527383


Ion Ratio Lower Upper

250 100

248 64.7 51.8 77.6

252 65.1 52.1 78.1









#77
 Carbazole
 Concen: 23.232 ng/ul
 RT: 17.750 min Scan# 2493
 Delta R.T. 0.000 min
 Lab File: BM036798.D
 Acq: 29 Sep 2022 13:00




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020028


Tgt Ion:167 Resp: 2066028
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.4 26.0
 139 13.5 10.8 16.2



#78
 Di-n-butylphthalate
 Concen: 23.660 ng/ul
 RT: 18.315 min Scan# 2589
 Delta R.T. 0.000 min
 Lab File: BM036798.D
 Acq: 29 Sep 2022 13:00

Tgt Ion:149 Resp: 2384823
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.3 10.9
 104 5.6 4.5 6.7









#80
Fluoranthene
Concen: 22.514 ng/ul
RT: 19.397 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:202 Resp: 2442715
Ion Ratio Lower Upper
202 100
101 13.1 10.5 15.7
100 10.2 8.2 12.2



#82
Pyrene
Concen: 22.904 ng/ul
RT: 19.756 min Scan# 2834
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:202 Resp: 2630495
Ion Ratio Lower Upper
202 100
101 15.8 12.6 19.0
100 13.1 10.5 15.7





#83

Butylbenzylphthalate

Concen: 24.690 ng/ul

RT: 20.656 min Scan# 2987

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00

Instrument :

BNA_M

ClientSampleId :

SSTD020028





Tgt Ion:149 Resp: 1074505

Ion Ratio Lower Upper

149 100

91 77.3 61.8 92.8

206 18.8 15.0 22.6



#84

3,3'-Dichlorobenzidine

Concen: 26.072 ng/ul

RT: 21.433 min Scan# 3119

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00




Tgt Ion:252 Resp: 916855


Ion Ratio Lower Upper

252 100

254 64.8 51.8 77.8

126 14.4 11.5 17.3





#85

Benzo(a)anthracene

Concen: 22.780 ng/ul

RT: 21.503 min Scan# 3131

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00

Instrument :

BNA_M

ClientSampleId :

SSTD020028





Tgt Ion:228 Resp: 2568789

Ion Ratio Lower Upper

228 100

229 19.8 15.8 23.8

226 27.6 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 26.385 ng/ul

RT: 21.427 min Scan# 3118

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00




Tgt Ion:149 Resp: 1677370


Ion Ratio Lower Upper

149 100

167 26.8 21.4 32.2

279 3.1 2.5 3.7





#87

Chrysene

Concen: 22.695 ng/ul

RT: 21.556 min Scan# 3140

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00

Instrument :

BNA_M

ClientSampleId :

SSTD020028





Tgt Ion:228 Resp: 2429221

Ion Ratio Lower Upper

228 100

226 30.4 24.3 36.5

229 19.7 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 25.313 ng/ul




RT: 22.362 min Scan# 3277


Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00

Tgt Ion:149 Resp: 2812355









#90
Benzo(b)fluoranthene
Concen: 22.281 ng/ul
RT: 23.197 min Scan# 3419
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00




Instrument :
BNA_M
ClientSampleId :
SSTD020028


Tgt Ion:252 Resp: 2603877
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 12.4 9.9 14.9



#91
Benzo(k)fluoranthene
Concen: 22.600 ng/ul
RT: 23.250 min Scan# 3428
Delta R.T. 0.000 min
Lab File: BM036798.D
Acq: 29 Sep 2022 13:00

Tgt Ion:252 Resp: 2639146
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 12.1 9.7 14.5





#93

Benzo(a)pyrene

Concen: 23.004 ng/ul

RT: 23.832 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00

Instrument :

BNA_M

ClientSampleId :

SSTD020028





Tgt Ion:252 Resp: 2363383

Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9

125 13.1 10.5 15.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 24.295 ng/ul

RT: 26.456 min Scan# 3973

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00




Tgt Ion:276 Resp: 2914753


Ion Ratio Lower Upper

276 100

138 30.0 24.0 36.0

277 24.6 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 24.149 ng/ul

RT: 26.473 min Scan# 3976

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00

Instrument :

BNA_M

ClientSampleId :

SSTD020028





Tgt Ion:278 Resp: 2627531

Ion Ratio Lower Upper

278 100

139 20.1 16.1 24.1

279 24.3 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 24.811 ng/ul

RT: 27.232 min Scan# 4105

Delta R.T. 0.000 min

Lab File: BM036798.D

Acq: 29 Sep 2022 13:00

Tgt Ion:276 Resp: 2580486

Ion Ratio Lower Upper

276 100

138 26.2 21.0 31.4

277 23.9 19.1 28.7

