

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM093019\
 Data File : BM022835.D
 Acq On : 30 Sep 2019 21:08
 Operator : JU
 Sample : K5006-05
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ETJW1

Quant Time: Sep 30 23:38:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM092719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 27 18:03:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	257938	20.00	ng/ul	0.00
18) Naphthalene-d8	10.60	136	1177476	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.44	164	785750	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.18	188	1662927	20.00	ng/ul	0.00
78) Chrysene-d12	21.36	240	1601186	20.00	ng/ul	0.00
86) Perylene-d12	23.63	264	1430667	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.27	96	28728	4.82	ng/uL	0.00
5) Phenol-d5	6.97	99	137932	6.05	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.14	67	304154	24.38	ng/ul	0.00
9) 2-Chlorophenol-d4	7.34	132	375480	20.46	ng/ul	0.00
13) 4-Methylphenol-d8	8.51	113	252632	13.54	ng/ul	0.00
19) Nitrobenzene-d5	8.96	128	217707	23.96	ng/ul	0.00
22) 2-Nitrophenol-d4	9.68	143	239850	24.17	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.22	165	443078	22.43	ng/ul	-0.01
29) 4-Chloroaniline-d4	10.73	131	149694	6.24	ng/ul	0.00
44) Dimethylphthalate-d6	13.85	166	1569916	25.74	ng/ul	0.00
47) Acenaphthylene-d8	14.13	160	1743433	24.76	ng/ul	0.00
52) 4-Nitrophenol-d4	14.63	143	68117	5.65	ng/ul	-0.01
58) Fluorene-d10	15.43	176	1382619	26.60	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.54	200	253125	23.39	ng/ul	-0.01
71) Anthracene-d10	17.28	188	2438855	30.03	ng/ul	0.00
79) Pyrene-d10	19.57	212	2844421	33.01	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.49	264	2343504	31.66	ng/ul	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM093019\
Data File : BM022835.D
Acq On : 30 Sep 2019 21:08
Operator : JU
Sample : K5006-05
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ETJW1

Quant Time: Sep 30 23:38:58 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM092719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Sep 27 18:03:27 2019
Response via : Initial Calibration

