

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM093021\
 Data File : BM032293.D
 Acq On : 30 Sep 2021 11:43
 Operator : CG/JU
 Sample : PB139117BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK117

Quant Time: Sep 30 12:18:03 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 30 10:17:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.650	152	6618	0.400	ng/ul	0.00
4) Naphthalene-d8	10.442	136	19978	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.296	164	8617	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.022	188	16124	0.400	ng/ul	0.00
17) Chrysene-d12	21.186	240	11521	0.400	ng/ul	0.00
23) Perylene-d12	23.348	264	9520	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.986	96	5199	0.731	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.035	152	9265	0.348	ng/ul	0.00
18) Fluoranthene-d10	19.042	212	13244	0.423	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM093021\
Data File : BM032293.D
Acq On : 30 Sep 2021 11:43
Operator : CG/JU
Sample : PB139117BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK117

Quant Time: Sep 30 12:18:03 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 30 10:17:25 2021
Response via : Initial Calibration

