



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM093021\
 Data File : BM032289.D
 Acq On : 30 Sep 2021 08:24
 Operator : CG/JU
 Sample : DFTPP017
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP017

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Sep 30 10:17:25 2021



AutoFind: Scans 2488, 2489, 2490; Background Corrected with Scan 2481

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.2	270612	PASS
68	69	0.00	2	1.6	4240	PASS
69	198	0.00	100	35.8	259921	PASS
70	69	0.00	2	0.6	1548	PASS
127	198	10	80	46.0	334528	PASS
197	198	0.00	2	0.4	2557	PASS
198	198	100	100	100.0	726656	PASS
199	198	5	9	6.8	49571	PASS
275	198	10	60	25.1	182571	PASS
365	198	1	100	3.1	22501	PASS
441	198	0.01	100	11.9	86205	PASS
442	198	50	100	77.0	559829	PASS
443	442	15	24	19.5	109085	PASS

m/z	Abundance
36.00	185.0
39.00	346.0
40.00	1823.0
41.00	595.0
42.00	158.0
43.00	570.0
44.00	7708.0
45.00	263.0
46.90	191.0
49.00	705.0
51.00	492.0
52.10	158.0
55.00	253.0
67.00	187.0
69.10	196.0
73.00	1764.0
75.10	197.0
76.00	166.0
77.00	425.0
78.00	348.0
79.00	187.0
81.10	186.0
82.90	152.0
83.80	483.0
85.80	424.0
91.10	365.0
95.90	227.0
97.10	201.0
103.90	222.0
105.10	162.0
115.20	249.0
116.90	161.0
119.00	242.0
133.00	308.0
134.90	188.0
147.00	742.0
149.00	179.0
165.00	206.0
167.10	168.0
177.00	173.0
178.90	167.0
191.00	445.0
193.10	251.0
207.10	3046.0
208.00	725.0
209.00	450.0
221.00	222.0
281.10	568.0
282.00	176.0
283.00	152.0
341.10	176.0
355.10	166.0

Instrument :
BNA_M
ClientSampleId :
DFTPP017

m/z	Abundance
36.00	381.0
37.10	1395.0
38.10	3815.0
39.10	21256.0
40.10	2289.0
41.10	870.0
42.00	222.0
43.00	385.0
44.00	5843.0
45.00	699.0
47.00	214.0
48.00	266.0
49.10	2558.0
50.10	82552.0
51.10	316160.0
52.10	15679.0
53.00	792.0
55.00	2055.0
56.00	8941.0
57.00	20304.0
58.00	997.0
59.00	271.0
60.00	348.0
61.00	4036.0
62.00	4412.0
63.10	12543.0
64.00	1518.0
65.10	5512.0
66.00	454.0
67.00	766.0
68.00	5256.0
69.00	300736.0
70.00	1773.0
70.90	199.0
73.00	3459.0
74.10	31792.0
75.00	49304.0
76.10	17024.0
77.10	365184.0
78.10	25512.0
79.00	22808.0
80.00	17576.0
81.00	24768.0
82.00	6221.0
83.00	5419.0
84.00	1382.0
85.00	4741.0
86.00	8107.0
87.10	3350.0
88.00	1742.0
89.00	531.0
90.10	237.0
91.00	5481.0
92.00	6293.0
93.00	41008.0
94.00	2839.0
95.10	784.0
96.00	2008.0
96.90	752.0
98.00	31016.0
99.00	28424.0
100.00	2791.0
101.00	13085.0
102.00	837.0
103.00	4312.0
104.00	7631.0
105.00	8108.0
106.10	2678.0
107.00	107400.0
108.10	17080.0
109.10	2609.0
110.00	192384.0
111.00	30160.0
112.00	3666.0
113.00	1136.0
114.10	264.0
115.10	672.0
116.00	5604.0
117.00	102072.0
118.00	7999.0
119.00	1247.0
120.00	1307.0
121.10	747.0
122.00	6575.0
123.00	11064.0
124.00	5266.0
125.00	4209.0
126.10	1552.0
127.00	361536.0
128.00	27152.0
129.00	136064.0
130.00	11956.0

Instrument :
BNA_M
ClientSampleId :
DFTPP017

131.00	2358.0
132.00	1203.0
133.00	733.0
134.00	4230.0
135.00	11206.0
136.00	4344.0
137.00	7390.0
138.10	1405.0
139.10	800.0
140.00	1692.0
141.00	18656.0
142.00	5492.0
143.00	4098.0
144.10	1143.0
145.10	1209.0
146.00	2892.0
147.00	9899.0
148.00	24768.0
149.00	5589.0
150.10	1369.0
151.20	1963.0
151.80	1117.0
153.00	5831.0
154.00	4557.0
155.10	10026.0
156.10	15457.0
157.10	3183.0
158.00	3173.0
159.10	2873.0
160.00	5357.0
161.00	8469.0
162.00	2278.0
163.00	892.0
164.00	1142.0
165.00	6824.0
166.00	5627.0
167.10	41448.0
168.00	31280.0
169.10	4186.0
170.00	1259.0
170.90	1495.0
172.00	2945.0
173.00	4071.0
174.00	8348.0
175.10	14960.0
176.00	4619.0
177.00	6581.0
178.00	2341.0
179.00	26600.0
180.00	19776.0
181.00	8390.0
182.00	1635.0
183.10	975.0
184.00	2301.0
185.00	13242.0
186.00	102384.0
187.10	30872.0
188.10	3379.0
189.00	7022.0
190.00	1303.0
191.10	3040.0
192.00	9894.0
193.10	10224.0
194.00	2436.0
195.10	1172.0
196.00	21592.0
196.80	4307.0
198.00	744960.0
199.00	51584.0
200.00	4049.0
201.60	2495.0
203.10	5008.0
204.00	26032.0
205.10	46112.0
206.10	189760.0
207.10	27656.0
208.00	6490.0
209.00	2803.0
210.20	2909.0
211.00	7023.0
211.80	1655.0
213.00	482.0
214.00	293.0
215.00	2066.0
216.00	4333.0
217.00	48656.0
218.00	6693.0
219.00	660.0
221.10	30424.0
223.00	10851.0
224.10	106992.0
225.10	27920.0
226.10	3633.0
227.00	43472.0
228.00	6305.0
229.00	8422.0

Instrument :
BNA_M
ClientSampleId :
DFTPP017

230.00	1342.0
231.00	3920.0
232.00	689.0
233.00	657.0
234.00	2481.0
235.10	3399.0
236.00	2109.0
237.00	4667.0
238.00	797.0
239.00	1409.0
240.00	1435.0
240.90	2537.0
242.10	5403.0
243.10	6313.0
244.10	82128.0
245.10	11264.0
246.00	14844.0
247.00	2882.0
248.00	684.0
249.00	3202.0
250.10	559.0
251.00	534.0
252.10	727.0
253.00	2155.0
254.00	2617.0
255.10	393472.0
256.00	58664.0
257.10	5515.0
258.00	22152.0
259.00	3735.0
260.10	822.0
261.00	600.0
262.90	162.0
263.90	399.0
265.00	8801.0
266.00	1419.0
267.10	233.0
270.00	227.0
271.10	1294.0
272.00	1077.0
273.00	11874.0
274.00	30464.0
275.10	175104.0
276.10	23864.0
277.10	16162.0
278.00	3005.0
279.00	556.0
281.00	747.0
282.10	530.0
283.00	1767.0
284.00	1122.0
285.10	2465.0
285.90	457.0
287.90	180.0
289.00	710.0
290.10	473.0
290.90	440.0
292.10	636.0
293.00	3299.0
293.90	1122.0
294.90	989.0
296.00	47016.0
297.00	7120.0
298.00	552.0
299.10	155.0
301.10	634.0
302.00	755.0
303.00	5887.0
304.10	1536.0
304.90	288.0
308.00	539.0
309.00	607.0
310.10	593.0
312.10	170.0
313.00	436.0
314.00	2082.0
315.00	5265.0
316.10	2947.0
317.00	702.0
321.00	1561.0
322.10	822.0
323.10	14514.0
324.10	2414.0
325.10	279.0
325.90	206.0
327.00	3062.0
328.00	1521.0
328.90	229.0
332.00	1013.0
333.00	1335.0
334.10	9250.0
335.10	2427.0
336.00	348.0
339.00	186.0
340.00	163.0
341.10	1693.0

Instrument :
BNA_M
ClientSampleId :
DFTPP017

342.10	486.0
343.00	156.0
346.00	2887.0
347.00	607.0
352.10	4178.0
353.10	2601.0
354.10	4615.0
355.10	1002.0
359.00	303.0
363.80	172.0
365.00	18480.0
366.00	2977.0
367.10	192.0
370.10	302.0
371.00	1109.0
372.10	7055.0
373.00	1933.0
374.10	227.0
377.10	279.0
383.00	2215.0
384.00	519.0
385.00	226.0
390.00	979.0
391.10	618.0
392.10	360.0
401.00	484.0
402.00	2283.0
403.10	3928.0
404.10	1788.0
405.10	352.0
410.00	153.0
415.00	166.0
421.10	3046.0
422.10	3669.0
423.10	22776.0
424.10	7586.0
425.10	1318.0
438.00	164.0
439.00	181.0
440.10	388.0
441.10	68856.0
442.10	438528.0
443.10	88144.0
444.10	7939.0
445.10	572.0

Instrument :
BNA_M
ClientSampleId :
DFTPP017

m/z	Abundance
36.10	225.0
37.00	1479.0
38.10	3074.0
39.10	21808.0
40.00	2275.0
41.10	981.0
43.00	639.0
44.00	6330.0
45.00	716.0
47.00	227.0
48.10	197.0
49.00	3001.0
50.10	79680.0
51.10	311616.0
52.10	15852.0
53.10	666.0
55.00	1972.0
56.00	8647.0
57.00	20688.0
57.90	866.0
59.10	385.0
60.00	290.0
61.00	3674.0
62.10	4496.0
63.10	12289.0
64.10	1726.0
65.10	6024.0
66.00	527.0
67.10	425.0
68.10	4708.0
69.00	302976.0
70.00	1684.0
71.00	248.0
72.10	194.0
73.10	3690.0
74.10	32856.0
75.00	49800.0
76.10	18328.0
77.10	373568.0
78.10	25968.0
79.00	22816.0
80.00	17440.0
81.00	25376.0
82.10	6356.0
83.00	5947.0
84.00	1142.0
85.00	4050.0
86.00	7130.0
87.00	3381.0
88.00	1365.0
89.00	641.0
91.00	5816.0
92.10	6318.0
93.00	42568.0
94.10	3030.0
95.00	1045.0
96.10	1991.0
96.80	543.0
97.20	645.0
98.00	32896.0
99.00	28072.0
100.00	2339.0
101.00	15153.0
101.90	883.0
103.00	5153.0
104.00	8727.0
105.00	8162.0
106.10	2979.0
107.00	113616.0
108.00	17520.0
109.10	3095.0
110.00	204288.0
111.00	30392.0
112.00	3282.0
113.00	1323.0
114.00	363.0
115.10	598.0
116.00	5724.0
117.00	107328.0
118.00	8389.0
119.00	1207.0
120.00	1772.0
120.90	616.0
122.00	7772.0
123.00	11138.0
124.00	5665.0
125.00	4442.0
126.10	1571.0
127.00	397696.0
128.00	29816.0
129.00	147008.0
130.00	12951.0

Instrument :
BNA_M
ClientSampleId :
DFTPP017

131.00	2828.0
132.10	1184.0
132.80	728.0
133.10	721.0
134.00	4272.0
135.00	12458.0
136.00	4665.0
137.10	7287.0
138.00	1431.0
138.90	651.0
140.00	2008.0
141.00	20632.0
142.00	7035.0
143.00	4560.0
144.00	1195.0
145.00	1244.0
146.00	3545.0
147.00	10364.0
148.00	26768.0
149.00	5858.0
150.00	1383.0
151.00	2872.0
151.70	1400.0
152.10	1394.0
153.00	5917.0
154.00	5063.0
155.00	11281.0
156.10	17312.0
157.10	3063.0
158.00	3734.0
159.00	2656.0
160.00	6110.0
161.00	9155.0
162.00	2733.0
163.00	907.0
163.90	1016.0
165.00	7587.0
166.00	6067.0
167.00	46432.0
168.00	30664.0
169.00	4183.0
170.00	1433.0
171.00	1600.0
172.00	4066.0
173.10	4547.0
174.00	8836.0
175.10	17488.0
176.10	4394.0
177.00	8201.0
178.00	2525.0
179.00	30752.0
180.00	22208.0
181.10	9701.0
182.00	1463.0
183.00	903.0
184.00	2317.0
185.10	14782.0
186.10	114944.0
187.10	33424.0
188.00	3687.0
189.00	7115.0
190.10	1497.0
191.00	3578.0
192.00	10450.0
193.10	11853.0
194.10	2581.0
195.10	1704.0
196.00	25240.0
197.00	4849.0
198.00	859648.0
199.00	59240.0
200.00	4455.0
201.50	3494.0
203.00	5933.0
204.10	30728.0
205.10	51384.0
206.10	223680.0
207.10	32288.0
208.00	7803.0
209.00	2794.0
210.20	3454.0
211.10	8412.0
211.80	1900.0
213.00	894.0
214.00	252.0
215.00	2688.0
216.10	5093.0
217.00	57648.0
218.00	7516.0
219.10	943.0
221.10	34648.0
221.80	8220.0
223.00	12929.0
224.10	124008.0
225.10	31312.0
226.10	3930.0

Instrument :
BNA_M
ClientSampleId :
DFTPP017

227.00	52064.0
228.10	8083.0
229.00	10873.0
230.10	1249.0
231.10	4482.0
232.00	863.0
233.00	859.0
234.10	3330.0
235.00	3634.0
236.00	2729.0
237.00	4852.0
238.10	727.0
239.00	2018.0
240.00	1652.0
241.00	2809.0
242.10	6984.0
243.00	6735.0
244.10	100968.0
245.10	13547.0
246.00	19192.0
247.00	4071.0
247.90	766.0
249.00	3816.0
249.90	673.0
251.00	891.0
252.10	935.0
252.90	2093.0
254.10	3630.0
255.10	492096.0
256.10	71304.0
257.10	5903.0
258.00	27184.0
259.00	4275.0
260.10	839.0
261.10	951.0
263.00	303.0
264.00	668.0
265.00	11127.0
266.00	1698.0
267.00	310.0
270.00	537.0
271.00	1270.0
272.00	1186.0
273.10	14735.0
274.00	40256.0
275.10	219200.0
276.10	29888.0
277.00	19280.0
278.00	3389.0
279.10	757.0
281.10	714.0
282.10	612.0
283.00	2480.0
284.00	1356.0
285.10	3501.0
286.00	560.0
287.90	194.0
289.00	717.0
289.90	602.0
291.00	374.0
292.00	989.0
293.10	4747.0
294.00	1441.0
295.10	1391.0
296.00	61584.0
297.00	8424.0
298.00	712.0
298.90	169.0
300.90	647.0
302.00	1223.0
303.10	7388.0
304.00	1822.0
305.00	287.0
308.00	948.0
309.00	683.0
309.90	693.0
310.90	172.0
312.10	212.0
313.10	562.0
314.00	3098.0
315.00	6485.0
316.10	4012.0
317.10	607.0
319.10	164.0
319.90	255.0
321.00	1834.0
322.10	898.0
323.10	19632.0
324.10	3536.0
325.00	439.0
325.90	508.0
327.00	3624.0
328.00	1738.0
329.10	343.0
332.10	1463.0
333.10	1781.0

Instrument :
BNA_M
ClientSampleId :
DFTPP017

334.10	12591.0
335.00	3326.0
336.00	329.0
338.90	314.0
339.90	247.0
341.00	2310.0
342.00	809.0
346.00	4444.0
347.00	1088.0
349.90	248.0
351.00	349.0
352.10	6140.0
353.00	3914.0
354.10	5736.0
355.00	1098.0
359.00	391.0
360.00	152.0
365.00	27792.0
366.00	3744.0
366.90	347.0
370.00	526.0
371.00	1878.0
372.10	10887.0
373.10	2711.0
374.00	303.0
376.90	337.0
383.00	2586.0
384.10	780.0
385.00	179.0
390.00	1603.0
391.00	961.0
392.10	658.0
401.00	646.0
402.00	3997.0
403.00	5663.0
404.10	2317.0
405.00	565.0
415.00	398.0
421.10	5149.0
422.00	4796.0
423.10	36184.0
424.10	10523.0
425.00	1120.0
437.70	221.0
438.20	171.0
439.00	262.0
439.90	321.0
441.10	103976.0
442.10	679360.0
443.10	131520.0
444.10	11543.0
445.00	700.0

Instrument :
BNA_M
ClientSampleId :
DFTPP017

m/z	Abundance
36.00	235.0
37.10	868.0
38.10	1820.0
39.10	13285.0
40.00	1982.0
41.00	898.0
42.00	316.0
43.10	515.0
44.00	7083.0
45.00	539.0
47.00	210.0
48.00	219.0
49.00	1937.0
50.10	47136.0
51.10	185536.0
52.10	8779.0
53.10	471.0
55.00	1604.0
56.00	4893.0
57.00	12483.0
58.10	559.0
59.10	241.0
60.10	204.0
61.00	2392.0
62.00	2364.0
63.00	6855.0
64.00	873.0
65.10	4246.0
65.90	317.0
67.00	559.0
68.00	2756.0
69.00	176640.0
70.00	1186.0
71.00	204.0
73.10	3084.0
74.10	18776.0
75.10	29888.0
76.10	10484.0
77.10	224384.0
78.10	15039.0
79.00	13912.0
80.00	10596.0
81.00	14639.0
82.00	3814.0
83.00	3370.0
83.90	910.0
85.00	2818.0
86.00	4819.0
87.00	2076.0
88.00	1106.0
89.00	503.0
89.90	162.0
91.00	3791.0
92.00	4197.0
93.00	25240.0
94.00	1793.0
95.00	635.0
96.00	1561.0
96.90	571.0
98.00	20544.0
99.00	17856.0
100.00	1493.0
101.00	8512.0
101.90	519.0
103.00	3142.0
104.00	5433.0
105.00	5667.0
106.10	1776.0
107.00	67232.0
108.00	9967.0
109.00	2118.0
110.00	125464.0
111.00	18664.0
112.00	2610.0
113.00	936.0
113.80	179.0
114.90	449.0
116.00	3738.0
117.00	65112.0
118.00	5187.0
119.00	832.0
120.00	972.0
121.10	588.0
122.00	4408.0
123.00	6994.0
124.00	3263.0
125.10	2465.0
126.10	937.0
127.00	244352.0
128.00	18408.0
129.00	91000.0
130.00	7808.0

Instrument :
BNA_M
ClientSampleId :
DFTPP017

131.00	1710.0
131.90	936.0
133.00	609.0
134.00	2471.0
135.00	8008.0
136.00	3221.0
137.00	5587.0
137.80	837.0
139.00	501.0
140.10	1241.0
141.00	12498.0
142.00	4179.0
143.00	2719.0
144.00	705.0
145.00	684.0
146.10	2427.0
147.00	7408.0
148.00	16568.0
149.10	3772.0
150.00	1041.0
151.00	1666.0
151.80	1008.0
153.00	3759.0
154.00	2962.0
155.00	6984.0
156.00	10269.0
157.10	2328.0
158.00	2245.0
159.00	1737.0
160.00	4357.0
161.00	5944.0
162.00	1656.0
163.00	693.0
164.00	720.0
165.00	5029.0
166.00	4093.0
167.00	29640.0
168.00	20136.0
169.00	2744.0
170.10	952.0
171.00	1155.0
172.00	2362.0
173.00	3458.0
174.10	5362.0
175.10	10912.0
176.00	2671.0
177.00	4633.0
178.00	1481.0
179.00	19608.0
180.00	14670.0
181.00	6022.0
182.00	1086.0
182.80	394.0
183.00	395.0
184.00	1608.0
185.00	9892.0
186.00	74232.0
187.00	21136.0
188.00	2107.0
189.00	5327.0
190.00	846.0
191.00	2415.0
192.00	6608.0
193.00	7221.0
194.00	1514.0
195.10	892.0
196.00	16189.0
197.00	2823.0
198.00	575360.0
199.00	37888.0
200.00	2633.0
201.60	2817.0
203.00	3904.0
204.00	19984.0
205.10	35440.0
206.10	147072.0
207.10	22000.0
208.00	5623.0
209.00	2035.0
210.20	2426.0
211.00	5932.0
212.90	559.0
214.10	221.0
215.00	1461.0
216.00	3656.0
217.00	39576.0
218.00	5487.0
218.90	389.0
221.10	24120.0
221.80	5508.0
223.00	9169.0
224.10	87632.0
225.10	21120.0
226.10	2418.0
227.00	34208.0
228.00	4698.0

Instrument :
BNA_M
ClientSampleId :
DFTPP017

229.00	7465.0
230.00	920.0
231.00	2977.0
232.00	542.0
233.00	766.0
234.00	2170.0
235.00	2580.0
236.00	1884.0
237.00	3337.0
237.90	583.0
239.00	1526.0
240.00	1186.0
241.00	1900.0
242.00	5128.0
243.10	5154.0
244.10	69504.0
245.10	8510.0
246.00	13711.0
247.00	2415.0
247.90	633.0
249.00	2468.0
250.10	436.0
251.00	584.0
252.00	698.0
253.00	1593.0
254.00	2260.0
255.00	343104.0
256.10	47736.0
257.10	3434.0
258.00	19168.0
259.00	3058.0
260.00	430.0
261.00	519.0
263.00	193.0
264.00	505.0
265.00	8547.0
266.10	1068.0
267.20	271.0
270.00	279.0
271.00	1140.0
272.00	899.0
273.00	10125.0
274.00	27536.0
275.10	153408.0
276.10	19720.0
277.00	12978.0
278.00	2369.0
279.00	475.0
281.00	676.0
282.00	516.0
283.10	1591.0
284.00	817.0
285.00	2252.0
286.00	345.0
288.00	168.0
289.00	571.0
289.90	343.0
291.00	396.0
292.10	617.0
293.00	3558.0
294.10	857.0
295.10	809.0
296.00	46160.0
297.10	6317.0
298.00	584.0
301.00	484.0
301.90	733.0
303.10	5482.0
304.10	1368.0
308.00	645.0
308.90	520.0
310.00	776.0
311.10	176.0
313.00	408.0
314.10	2027.0
315.00	5753.0
316.00	2531.0
317.10	524.0
320.00	181.0
321.00	1649.0
322.10	631.0
323.10	15157.0
324.00	2625.0
324.90	260.0
326.10	256.0
327.00	2895.0
328.00	1417.0
329.00	297.0
332.00	1083.0
333.00	1258.0
334.10	9095.0
335.10	2300.0
336.00	447.0
339.00	292.0
340.00	274.0
341.00	1469.0

Instrument :
BNA_M
ClientSampleId :
DFTPP017

342.00	713.0
343.10	195.0
346.00	3831.0
347.00	620.0
350.90	300.0
352.10	4595.0
353.10	2938.0
354.10	5228.0
355.10	1060.0
359.20	323.0
365.00	21232.0
366.00	2977.0
367.00	271.0
370.10	434.0
371.10	992.0
372.10	7717.0
373.10	2176.0
374.10	183.0
376.90	264.0
383.00	2553.0
384.10	686.0
385.00	245.0
390.10	1119.0
391.10	712.0
392.00	563.0
401.00	496.0
402.00	3677.0
403.00	4403.0
404.10	1252.0
404.90	300.0
410.10	174.0
415.00	238.0
421.00	3830.0
422.00	3890.0
423.10	28384.0
424.10	7340.0
425.00	737.0
438.00	187.0
439.10	240.0
440.10	183.0
441.10	85784.0
442.10	561600.0
443.10	107592.0
444.10	8900.0
445.00	675.0

Instrument :
BNA_M
ClientSampleId :
DFTPP017