

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100221\
 Data File : BM032308.D
 Acq On : 01 Oct 2021 10:31
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC010

Quant Time: Oct 01 12:23:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM100221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 12:21:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.637	152	325069	20.000	ng	0.00
21) Naphthalene-d8	10.431	136	1280016	20.000	ng	# 0.00
39) Acenaphthene-d10	14.283	164	813375	20.000	ng	0.00
64) Phenanthrene-d10	17.013	188	1686959	20.000	ng	# 0.00
76) Chrysene-d12	21.177	240	1649539	20.000	ng	0.00
86) Perylene-d12	23.336	264	1563611	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.207	112	421271	21.899	ng	0.00
7) Phenol-d6	6.825	99	579426	20.740	ng	0.00
23) Nitrobenzene-d5	8.790	82	476619	20.926	ng	-0.01
42) 2,4,6-Tribromophenol	15.771	330	193872	21.957	ng	0.00
45) 2-Fluorobiphenyl	12.901	172	1207501	23.993	ng	0.00
79) Terphenyl-d14	19.624	244	1821693	25.554	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.013	88	89707	11.158	ng	92
3) Pyridine	3.425	79	240378	10.378	ng	92
4) n-Nitrosodimethylamine	3.337	42	96108	10.170	ng	# 65
6) Aniline	6.960	93	311264	9.916	ng	91
8) 2-Chlorophenol	7.207	128	228631	10.656	ng	92
9) Benzaldehyde	6.766	77	173281	10.710	ng	85
10) Phenol	6.848	94	275996	10.254	ng	81
11) bis(2-Chloroethyl)ether	7.054	93	219725	10.241	ng	94
12) 1,3-Dichlorobenzene	7.531	146	258668	10.873	ng	# 91
13) 1,4-Dichlorobenzene	7.672	146	264555	10.918	ng	97
14) 1,2-Dichlorobenzene	7.995	146	256040	10.866	ng	97
15) Benzyl Alcohol	7.878	79	190362	9.892	ng	88
16) 2,2'-oxybis(1-Chloropr...	8.172	45	278469	9.757	ng	98
17) 2-Methylphenol	8.095	107	181839	10.026	ng	# 90
18) Hexachloroethane	8.725	117	86634	10.737	ng	94
19) n-Nitroso-di-n-propyla...	8.443	70	154344	9.784	ng	# 85
20) 3+4-Methylphenols	8.425	107	249739	10.005	ng	93
22) Acetophenone	8.454	105	332326	10.590	ng	# 88
24) Nitrobenzene	8.831	77	229493	10.371	ng	# 90
25) Isophorone	9.354	82	386392	9.912	ng	# 94
26) 2-Nitrophenol	9.548	139	95422	9.645	ng	# 77
27) 2,4-Dimethylphenol	9.619	122	181357	10.397	ng	93
28) bis(2-Chloroethoxy)met...	9.837	93	292710	10.357	ng	99
29) 2,4-Dichlorophenol	10.084	162	199832	10.472	ng	98
30) 1,2,4-Trichlorobenzene	10.301	180	235918	11.148	ng	98
31) Naphthalene	10.478	128	711112	10.841	ng	99
32) Benzoic acid	9.719	122	90492	7.243	ng	# 82
33) 4-Chloroaniline	10.584	127	283204	10.150	ng	# 94
34) Hexachlorobutadiene	10.789	225	138673	11.245	ng	98
35) Caprolactam	11.307	113	55448	7.961	ng	# 84
36) 4-Chloro-3-methylphenol	11.725	107	215842	10.113	ng	96
37) 2-Methylnaphthalene	12.313	142	472524	10.440	ng	95
38) 1-Methylnaphthalene	12.313	142	472524	10.662	ng	# 95
40) 1,2,4,5-Tetrachloroben...	12.478	216	251546	11.056	ng	# 95
41) Hexachlorocyclopentadiene	12.472	237	105356	10.565	ng	99
43) 2,4,6-Trichlorophenol	12.719	196	165331	10.479	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100221\
 Data File : BM032308.D
 Acq On : 01 Oct 2021 10:31
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC010

Quant Time: Oct 01 12:23:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM100221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 12:21:14 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.783	196	177788	10.411	ng	# 91
46) 1,1'-Biphenyl	13.113	154	652114	10.920	ng	99
47) 2-Chloronaphthalene	13.154	162	491289	10.800	ng	97
48) 2-Nitroaniline	13.354	65	117548	9.366	ng	# 85
49) Acenaphthylene	14.001	152	757014	10.675	ng	100
50) Dimethylphthalate	13.736	163	613931	10.793	ng	99
51) 2,6-Dinitrotoluene	13.848	165	116694	10.297	ng	# 65
52) Acenaphthene	14.348	154	497767	10.513	ng	98
53) 3-Nitroaniline	14.177	138	126959	9.570	ng	87
54) 2,4-Dinitrophenol	14.377	184	50098	7.576	ng	# 72
55) Dibenzofuran	14.683	168	786903	10.940	ng	91
56) 4-Nitrophenol	14.495	139	103571	9.186	ng	# 69
57) 2,4-Dinitrotoluene	14.636	165	149655	9.929	ng	# 62
58) Fluorene	15.330	166	606552	11.275	ng	97
59) 2,3,4,6-Tetrachlorophenol	14.913	232	157006	10.540	ng	# 83
60) Diethylphthalate	15.095	149	604232	10.747	ng	96
61) 4-Chlorophenyl-phenyle...	15.319	204	312084	11.466	ng	92
62) 4-Nitroaniline	15.336	138	127069	9.371	ng	# 68
63) Azobenzene	15.613	77	564791	10.392	ng	84
65) 4,6-Dinitro-2-methylph...	15.401	198	81754	8.464	ng	# 77
66) n-Nitrosodiphenylamine	15.530	169	531199	10.772	ng	97
67) 4-Bromophenyl-phenylether	16.207	248	184749	10.970	ng	92
68) Hexachlorobenzene	16.342	284	207230	11.123	ng	# 83
69) Atrazine	16.471	200	176833	10.698	ng	97
70) Pentachlorophenol	16.677	266	114571	9.940	ng	96
71) Phenanthrene	17.054	178	984586	11.049	ng	99
72) Anthracene	17.142	178	944579	10.939	ng	100
73) Carbazole	17.407	167	940913	10.677	ng	98
74) Di-n-butylphthalate	17.965	149	945164	10.551	ng	# 97
75) Fluoranthene	19.059	202	1110273	11.010	ng	98
77) Benzidine	19.236	184	378173	9.285	ng	98
78) Pyrene	19.424	202	1162076	11.311	ng	100
80) Butylbenzylphthalate	20.312	149	384400	9.799	ng	93
81) Benzo(a)anthracene	21.159	228	1114056	10.979	ng	99
82) 3,3'-Dichlorobenzidine	21.095	252	341192	10.476	ng	99
83) Chrysene	21.212	228	1095202	11.072	ng	100
84) Bis(2-ethylhexyl)phtha...	21.083	149	554310	10.448	ng	95
85) Di-n-octyl phthalate	21.930	149	899517	9.446	ng	# 92
87) Indeno(1,2,3-cd)pyrene	25.465	276	1029564	9.868	ng	96
88) Benzo(b)fluoranthene	22.689	252	1030539	11.718	ng	99
89) Benzo(k)fluoranthene	22.730	252	996120	11.144	ng	99
90) Benzo(a)pyrene	23.236	252	807066	10.756	ng	98
91) Dibenzo(a,h)anthracene	25.465	278	874157	10.108	ng	# 94
92) Benzo(g,h,i)perylene	26.124	276	843732	9.419	ng	# 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100221\
 Data File : BM032308.D
 Acq On : 01 Oct 2021 10:31
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC010

Quant Time: Oct 01 12:23:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM100221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 12:21:14 2021
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.637 min Scan# 816
Delta R.T. -0.005 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:152 Resp: 325069
Ion Ratio Lower Upper
152 100
150 157.3 122.4 183.6
115 59.2 53.8 80.8



#2
1,4-Dioxane
Concen: 11.158 ng
RT: 3.013 min Scan# 30
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion: 88 Resp: 89707
Ion Ratio Lower Upper
88 100
58 71.5 50.4 75.6
43 26.7 22.2 33.2









#3
Pyridine
Concen: 10.378 ng
RT: 3.425 min Scan# 100
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 79 Resp: 240378
Ion Ratio Lower Upper
79 100
52 69.3 48.6 73.0
51 35.1 27.4 41.2



#4
n-Nitrosodimethylamine
Concen: 10.170 ng
RT: 3.337 min Scan# 85
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion: 42 Resp: 96108
Ion Ratio Lower Upper
42 100
74 137.0 99.5 149.3
44 8.3 57.3 85.9









#5
2-Fluorophenol
Concen: 21.899 ng
RT: 5.207 min Scan# 403
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:112 Resp: 421271
Ion Ratio Lower Upper
112 100
64 56.6 53.7 80.5
63 29.8 33.9 50.9#



#6
Aniline
Concen: 9.916 ng
RT: 6.960 min Scan# 701
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion: 93 Resp: 311264
Ion Ratio Lower Upper
93 100
66 37.5 35.3 52.9
65 19.7 18.7 28.1









#7
Phenol-d6
Concen: 20.740 ng
RT: 6.825 min Scan# 678
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion: 99 Resp: 579426
Ion Ratio Lower Upper
99 100
42 16.3 23.0 34.6#
71 32.6 37.8 56.8#



#8
2-Chlorophenol
Concen: 10.656 ng
RT: 7.207 min Scan# 743
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion: 128 Resp: 228631
Ion Ratio Lower Upper
128 100
130 31.7 13.0 53.0
64 45.7 33.4 73.4









#9
Benzaldehyde
Concen: 10.710 ng
RT: 6.766 min Scan# 668
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion: 77 Resp: 173281
Ion Ratio Lower Upper
77 100
105 100.4 68.9 108.9
106 97.8 61.2 101.2



#10
Phenol
Concen: 10.254 ng
RT: 6.848 min Scan# 682
Delta R.T. -0.012 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion: 94 Resp: 275996
Ion Ratio Lower Upper
94 100
65 27.5 16.9 56.9
66 35.9 30.0 70.0









#11
bis(2-Chloroethyl)ether
Concen: 10.241 ng
RT: 7.054 min Scan# 717
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion: 93 Resp: 219725
Ion Ratio Lower Upper
93 100
63 76.8 63.7 103.7
95 32.8 12.4 52.4



#12
1,3-Dichlorobenzene
Concen: 10.873 ng
RT: 7.531 min Scan# 798
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:146 Resp: 258668
Ion Ratio Lower Upper
146 100
148 64.6 52.3 78.5
75 28.9 34.2 51.2#









#13
1,4-Dichlorobenzene
Concen: 10.918 ng
RT: 7.672 min Scan# 822
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:146 Resp: 264555
Ion Ratio Lower Upper
146 100
148 63.4 51.1 76.7
111 39.8 35.3 52.9



#14
1,2-Dichlorobenzene
Concen: 10.866 ng
RT: 7.995 min Scan# 877
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:146 Resp: 256040
Ion Ratio Lower Upper
146 100
148 64.6 51.7 77.5
111 42.8 38.2 57.2









#15
Benzyl Alcohol
Concen: 9.892 ng
RT: 7.878 min Scan# 857
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion: 79 Resp: 190362
Ion Ratio Lower Upper
79 100
108 77.5 53.6 80.4
77 62.8 57.7 86.5



#16
2,2'-oxybis(1-Chloropropane)
Concen: 9.757 ng
RT: 8.172 min Scan# 907
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion: 45 Resp: 278469
Ion Ratio Lower Upper
45 100
77 15.1 0.0 36.1
79 11.5 0.0 30.5









#17
2-Methylphenol
Concen: 10.026 ng
RT: 8.095 min Scan# 894
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:	107	Resp:	181839
Ion	Ratio	Lower	Upper
107	100		
108	113.2	87.9	131.9
77	45.5	48.2	72.2#
79	45.1	44.7	67.1



#18
Hexachloroethane
Concen: 10.737 ng
RT: 8.725 min Scan# 1001
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:	117	Resp:	86634
Ion	Ratio	Lower	Upper
117	100		
119	93.8	77.7	116.5
201	96.5	70.2	105.2









#19
n-Nitroso-di-n-propylamine
Concen: 9.784 ng
RT: 8.443 min Scan# 953
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 70 Resp: 154344
Ion Ratio Lower Upper
70 100
42 54.8 56.3 84.5#
101 9.9 7.8 11.6
130 20.4 14.2 21.4



#20
3+4-Methylphenols
Concen: 10.005 ng
RT: 8.425 min Scan# 950
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion: 107 Resp: 249739
Ion Ratio Lower Upper
107 100
108 89.4 67.6 107.6
77 33.8 21.9 61.9
79 28.0 15.3 55.3









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.431 min Scan# 1291
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:136 Resp: 1280016
Ion Ratio Lower Upper
136 100
137 10.5 8.7 13.1
54 7.8 7.4 11.0
68 6.2 3.8 5.8#



#22
Acetophenone
Concen: 10.590 ng
RT: 8.454 min Scan# 955
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:105 Resp: 332326
Ion Ratio Lower Upper
105 100
71 3.8 3.8 5.6#
51 26.6 29.5 44.3#
120 23.2 16.3 24.5









#23
Nitrobenzene-d5
Concen: 20.926 ng
RT: 8.790 min Scan# 1012
Delta R.T. -0.012 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion: 82 Resp: 476619
Ion Ratio Lower Upper
82 100
128 43.0 29.0 43.6
54 51.1 45.8 68.8



#24
Nitrobenzene
Concen: 10.371 ng
RT: 8.831 min Scan# 1019
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion: 77 Resp: 229493
Ion Ratio Lower Upper
77 100
123 45.5 30.1 45.1#
65 13.3 11.7 17.5





#25

Isophorone

Concen: 9.912 ng

RT: 9.354 min Scan# 1108

Delta R.T. 0.000 min

Lab File: BM032308.D

Acq: 01 Oct 2021 10:31

Tgt Ion: 82 Resp: 386392

Ion Ratio Lower Upper

82 100

95 7.2 6.4 9.6





138 18.7 12.4 18.6#

Instrument :

BNA_M

ClientSampleId :

SSTDIC010



#26

2-Nitrophenol

Concen: 9.645 ng

RT: 9.548 min Scan# 1141

Delta R.T. 0.000 min

Lab File: BM032308.D

Acq: 01 Oct 2021 10:31




Tgt Ion:139 Resp: 95422


Ion Ratio Lower Upper

139 100

109 31.7 34.8 52.2#

65 52.6 59.9 89.9#









#27
2,4-Dimethylphenol
Concen: 10.397 ng
RT: 9.619 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:122 Resp: 181357
Ion Ratio Lower Upper
122 100
107 114.7 100.6 150.8
121 58.4 47.3 70.9



#28
bis(2-Chloroethoxy)methane
Concen: 10.357 ng
RT: 9.837 min Scan# 1190
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion: 93 Resp: 292710
Ion Ratio Lower Upper
93 100
95 33.1 25.8 38.8
123 11.3 8.8 13.2









#29
2,4-Dichlorophenol
Concen: 10.472 ng
RT: 10.084 min Scan# 1232
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:162 Resp: 199832
Ion Ratio Lower Upper
162 100
164 65.5 47.4 87.4
98 38.0 18.9 58.9



#30
1,2,4-Trichlorobenzene
Concen: 11.148 ng
RT: 10.301 min Scan# 1269
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:180 Resp: 235918
Ion Ratio Lower Upper
180 100
182 95.3 75.5 113.3
145 29.7 26.2 39.2









#31
Naphthalene
Concen: 10.841 ng
RT: 10.478 min Scan# 1299
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:128 Resp: 711112
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 12.5 10.4 15.6



#32
Benzoic acid
Concen: 7.243 ng
RT: 9.719 min Scan# 1170
Delta R.T. -0.076 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:122 Resp: 90492
Ion Ratio Lower Upper
122 100
105 122.2 120.3 160.3
77 95.6 100.0 140.0#









#33
4-Chloroaniline
Concen: 10.150 ng
RT: 10.584 min Scan# 1317
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:	127	Resp:	283204
Ion	Ratio	Lower	Upper
127	100		
129	31.8	25.6	38.4
65	30.0	30.4	45.6#
92	19.8	16.2	24.4



#34
Hexachlorobutadiene
Concen: 11.245 ng
RT: 10.789 min Scan# 1352
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:	225	Resp:	138673
Ion	Ratio	Lower	Upper
225	100		
223	63.9	52.6	78.8
227	65.5	51.2	76.8









#35
Caprolactam
Concen: 7.961 ng
RT: 11.307 min Scan# 1440
Delta R.T. -0.041 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:113 Resp: 55448
Ion Ratio Lower Upper
113 100
55 164.6 170.5 210.5#
56 127.6 125.2 165.2



#36
4-Chloro-3-methylphenol
Concen: 10.113 ng
RT: 11.725 min Scan# 1511
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:107 Resp: 215842
Ion Ratio Lower Upper
107 100
144 25.2 19.3 28.9
142 79.9 60.5 90.7









#37
2-Methylnaphthalene
Concen: 10.440 ng
RT: 12.313 min Scan# 1611
Delta R.T. 0.218 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:142 Resp: 472524
Ion Ratio Lower Upper
142 100
141 89.9 71.8 107.6
115 36.4 36.3 54.5



#38
1-Methylnaphthalene
Concen: 10.662 ng
RT: 12.313 min Scan# 1611
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:142 Resp: 472524
Ion Ratio Lower Upper
142 100
141 89.9 75.6 113.4
116 3.8 4.0 6.0#









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.283 min Scan# 1946
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:164 Resp: 813375
Ion Ratio Lower Upper
164 100
162 99.2 80.6 121.0
160 43.5 35.1 52.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 11.056 ng
RT: 12.478 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:216 Resp: 251546
Ion Ratio Lower Upper
216 100
214 79.8 62.5 93.7
179 20.7 18.7 28.1
108 19.0 21.0 31.4#









#41
Hexachlorocyclopentadiene
Concen: 10.565 ng
RT: 12.472 min Scan# 1638
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:	237	Resp:	105356
Ion Ratio	Lower	Upper	
237	100		
235	64.4	44.4	84.4
272	13.6	0.0	31.4



#42
2,4,6-Tribromophenol
Concen: 21.957 ng
RT: 15.771 min Scan# 2199
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:	330	Resp:	193872
Ion Ratio	Lower	Upper	
330	100		
332	96.3	75.7	113.5
141	37.7	50.7	76.1#









#43
2,4,6-Trichlorophenol
Concen: 10.479 ng
RT: 12.719 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:196 Resp: 165331
Ion Ratio Lower Upper
196 100
198 94.6 73.6 110.4
200 30.2 22.1 33.1



#44
2,4,5-Trichlorophenol
Concen: 10.411 ng
RT: 12.783 min Scan# 1691
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:196 Resp: 177788
Ion Ratio Lower Upper
196 100
198 96.2 77.0 115.4
97 49.1 53.9 80.9#
132 26.0 23.8 35.8









#45
2-Fluorobiphenyl
Concen: 23.993 ng
RT: 12.901 min Scan# 1711
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:172 Resp: 1207501
Ion Ratio Lower Upper
172 100
171 34.8 28.1 42.1
170 22.8 18.6 27.8



#46
1,1'-Biphenyl
Concen: 10.920 ng
RT: 13.113 min Scan# 1747
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:154 Resp: 652114
Ion Ratio Lower Upper
154 100
153 40.1 20.3 60.3
76 12.5 0.0 33.5









#47
2-Chloronaphthalene
Concen: 10.800 ng
RT: 13.154 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:162 Resp: 491289
Ion Ratio Lower Upper
162 100
127 40.0 33.2 49.8
164 33.8 25.7 38.5



#48
2-Nitroaniline
Concen: 9.366 ng
RT: 13.354 min Scan# 1788
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion: 65 Resp: 117548
Ion Ratio Lower Upper
65 100
92 68.3 51.2 76.8
138 104.4 66.8 100.2#









#49
Acenaphthylene
Concen: 10.675 ng
RT: 14.001 min Scan# 1898
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:152 Resp: 757014
Ion Ratio Lower Upper
152 100
151 19.5 15.7 23.5
153 13.2 10.6 15.8



#50
Dimethylphthalate
Concen: 10.793 ng
RT: 13.736 min Scan# 1853
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:163 Resp: 613931
Ion Ratio Lower Upper
163 100
194 4.0 2.9 4.3
164 9.9 8.0 12.0









#51
2,6-Dinitrotoluene
Concen: 10.297 ng
RT: 13.848 min Scan# 1872
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:165 Resp: 116694
Ion Ratio Lower Upper
165 100
63 55.6 72.4 108.6#
89 56.5 69.3 103.9#



#52
Acenaphthene
Concen: 10.513 ng
RT: 14.348 min Scan# 1957
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:154 Resp: 497767
Ion Ratio Lower Upper
154 100
153 113.4 92.4 138.6
152 53.8 43.7 65.5









#53
3-Nitroaniline
Concen: 9.570 ng
RT: 14.177 min Scan# 1928
Delta R.T. -0.012 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:138 Resp: 126959
Ion Ratio Lower Upper
138 100
108 11.0 8.3 12.5
92 115.4 105.2 157.8



#54
2,4-Dinitrophenol
Concen: 7.576 ng
RT: 14.377 min Scan# 1962
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:184 Resp: 50098
Ion Ratio Lower Upper
184 100
63 71.5 93.8 140.6#
154 72.1 64.0 96.0









#55
Dibenzenofuran
Concen: 10.940 ng
RT: 14.683 min Scan# 2014
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:168 Resp: 786903
Ion Ratio Lower Upper
168 100
139 36.8 35.0 52.4
169 12.8 10.8 16.2



#56
4-Nitrophenol
Concen: 9.186 ng
RT: 14.495 min Scan# 1982
Delta R.T. -0.012 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:139 Resp: 103571
Ion Ratio Lower Upper
139 100
109 75.5 92.3 132.3#
65 98.2 110.5 150.5#









#57
2,4-Dinitrotoluene
Concen: 9.929 ng
RT: 14.636 min Scan# 2006
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:165 Resp: 149655
Ion Ratio Lower Upper
165 100
63 42.2 70.3 105.5#
89 73.9 83.1 124.7#
182 2.6 2.2 3.4



#58
Fluorene
Concen: 11.275 ng
RT: 15.330 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:166 Resp: 606552
Ion Ratio Lower Upper
166 100
165 96.3 79.3 118.9
167 12.9 10.6 16.0









#59
2,3,4,6-Tetrachlorophenol
Concen: 10.540 ng
RT: 14.913 min Scan# 2053
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:232 Resp: 157006
Ion Ratio Lower Upper
232 100
131 46.0 50.6 75.8#
130 2.5 3.0 4.4#
166 31.2 28.8 43.2



#60
Diethylphthalate
Concen: 10.747 ng
RT: 15.095 min Scan# 2084
Delta R.T. -0.012 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:149 Resp: 604232
Ion Ratio Lower Upper
149 100
177 22.2 15.7 23.5
150 12.6 9.6 14.4









#61
4-Chlorophenyl-phenylether
Concen: 11.466 ng
RT: 15.319 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:204 Resp: 312084
Ion Ratio Lower Upper
204 100
206 33.1 26.7 40.1
141 60.6 55.8 83.8



#62
4-Nitroaniline
Concen: 9.371 ng
RT: 15.336 min Scan# 2125
Delta R.T. -0.012 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:138 Resp: 127069
Ion Ratio Lower Upper
138 100
92 50.6 40.6 80.6
108 78.0 103.8 143.8#









#63
Azobenzene
Concen: 10.392 ng
RT: 15.613 min Scan# 2172
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion: 77	Resp: 564791
Ion Ratio	Lower Upper
77 100	
182 27.2	0.0 39.3
105 22.5	0.0 36.1
51 29.4	19.1 59.1



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.013 min Scan# 2410
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:188	Resp: 1686959		
Ion Ratio	Lower	Upper	
188 100			
94 9.5	6.3	9.5#	
80 10.3	7.3	10.9	









#65
4,6-Dinitro-2-methylphenol
Concen: 8.464 ng
RT: 15.401 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:198 Resp: 81754
Ion Ratio Lower Upper
198 100
51 45.0 49.7 89.7#
105 44.3 34.9 74.9



#66
n-Nitrosodiphenylamine
Concen: 10.772 ng
RT: 15.530 min Scan# 2158
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:169 Resp: 531199
Ion Ratio Lower Upper
169 100
168 66.0 55.0 82.4
167 36.3 29.8 44.8









#67
4-Bromophenyl-phenylether
Concen: 10.970 ng
RT: 16.207 min Scan# 2273
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:248 Resp: 184749
Ion Ratio Lower Upper
248 100
250 98.6 77.4 116.2
141 79.0 74.1 111.1



#68
Hexachlorobenzene
Concen: 11.123 ng
RT: 16.342 min Scan# 2296
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:284 Resp: 207230
Ion Ratio Lower Upper
284 100
142 35.8 43.0 64.6#
249 32.2 27.6 41.4





#69

Atrazine

Concen: 10.698 ng

RT: 16.471 min Scan# 2318

Delta R.T. -0.006 min

Lab File: BM032308.D

Acq: 01 Oct 2021 10:31

Instrument :

BNA_M

ClientSampleId :

SSTDIC010





Tgt Ion:200 Resp: 176833

Ion Ratio Lower Upper

200 100

173 25.7 5.8 45.8

215 46.3 23.6 63.6



#70

Pentachlorophenol

Concen: 9.940 ng

RT: 16.677 min Scan# 2353

Delta R.T. 0.000 min

Lab File: BM032308.D

Acq: 01 Oct 2021 10:31




Tgt Ion:266 Resp: 114571


Ion Ratio Lower Upper

266 100

268 64.9 50.3 75.5

264 61.6 52.1 78.1









#71
Phenanthrene
Concen: 11.049 ng
RT: 17.054 min Scan# 2417
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:178 Resp: 984586
Ion Ratio Lower Upper
178 100
176 18.9 15.6 23.4
179 15.4 12.4 18.6



#72
Anthracene
Concen: 10.939 ng
RT: 17.142 min Scan# 2432
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:178 Resp: 944579
Ion Ratio Lower Upper
178 100
176 18.5 14.8 22.2
179 15.3 12.3 18.5





#73

Carbazole

Concen: 10.677 ng

RT: 17.407 min Scan# 2477

Delta R.T. -0.006 min

Lab File: BM032308.D

Acq: 01 Oct 2021 10:31





Tgt Ion:167 Resp: 940913

Ion Ratio Lower Upper

167 100

166 21.5 17.4 26.0

139 12.8 11.9 17.9



#74

Di-n-butylphthalate

Concen: 10.551 ng

RT: 17.965 min Scan# 2572

Delta R.T. -0.006 min

Lab File: BM032308.D

Acq: 01 Oct 2021 10:31




Tgt Ion:149 Resp: 945164


Ion Ratio Lower Upper

149 100

150 9.5 7.3 10.9

104 5.1 5.6 8.4#









#75
Fluoranthene
Concen: 11.010 ng
RT: 19.059 min Scan# 2758
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:202 Resp: 1110273
Ion Ratio Lower Upper
202 100
101 11.3 0.0 29.9
203 17.2 0.0 36.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.177 min Scan# 3118
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:240 Resp: 1649539
Ion Ratio Lower Upper
240 100
120 11.3 7.8 11.6
236 25.6 20.9 31.3









#77
Benzidine
Concen: 9.285 ng
RT: 19.236 min Scan# 2788
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:184 Resp: 378173
Ion Ratio Lower Upper
184 100
185 14.6 10.6 16.0
183 11.6 9.0 13.4



#78
Pyrene
Concen: 11.311 ng
RT: 19.424 min Scan# 2820
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:202 Resp: 1162076
Ion Ratio Lower Upper
202 100
200 20.6 16.6 25.0
203 17.3 13.7 20.5









#79
Terphenyl-d14
Concen: 25.554 ng
RT: 19.624 min Scan# 2854
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:244 Resp: 1821693
Ion Ratio Lower Upper
244 100
212 7.1 6.5 9.7
122 11.1 8.7 13.1



#80
Butylbenzylphthalate
Concen: 9.799 ng
RT: 20.312 min Scan# 2971
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:149 Resp: 384400
Ion Ratio Lower Upper
149 100
91 73.5 64.0 96.0
206 21.4 14.4 21.6









#81
Benzo(a)anthracene
Concen: 10.979 ng
RT: 21.159 min Scan# 3115
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:228 Resp: 1114056
Ion Ratio Lower Upper
228 100
226 27.1 21.4 32.2
229 19.8 15.3 22.9



#82
3,3'-Dichlorobenzidine
Concen: 10.476 ng
RT: 21.095 min Scan# 3104
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:252 Resp: 341192
Ion Ratio Lower Upper
252 100
254 63.8 51.6 77.4
126 10.8 9.7 14.5









#83
Chrysene
Concen: 11.072 ng
RT: 21.212 min Scan# 3124
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:228 Resp: 1095202
Ion Ratio Lower Upper
228 100
226 29.9 24.0 36.0
229 19.5 16.0 24.0



#84
Bis(2-ethylhexyl)phthalate
Concen: 10.448 ng
RT: 21.083 min Scan# 3102
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:149 Resp: 554310
Ion Ratio Lower Upper
149 100
167 27.5 19.7 29.5
279 3.7 2.6 4.0









#85
Di-n-octyl phthalate
Concen: 9.446 ng
RT: 21.930 min Scan# 3246
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:149 Resp: 899517
Ion Ratio Lower Upper
149 100
167 1.6 1.2 1.8
43 9.3 10.4 15.6#



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.336 min Scan# 3485
Delta R.T. 0.000 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:264 Resp: 1563611
Ion Ratio Lower Upper
264 100
260 24.0 19.3 28.9
265 21.9 18.3 27.5









#87
Indeno(1,2,3-cd)pyrene
Concen: 9.868 ng
RT: 25.465 min Scan# 3847
Delta R.T. -0.012 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:276 Resp: 1029564
Ion Ratio Lower Upper
276 100
138 26.5 18.4 27.6
277 25.7 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 11.718 ng
RT: 22.689 min Scan# 3375
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:252 Resp: 1030539
Ion Ratio Lower Upper
252 100
253 22.4 18.0 27.0
125 10.2 7.5 11.3









#89
Benzo(k)fluoranthene
Concen: 11.144 ng
RT: 22.730 min Scan# 3382
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:252 Resp: 996120
Ion Ratio Lower Upper
252 100
253 22.2 17.3 25.9
125 10.0 7.5 11.3



#90
Benzo(a)pyrene
Concen: 10.756 ng
RT: 23.236 min Scan# 3468
Delta R.T. -0.006 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:252 Resp: 807066
Ion Ratio Lower Upper
252 100
253 21.4 17.4 26.2
125 11.4 8.1 12.1









#91
Dibenzo(a,h)anthracene
Concen: 10.108 ng
RT: 25.465 min Scan# 3847
Delta R.T. -0.018 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Instrument :
BNA_M
ClientSampleId :
SSTDIC010

Tgt Ion:278 Resp: 874157
Ion Ratio Lower Upper
278 100
139 17.4 10.9 16.3#
279 24.4 18.2 27.2



#92
Benzo(g,h,i)perylene
Concen: 9.419 ng
RT: 26.124 min Scan# 3959
Delta R.T. -0.012 min
Lab File: BM032308.D
Acq: 01 Oct 2021 10:31

Tgt Ion:276 Resp: 843732
Ion Ratio Lower Upper
276 100
277 23.6 19.4 29.0
138 23.1 14.6 22.0#

