

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100221\
 Data File : BM032309.D
 Acq On : 01 Oct 2021 11:07
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC020

Quant Time: Oct 01 12:23:57 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM100221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 12:21:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.637	152	341022	20.000	ng	0.00
21) Naphthalene-d8	10.431	136	1309978	20.000	ng	# 0.00
39) Acenaphthene-d10	14.289	164	797903	20.000	ng	0.00
64) Phenanthrene-d10	17.012	188	1557882	20.000	ng	0.00
76) Chrysene-d12	21.177	240	1297051	20.000	ng	0.00
86) Perylene-d12	23.330	264	1148246	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.213	112	838404	41.545	ng	0.00
7) Phenol-d6	6.825	99	1139637	38.884	ng	0.00
23) Nitrobenzene-d5	8.795	82	962251	41.282	ng	0.00
42) 2,4,6-Tribromophenol	15.771	330	357617	41.287	ng	0.00
45) 2-Fluorobiphenyl	12.901	172	2206416	44.691	ng	0.00
79) Terphenyl-d14	19.624	244	2812369	50.172	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.013	88	175140	20.766	ng	90
3) Pyridine	3.425	79	471228	19.393	ng	94
4) n-Nitrosodimethylamine	3.337	42	192304	19.397	ng	# 62
6) Aniline	6.966	93	619283	18.805	ng	91
8) 2-Chlorophenol	7.213	128	451506	20.058	ng	91
9) Benzaldehyde	6.766	77	330010	19.444	ng	85
10) Phenol	6.854	94	543673	19.253	ng	82
11) bis(2-Chloroethyl)ether	7.060	93	436572	19.396	ng	91
12) 1,3-Dichlorobenzene	7.531	146	511499	20.495	ng	# 91
13) 1,4-Dichlorobenzene	7.678	146	518863	20.411	ng	97
14) 1,2-Dichlorobenzene	7.995	146	496232	20.074	ng	96
15) Benzyl Alcohol	7.884	79	379962	18.822	ng	# 87
16) 2,2'-oxybis(1-Chloropr...	8.172	45	550761	18.394	ng	98
17) 2-Methylphenol	8.101	107	365855	19.229	ng	# 88
18) Hexachloroethane	8.731	117	173377	20.482	ng	93
19) n-Nitroso-di-n-propyla...	8.442	70	308677	18.652	ng	# 85
20) 3+4-Methylphenols	8.425	107	492698	18.816	ng	93
22) Acetophenone	8.454	105	652545	20.318	ng	# 88
24) Nitrobenzene	8.836	77	462014	20.402	ng	# 89
25) Isophorone	9.354	82	779488	19.538	ng	96
26) 2-Nitrophenol	9.548	139	206373	20.382	ng	# 73
27) 2,4-Dimethylphenol	9.619	122	360889	20.215	ng	94
28) bis(2-Chloroethoxy)met...	9.836	93	569878	19.704	ng	100
29) 2,4-Dichlorophenol	10.089	162	400990	20.534	ng	95
30) 1,2,4-Trichlorobenzene	10.301	180	460646	21.269	ng	98
31) Naphthalene	10.478	128	1367151	20.365	ng	99
32) Benzoic acid	9.748	122	210913	16.496	ng	# 82
33) 4-Chloroaniline	10.583	127	557181	19.513	ng	# 93
34) Hexachlorobutadiene	10.789	225	274239	21.730	ng	96
35) Caprolactam	11.319	113	113466	15.918	ng	# 82
36) 4-Chloro-3-methylphenol	11.725	107	423029	19.366	ng	96
37) 2-Methylnaphthalene	12.319	142	903414	19.503	ng	94
38) 1-Methylnaphthalene	12.319	142	903795	19.927	ng	# 97
40) 1,2,4,5-Tetrachloroben...	12.477	216	477651	21.401	ng	# 95
41) Hexachlorocyclopentadiene	12.472	237	223730	22.871	ng	97
43) 2,4,6-Trichlorophenol	12.719	196	323334	20.890	ng	95

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100221\
 Data File : BM032309.D
 Acq On : 01 Oct 2021 11:07
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC020


Quant Time: Oct 01 12:23:57 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM100221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 12:21:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.789	196	344646	20.573	ng	# 90
46) 1,1'-Biphenyl	13.113	154	1225212	20.914	ng	99
47) 2-Chloronaphthalene	13.154	162	922718	20.678	ng	98
48) 2-Nitroaniline	13.354	65	238290	19.356	ng	# 80
49) Acenaphthylene	14.007	152	1429313	20.546	ng	100
50) Dimethylphthalate	13.736	163	1136301	20.364	ng	99
51) 2,6-Dinitrotoluene	13.848	165	225674	20.300	ng	# 68
52) Acenaphthene	14.348	154	944070	20.326	ng	97
53) 3-Nitroaniline	14.183	138	250314	19.235	ng	85
54) 2,4-Dinitrophenol	14.383	184	109953	16.950	ng	# 63
55) Dibenzofuran	14.683	168	1439320	20.399	ng	92
56) 4-Nitrophenol	14.501	139	205771	18.605	ng	# 65
57) 2,4-Dinitrotoluene	14.636	165	291772	19.734	ng	# 62
58) Fluorene	15.330	166	1074388	20.359	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.913	232	293493	20.084	ng	# 83
60) Diethylphthalate	15.101	149	1097241	19.893	ng	95
61) 4-Chlorophenyl-phenyle...	15.318	204	558906	20.933	ng	92
62) 4-Nitroaniline	15.342	138	245694	18.471	ng	# 67
63) Azobenzene	15.613	77	1063241	19.943	ng	85
65) 4,6-Dinitro-2-methylph...	15.401	198	167493	18.777	ng	# 79
66) n-Nitrosodiphenylamine	15.536	169	961004	21.102	ng	97
67) 4-Bromophenyl-phenylether	16.207	248	333237	21.427	ng	93
68) Hexachlorobenzene	16.342	284	365815	21.261	ng	# 83
69) Atrazine	16.471	200	316958	20.764	ng	93
70) Pentachlorophenol	16.677	266	216197	20.310	ng	99
71) Phenanthrene	17.054	178	1696329	20.613	ng	100
72) Anthracene	17.142	178	1649922	20.691	ng	99
73) Carbazole	17.407	167	1623147	19.946	ng	98
74) Di-n-butylphthalate	17.965	149	1705038	20.610	ng	# 97
75) Fluoranthene	19.059	202	1827760	19.627	ng	97
77) Benzidine	19.242	184	650240	20.304	ng	98
78) Pyrene	19.424	202	1882968	23.309	ng	99
80) Butylbenzylphthalate	20.312	149	655332	21.245	ng	91
81) Benzo(a)anthracene	21.159	228	1676687	21.013	ng	99
82) 3,3'-Dichlorobenzidine	21.095	252	522429	20.400	ng	99
83) Chrysene	21.212	228	1621487	20.847	ng	100
84) Bis(2-ethylhexyl)phtha...	21.083	149	891346	21.366	ng	95
85) Di-n-octyl phthalate	21.930	149	1388582	18.544	ng	# 91
87) Indeno(1,2,3-cd)pyrene	25.465	276	1473531	19.233	ng	96
88) Benzo(b)fluoranthene	22.694	252	1435845	22.233	ng	99
89) Benzo(k)fluoranthene	22.730	252	1439123	21.924	ng	99
90) Benzo(a)pyrene	23.241	252	1155460	20.969	ng	99
91) Dibenzo(a,h)anthracene	25.471	278	1243344	19.577	ng	# 96
92) Benzo(g,h,i)perylene	26.124	276	1210326	18.398	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 01 12:23:57 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM100221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 01 12:21:14 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.637 min Scan# 816
Delta R.T. -0.005 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:152 Resp: 341022
Ion Ratio Lower Upper
152 100
150 155.1 122.4 183.6
115 59.0 53.8 80.8



#2
1,4-Dioxane
Concen: 20.766 ng
RT: 3.013 min Scan# 30
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion: 88 Resp: 175140
Ion Ratio Lower Upper
88 100
58 73.5 50.4 75.6
43 26.8 22.2 33.2









#3
Pyridine
Concen: 19.393 ng
RT: 3.425 min Scan# 100
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion: 79 Resp: 471228
Ion Ratio Lower Upper
79 100
52 67.7 48.6 73.0
51 34.1 27.4 41.2



#4
n-Nitrosodimethylamine
Concen: 19.397 ng
RT: 3.337 min Scan# 85
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion: 42 Resp: 192304
Ion Ratio Lower Upper
42 100
74 141.1 99.5 149.3
44 7.4 57.3 85.9#









#5
2-Fluorophenol
Concen: 41.545 ng
RT: 5.213 min Scan# 404
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:112 Resp: 838404
Ion Ratio Lower Upper
112 100
64 54.3 53.7 80.5
63 29.0 33.9 50.9#



#6
Aniline
Concen: 18.805 ng
RT: 6.966 min Scan# 702
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion: 93 Resp: 619283
Ion Ratio Lower Upper
93 100
66 38.0 35.3 52.9
65 19.0 18.7 28.1









#7
Phenol-d6
Concen: 38.884 ng
RT: 6.825 min Scan# 678
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion: 99 Resp: 1139637
Ion Ratio Lower Upper
99 100
42 16.7 23.0 34.6#
71 32.9 37.8 56.8#



#8
2-Chlorophenol
Concen: 20.058 ng
RT: 7.213 min Scan# 744
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion: 128 Resp: 451506
Ion Ratio Lower Upper
128 100
130 32.1 13.0 53.0
64 43.8 33.4 73.4









#9
Benzaldehyde
Concen: 19.444 ng
RT: 6.766 min Scan# 668
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion: 77 Resp: 330010
Ion Ratio Lower Upper
77 100
105 100.3 68.9 108.9
106 97.0 61.2 101.2



#10
Phenol
Concen: 19.253 ng
RT: 6.854 min Scan# 683
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion: 94 Resp: 543673
Ion Ratio Lower Upper
94 100
65 26.9 16.9 56.9
66 36.7 30.0 70.0









#11
bis(2-Chloroethyl)ether
Concen: 19.396 ng
RT: 7.060 min Scan# 718
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion: 93 Resp: 436572
Ion Ratio Lower Upper
93 100
63 73.3 63.7 103.7
95 33.2 12.4 52.4



#12
1,3-Dichlorobenzene
Concen: 20.495 ng
RT: 7.531 min Scan# 798
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:146 Resp: 511499
Ion Ratio Lower Upper
146 100
148 64.2 52.3 78.5
75 29.8 34.2 51.2#









#13
1,4-Dichlorobenzene
Concen: 20.411 ng
RT: 7.678 min Scan# 823
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:146 Resp: 518863
Ion Ratio Lower Upper
146 100
148 64.7 51.1 76.7
111 40.0 35.3 52.9



#14
1,2-Dichlorobenzene
Concen: 20.074 ng
RT: 7.995 min Scan# 877
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:146 Resp: 496232
Ion Ratio Lower Upper
146 100
148 64.3 51.7 77.5
111 42.5 38.2 57.2









#15
Benzyl Alcohol
Concen: 18.822 ng
RT: 7.884 min Scan# 858
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion: 79 Resp: 379962
Ion Ratio Lower Upper
79 100
108 80.7 53.6 80.4#
77 64.4 57.7 86.5



#16
2,2'-oxybis(1-Chloropropane)
Concen: 18.394 ng
RT: 8.172 min Scan# 907
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion: 45 Resp: 550761
Ion Ratio Lower Upper
45 100
77 15.1 0.0 36.1
79 11.3 0.0 30.5









#17
2-Methylphenol
Concen: 19.229 ng
RT: 8.101 min Scan# 895
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:107 Resp: 365855
Ion Ratio Lower Upper
107 100
108 113.8 87.9 131.9
77 43.5 48.2 72.2#
79 43.6 44.7 67.1#



#18
Hexachloroethane
Concen: 20.482 ng
RT: 8.731 min Scan# 1002
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:117 Resp: 173377
Ion Ratio Lower Upper
117 100
119 96.6 77.7 116.5
201 100.3 70.2 105.2









#19
n-Nitroso-di-n-propylamine
Concen: 18.652 ng
RT: 8.442 min Scan# 953
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion: 70 Resp: 308677
Ion Ratio Lower Upper
70 100
42 55.3 56.3 84.5#
101 10.2 7.8 11.6
130 20.8 14.2 21.4



#20
3+4-Methylphenols
Concen: 18.816 ng
RT: 8.425 min Scan# 950
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion: 107 Resp: 492698
Ion Ratio Lower Upper
107 100
108 87.6 67.6 107.6
77 32.1 21.9 61.9
79 27.9 15.3 55.3









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.431 min Scan# 1291
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:136 Resp: 1309978
Ion Ratio Lower Upper
136 100
137 10.8 8.7 13.1
54 7.6 7.4 11.0
68 6.1 3.8 5.8#



#22
Acetophenone
Concen: 20.318 ng
RT: 8.454 min Scan# 955
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:105 Resp: 652545
Ion Ratio Lower Upper
105 100
71 5.1 3.8 5.6
51 25.9 29.5 44.3#
120 22.1 16.3 24.5









#23
Nitrobenzene-d5
Concen: 41.282 ng
RT: 8.795 min Scan# 1013
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion: 82 Resp: 962251
Ion Ratio Lower Upper
82 100
128 44.9 29.0 43.6#
54 49.2 45.8 68.8



#24
Nitrobenzene
Concen: 20.402 ng
RT: 8.836 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion: 77 Resp: 462014
Ion Ratio Lower Upper
77 100
123 46.3 30.1 45.1#
65 13.4 11.7 17.5









#25
Isophorone
Concen: 19.538 ng
RT: 9.354 min Scan# 1108
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion: 82 Resp: 779488
Ion Ratio Lower Upper
82 100
95 7.3 6.4 9.6
138 17.6 12.4 18.6



#26
2-Nitrophenol
Concen: 20.382 ng
RT: 9.548 min Scan# 1141
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:139 Resp: 206373
Ion Ratio Lower Upper
139 100
109 30.2 34.8 52.2#
65 49.4 59.9 89.9#









#27
2,4-Dimethylphenol
Concen: 20.215 ng
RT: 9.619 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:122 Resp: 360889
Ion Ratio Lower Upper
122 100
107 115.5 100.6 150.8
121 59.2 47.3 70.9



#28
bis(2-Chloroethoxy)methane
Concen: 19.704 ng
RT: 9.836 min Scan# 1190
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion: 93 Resp: 569878
Ion Ratio Lower Upper
93 100
95 32.2 25.8 38.8
123 11.2 8.8 13.2









#29
2,4-Dichlorophenol
Concen: 20.534 ng
RT: 10.089 min Scan# 1233
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:162 Resp: 400990
Ion Ratio Lower Upper
162 100
164 64.4 47.4 87.4
98 35.4 18.9 58.9



#30
1,2,4-Trichlorobenzene
Concen: 21.269 ng
RT: 10.301 min Scan# 1269
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:180 Resp: 460646
Ion Ratio Lower Upper
180 100
182 95.0 75.5 113.3
145 30.0 26.2 39.2









#31
Naphthalene
Concen: 20.365 ng
RT: 10.478 min Scan# 1299
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:128 Resp: 1367151
Ion Ratio Lower Upper
128 100
129 10.9 8.6 13.0
127 13.3 10.4 15.6



#32
Benzoic acid
Concen: 16.496 ng
RT: 9.748 min Scan# 1175
Delta R.T. -0.047 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:122 Resp: 210913
Ion Ratio Lower Upper
122 100
105 123.8 120.3 160.3
77 93.0 100.0 140.0#









#33
4-Chloroaniline
Concen: 19.513 ng
RT: 10.583 min Scan# 1317
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:127	Resp: 557181
Ion Ratio	Lower Upper
127 100	
129 31.1	25.6 38.4
65 29.8	30.4 45.6#
92 19.1	16.2 24.4



#34
Hexachlorobutadiene
Concen: 21.730 ng
RT: 10.789 min Scan# 1352
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:225	Resp: 274239
Ion Ratio	Lower Upper
225 100	
223 61.0	52.6 78.8
227 63.1	51.2 76.8









#35
Caprolactam
Concen: 15.918 ng
RT: 11.319 min Scan# 1442
Delta R.T. -0.029 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:113 Resp: 113466
Ion Ratio Lower Upper
113 100
55 165.3 170.5 210.5#
56 122.7 125.2 165.2#



#36
4-Chloro-3-methylphenol
Concen: 19.366 ng
RT: 11.725 min Scan# 1511
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:107 Resp: 423029
Ion Ratio Lower Upper
107 100
144 26.5 19.3 28.9
142 78.8 60.5 90.7









#37
2-Methylnaphthalene
Concen: 19.503 ng
RT: 12.319 min Scan# 1612
Delta R.T. 0.224 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:142 Resp: 903414
Ion Ratio Lower Upper
142 100
141 91.4 71.8 107.6
115 36.4 36.3 54.5



#38
1-Methylnaphthalene
Concen: 19.927 ng
RT: 12.319 min Scan# 1612
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:142 Resp: 903795
Ion Ratio Lower Upper
142 100
141 91.4 75.6 113.4
116 3.7 4.0 6.0#









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.289 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:164 Resp: 797903
Ion Ratio Lower Upper
164 100
162 99.9 80.6 121.0
160 43.7 35.1 52.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 21.401 ng
RT: 12.477 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:216 Resp: 477651
Ion Ratio Lower Upper
216 100
214 79.0 62.5 93.7
179 20.8 18.7 28.1
108 18.9 21.0 31.4#









#41
Hexachlorocyclopentadiene
Concen: 22.871 ng
RT: 12.472 min Scan# 1638
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:237 Resp: 223730
Ion Ratio Lower Upper
237 100
235 62.1 44.4 84.4
272 12.1 0.0 31.4



#42
2,4,6-Tribromophenol
Concen: 41.287 ng
RT: 15.771 min Scan# 2199
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:330 Resp: 357617
Ion Ratio Lower Upper
330 100
332 97.0 75.7 113.5
141 37.6 50.7 76.1#









#43
2,4,6-Trichlorophenol
Concen: 20.890 ng
RT: 12.719 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:196 Resp: 323334
Ion Ratio Lower Upper
196 100
198 96.0 73.6 110.4
200 30.5 22.1 33.1



#44
2,4,5-Trichlorophenol
Concen: 20.573 ng
RT: 12.789 min Scan# 1692
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:196 Resp: 344646
Ion Ratio Lower Upper
196 100
198 95.4 77.0 115.4
97 47.6 53.9 80.9#
132 26.3 23.8 35.8









#45
2-Fluorobiphenyl
Concen: 44.691 ng
RT: 12.901 min Scan# 1711
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:172 Resp: 2206416
Ion Ratio Lower Upper
172 100
171 35.5 28.1 42.1
170 23.5 18.6 27.8



#46
1,1'-Biphenyl
Concen: 20.914 ng
RT: 13.113 min Scan# 1747
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:154 Resp: 1225212
Ion Ratio Lower Upper
154 100
153 39.6 20.3 60.3
76 12.5 0.0 33.5









#47
2-Chloronaphthalene
Concen: 20.678 ng
RT: 13.154 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:162 Resp: 922718
Ion Ratio Lower Upper
162 100
127 40.0 33.2 49.8
164 32.9 25.7 38.5



#48
2-Nitroaniline
Concen: 19.356 ng
RT: 13.354 min Scan# 1788
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion: 65 Resp: 238290
Ion Ratio Lower Upper
65 100
92 72.0 51.2 76.8
138 109.1 66.8 100.2#









#49
Acenaphthylene
Concen: 20.546 ng
RT: 14.007 min Scan# 1899
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:152 Resp: 1429313
Ion Ratio Lower Upper
152 100
151 19.6 15.7 23.5
153 13.1 10.6 15.8



#50
Dimethylphthalate
Concen: 20.364 ng
RT: 13.736 min Scan# 1853
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:163 Resp: 1136301
Ion Ratio Lower Upper
163 100
194 4.0 2.9 4.3
164 10.2 8.0 12.0









#51
2,6-Dinitrotoluene
Concen: 20.300 ng
RT: 13.848 min Scan# 1872
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:165 Resp: 225674
Ion Ratio Lower Upper
165 100
63 57.4 72.4 108.6#
89 59.8 69.3 103.9#



#52
Acenaphthene
Concen: 20.326 ng
RT: 14.348 min Scan# 1957
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:154 Resp: 944070
Ion Ratio Lower Upper
154 100
153 111.6 92.4 138.6
152 53.7 43.7 65.5









#53
3-Nitroaniline
Concen: 19.235 ng
RT: 14.183 min Scan# 1929
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:138 Resp: 250314
Ion Ratio Lower Upper
138 100
108 10.8 8.3 12.5
92 112.0 105.2 157.8



#54
2,4-Dinitrophenol
Concen: 16.950 ng
RT: 14.383 min Scan# 1963
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:184 Resp: 109953
Ion Ratio Lower Upper
184 100
63 62.3 93.8 140.6#
154 63.6 64.0 96.0#









#55
Dibenzenofuran
Concen: 20.399 ng
RT: 14.683 min Scan# 2014
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:168 Resp: 1439320
Ion Ratio Lower Upper
168 100
139 37.5 35.0 52.4
169 13.1 10.8 16.2



#56
4-Nitrophenol
Concen: 18.605 ng
RT: 14.501 min Scan# 1983
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:139 Resp: 205771
Ion Ratio Lower Upper
139 100
109 72.3 92.3 132.3#
65 91.1 110.5 150.5#









#57
2,4-Dinitrotoluene
Concen: 19.734 ng
RT: 14.636 min Scan# 2006
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:165 Resp: 291772
Ion Ratio Lower Upper
165 100
63 43.7 70.3 105.5#
89 72.1 83.1 124.7#
182 2.9 2.2 3.4



#58
Fluorene
Concen: 20.359 ng
RT: 15.330 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:166 Resp: 1074388
Ion Ratio Lower Upper
166 100
165 97.7 79.3 118.9
167 13.1 10.6 16.0









#59
2,3,4,6-Tetrachlorophenol
Concen: 20.084 ng
RT: 14.913 min Scan# 2053
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:232 Resp: 293493
Ion Ratio Lower Upper
232 100
131 45.5 50.6 75.8#
130 2.4 3.0 4.4#
166 31.1 28.8 43.2



#60
Diethylphthalate
Concen: 19.893 ng
RT: 15.101 min Scan# 2085
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:149 Resp: 1097241
Ion Ratio Lower Upper
149 100
177 23.2 15.7 23.5
150 12.3 9.6 14.4









#61
4-Chlorophenyl-phenylether
Concen: 20.933 ng
RT: 15.318 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:204 Resp: 558906
Ion Ratio Lower Upper
204 100
206 32.5 26.7 40.1
141 60.2 55.8 83.8



#62
4-Nitroaniline
Concen: 18.471 ng
RT: 15.342 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:138 Resp: 245694
Ion Ratio Lower Upper
138 100
92 48.5 40.6 80.6
108 77.1 103.8 143.8#









#63
Azobenzene
Concen: 19.943 ng
RT: 15.613 min Scan# 2172
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Instrument :
BNA_M
ClientSampleId :
SSTDIC020

Tgt Ion: 77 Resp: 1063241




Ion	Ratio	Lower	Upper
77	100		
182	26.4	0.0	39.3
105	21.8	0.0	36.1
51	29.3	19.1	59.1




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.012 min Scan# 2410
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion: 188 Resp: 1557882

Ion	Ratio	Lower	Upper
188	100		
94	9.4	6.3	9.5
80	10.5	7.3	10.9









#65
4,6-Dinitro-2-methylphenol
Concen: 18.777 ng
RT: 15.401 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:198 Resp: 167493
Ion Ratio Lower Upper
198 100
51 46.9 49.7 89.7#
105 46.5 34.9 74.9



#66
n-Nitrosodiphenylamine
Concen: 21.102 ng
RT: 15.536 min Scan# 2159
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:169 Resp: 961004
Ion Ratio Lower Upper
169 100
168 66.3 55.0 82.4
167 35.9 29.8 44.8









#67
4-Bromophenyl-phenylether
Concen: 21.427 ng
RT: 16.207 min Scan# 2273
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:248 Resp: 333237
Ion Ratio Lower Upper
248 100
250 96.9 77.4 116.2
141 78.7 74.1 111.1



#68
Hexachlorobenzene
Concen: 21.261 ng
RT: 16.342 min Scan# 2296
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:284 Resp: 365815
Ion Ratio Lower Upper
284 100
142 36.6 43.0 64.6#
249 31.4 27.6 41.4





#69

Atrazine

Concen: 20.764 ng

RT: 16.471 min Scan# 2318

Delta R.T. -0.006 min

Lab File: BM032309.D

Acq: 01 Oct 2021 11:07

Tgt Ion:200 Resp: 316958

Ion Ratio Lower Upper

200 100

173 28.4 5.8 45.8





215 48.7 23.6 63.6

Instrument :

BNA_M

ClientSampleId :

SSTDIC020



#70

Pentachlorophenol

Concen: 20.310 ng

RT: 16.677 min Scan# 2353

Delta R.T. 0.000 min

Lab File: BM032309.D

Acq: 01 Oct 2021 11:07




Tgt Ion:266 Resp: 216197


Ion Ratio Lower Upper

266 100

268 62.8 50.3 75.5

264 63.5 52.1 78.1









#71
Phenanthrene
Concen: 20.613 ng
RT: 17.054 min Scan# 2417
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:178 Resp: 1696329
Ion Ratio Lower Upper
178 100
176 19.6 15.6 23.4
179 15.4 12.4 18.6



#72
Anthracene
Concen: 20.691 ng
RT: 17.142 min Scan# 2432
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:178 Resp: 1649922
Ion Ratio Lower Upper
178 100
176 18.8 14.8 22.2
179 15.7 12.3 18.5









#73
 Carbazole
 Concen: 19.946 ng
 RT: 17.407 min Scan# 2477
 Delta R.T. -0.006 min
 Lab File: BM032309.D
 Acq: 01 Oct 2021 11:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC020


Tgt Ion:167 Resp: 1623147
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.4 26.0
 139 13.1 11.9 17.9



#74
 Di-n-butylphthalate
 Concen: 20.610 ng
 RT: 17.965 min Scan# 2572
 Delta R.T. -0.006 min
 Lab File: BM032309.D
 Acq: 01 Oct 2021 11:07

Tgt Ion:149 Resp: 1705038
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.3 10.9
 104 5.1 5.6 8.4#









#75
Fluoranthene
Concen: 19.627 ng
RT: 19.059 min Scan# 2758
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:202 Resp: 1827760
Ion Ratio Lower Upper
202 100
101 11.8 0.0 29.9
203 17.6 0.0 36.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.177 min Scan# 3118
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:240 Resp: 1297051
Ion Ratio Lower Upper
240 100
120 11.3 7.8 11.6
236 25.9 20.9 31.3









#77
Benzidine
Concen: 20.304 ng
RT: 19.242 min Scan# 2789
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:184 Resp: 650240
Ion Ratio Lower Upper
184 100
185 14.3 10.6 16.0
183 11.5 9.0 13.4



#78
Pyrene
Concen: 23.309 ng
RT: 19.424 min Scan# 2820
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:202 Resp: 1882968
Ion Ratio Lower Upper
202 100
200 21.2 16.6 25.0
203 17.3 13.7 20.5









#79
 Terphenyl-d14
 Concen: 50.172 ng
 RT: 19.624 min Scan# 2854
 Delta R.T. -0.006 min
 Lab File: BM032309.D
 Acq: 01 Oct 2021 11:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC020


Tgt Ion:244 Resp: 2812369
 Ion Ratio Lower Upper
 244 100
 212 7.4 6.5 9.7
 122 11.8 8.7 13.1



#80
 Butylbenzylphthalate
 Concen: 21.245 ng
 RT: 20.312 min Scan# 2971
 Delta R.T. 0.000 min
 Lab File: BM032309.D
 Acq: 01 Oct 2021 11:07

Tgt Ion:149 Resp: 655332
 Ion Ratio Lower Upper
 149 100
 91 71.6 64.0 96.0
 206 20.5 14.4 21.6









#81
Benzo(a)anthracene
Concen: 21.013 ng
RT: 21.159 min Scan# 3115
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:228 Resp: 1676687
Ion Ratio Lower Upper
228 100
226 27.3 21.4 32.2
229 20.0 15.3 22.9



#82
3,3'-Dichlorobenzidine
Concen: 20.400 ng
RT: 21.095 min Scan# 3104
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:252 Resp: 522429
Ion Ratio Lower Upper
252 100
254 64.3 51.6 77.4
126 11.2 9.7 14.5





#83

Chrysene

Concen: 20.847 ng

RT: 21.212 min Scan# 3124

Delta R.T. -0.006 min

Lab File: BM032309.D

Acq: 01 Oct 2021 11:07

Tgt Ion:228 Resp: 1621487





Ion	Ratio	Lower	Upper
228	100		
226	29.8	24.0	36.0
229	19.8	16.0	24.0

Instrument :

BNA_M

ClientSampleId :

SSTDIC020



#84

Bis(2-ethylhexyl)phthalate

Concen: 21.366 ng

RT: 21.083 min Scan# 3102




Delta R.T. -0.006 min


Lab File: BM032309.D

Acq: 01 Oct 2021 11:07

Tgt Ion:149 Resp: 891346

Ion	Ratio	Lower	Upper
149	100		
167	27.5	19.7	29.5
279	3.6	2.6	4.0









#85
Di-n-octyl phthalate
Concen: 18.544 ng
RT: 21.930 min Scan# 3246
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:149 Resp: 1388582
Ion Ratio Lower Upper
149 100
167 1.7 1.2 1.8
43 9.2 10.4 15.6#



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.330 min Scan# 3484
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:264 Resp: 1148246
Ion Ratio Lower Upper
264 100
260 23.9 19.3 28.9
265 22.1 18.3 27.5









#87
Indeno(1,2,3-cd)pyrene
Concen: 19.233 ng
RT: 25.465 min Scan# 3847
Delta R.T. -0.012 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:276 Resp: 1473531
Ion Ratio Lower Upper
276 100
138 26.7 18.4 27.6
277 25.5 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 22.233 ng
RT: 22.694 min Scan# 3376
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:252 Resp: 1435845
Ion Ratio Lower Upper
252 100
253 21.9 18.0 27.0
125 9.8 7.5 11.3









#89
Benzo(k)fluoranthene
Concen: 21.924 ng
RT: 22.730 min Scan# 3382
Delta R.T. -0.006 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:252 Resp: 1439123
Ion Ratio Lower Upper
252 100
253 21.6 17.3 25.9
125 10.8 7.5 11.3



#90
Benzo(a)pyrene
Concen: 20.969 ng
RT: 23.241 min Scan# 3469
Delta R.T. 0.000 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:252 Resp: 1155460
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.2
125 10.5 8.1 12.1









#91
Dibenzo(a,h)anthracene
Concen: 19.577 ng
RT: 25.471 min Scan# 3848
Delta R.T. -0.012 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Instrument :
BNA_M
ClientSampleId :
SSTDIC020

Tgt Ion:278 Resp: 1243344
Ion Ratio Lower Upper
278 100
139 17.1 10.9 16.3#
279 23.6 18.2 27.2



#92
Benzo(g,h,i)perylene
Concen: 18.398 ng
RT: 26.124 min Scan# 3959
Delta R.T. -0.012 min
Lab File: BM032309.D
Acq: 01 Oct 2021 11:07

Tgt Ion:276 Resp: 1210326
Ion Ratio Lower Upper
276 100
277 24.0 19.4 29.0
138 22.7 14.6 22.0#

