

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM100319\
 Data File : BM022977.D
 Acq On : 05 Oct 2019 02:53
 Operator : JU
 Sample : PB123422BL
 Misc :
 ALS Vial : 63 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Oct 05 03:28:37 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM092719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 04 16:45:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	231202	20.00	ng/ul	0.00
18) Naphthalene-d8	10.58	136	1011743	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.43	164	616687	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.16	188	1091843	20.00	ng/ul	0.00
78) Chrysene-d12	21.34	240	804979	20.00	ng/ul	0.00
86) Perylene-d12	23.60	264	923077	20.00	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.26	96	27434	5.13	ng/uL	0.00
5) Phenol-d5	6.96	99	477802	23.39	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.12	67	280786	25.11	ng/ul	0.00
9) 2-Chlorophenol-d4	7.32	132	399182	24.27	ng/ul	0.00
13) 4-Methylphenol-d8	8.49	113	390396	23.34	ng/ul	-0.01
19) Nitrobenzene-d5	8.95	128	193635	24.80	ng/ul	0.00
22) 2-Nitrophenol-d4	9.66	143	209147	24.53	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.20	165	396999	23.39	ng/ul	0.00
29) 4-Chloroaniline-d4	10.71	131	560352	27.17	ng/ul	0.00
44) Dimethylphthalate-d6	13.83	166	1176417	24.58	ng/ul	0.00
47) Acenaphthylene-d8	14.12	160	1378986	24.96	ng/ul	0.00
52) 4-Nitrophenol-d4	14.62	143	143230	15.15	ng/ul	0.00
58) Fluorene-d10	15.42	176	967728	23.72	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.53	200	102194	14.38	ng/ul	0.00
71) Anthracene-d10	17.26	188	1331598	24.97	ng/ul	0.00
79) Pyrene-d10	19.56	212	1234293	28.49	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.46	264	1202381	25.18	ng/ul	-0.01

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM100319\
Data File : BM022977.D
Acq On : 05 Oct 2019 02:53
Operator : JU
Sample : PB123422BL
Misc :
ALS Vial : 63 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Oct 05 03:28:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM092719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Oct 04 16:45:59 2019
Response via : Initial Calibration

