

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100323\
 Data File : BM042160.D
 Acq On : 03 Oct 2023 12:46
 Operator : MA/JU
 Sample : 04588-07
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BBJM4

Quant Time: Oct 03 13:52:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM100223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 02 19:28:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.937	152	4188	0.400	ng/u1	0.00
4) Naphthalene-d8	10.746	136	10050	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.583	164	4535	0.400	ng/u1	# 0.00
13) Phenanthrene-d10	17.337	188	9320	0.400	ng/u1	0.00
17) Chrysene-d12	21.512	240	4870	0.400	ng/u1	0.00
23) Perylene-d12	23.912	264	5595	0.400	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.359	96	23116	4.145	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.330	152	4157	0.302	ng/u1	0.00
18) Fluoranthene-d10	19.357	212	7293	0.399	ng/u1	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.397	88	1590	0.273	ng/u1	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100323\
Data File : BM042160.D
Acq On : 03 Oct 2023 12:46
Operator : MA/JU
Sample : 04588-07
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BBJM4

Quant Time: Oct 03 13:52:35 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM100223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 02 19:28:41 2023
Response via : Initial Calibration

