

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100323\
 Data File : BM042177.D
 Acq On : 04 Oct 2023 00:27
 Operator : MA/JU
 Sample : 04480-09
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BBJY5

Quant Time: Oct 04 03:14:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM100223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 02 19:28:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.933	152	4608	0.400	ng/ul	# 0.00
4) Naphthalene-d8	10.746	136	11290	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.578	164	5711	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.333	188	12345	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.507	240	6333	0.400	ng/ul	0.00
23) Perylene-d12	23.903	264	6739	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.360	96	27110	4.418	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.324	152	4890	0.316	ng/ul	-0.01
18) Fluoranthene-d10	19.357	212	8676	0.365	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.397	88	7247	1.131	ng/ul	95
5) Naphthalene	10.796	128	289286	7.680	ng/ul	98
7) 2-Methylnaphthalene	12.401	142	8235	0.365	ng/ul	99
8) 1-Methylnaphthalene	12.621	142	71402	3.143	ng/ul	100
11) Acenaphthene	14.643	153	3152	0.115	ng/ul	94
12) Fluorene	15.628	166	5378	0.175	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100323\
Data File : BM042177.D
Acq On : 04 Oct 2023 00:27
Operator : MA/JU
Sample : 04480-09
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BBJY5

Quant Time: Oct 04 03:14:00 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM100223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 02 19:28:41 2023
Response via : Initial Calibration

