

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM100323\  
 Data File : BM042180.D  
 Acq On : 04 Oct 2023 02:17  
 Operator : MA/JU  
 Sample : PB155824BS  
 Misc :  
 ALS Vial : 27 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SLCS824

Quant Time: Oct 04 03:14:24 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\SFAM-EPA-SIM-BM100223.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Oct 02 19:28:41 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|-------|----------|
| -----                       |        |      |          |       |       |          |
| Internal Standards          |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.933  | 152  | 3943     | 0.400 | ng/ul | 0.00     |
| 4) Naphthalene-d8           | 10.746 | 136  | 9494     | 0.400 | ng/ul | 0.00     |
| 9) Acenaphthene-d10         | 14.597 | 164  | 3514     | 0.400 | ng/ul | 0.00     |
| 13) Phenanthrene-d10        | 17.354 | 188  | 7277     | 0.400 | ng/ul | 0.01     |
| 17) Chrysene-d12            | 21.512 | 240  | 4941     | 0.400 | ng/ul | 0.00     |
| 23) Perylene-d12            | 23.918 | 264  | 5178     | 0.400 | ng/ul | # 0.00   |
| System Monitoring Compounds |        |      |          |       |       |          |
| 3) 1,4-Dioxane-d8           | 3.360  | 96   | 3272     | 0.623 | ng/ul | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.346 | 152  | 3559     | 0.273 | ng/ul | 0.01     |
| 18) Fluoranthene-d10        | 19.366 | 212  | 6143     | 0.331 | ng/ul | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM100323\  
Data File : BM042180.D  
Acq On : 04 Oct 2023 02:17  
Operator : MA/JU  
Sample : PB155824BS  
Misc :  
ALS Vial : 27 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
SLCS824

Quant Time: Oct 04 03:14:24 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\SFAM-EPA-SIM-BM100223.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Oct 02 19:28:41 2023  
Response via : Initial Calibration

