

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100423\
 Data File : BM042209.D
 Acq On : 05 Oct 2023 09:31
 Operator : MA/JU
 Sample : PB155938BL
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK938

Quant Time: Oct 06 05:26:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM100223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 06 05:25:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.920	152	5810	0.400	ng/ul	0.00
4) Naphthalene-d8	10.735	136	14987	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.573	164	6133	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.329	188	13196	0.400	ng/ul	-0.01
17) Chrysene-d12	21.509	240	6576	0.400	ng/ul	0.00
23) Perylene-d12	23.906	264	6787	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.351	96	66130	8.547	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.319	152	8089	0.394	ng/ul	-0.01
18) Fluoranthene-d10	19.352	212	10949	0.443	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100423\
Data File : BM042209.D
Acq On : 05 Oct 2023 09:31
Operator : MA/JU
Sample : PB155938BL
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK938

Quant Time: Oct 06 05:26:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM100223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 06 05:25:26 2023
Response via : Initial Calibration

