


Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM100719\
 Data File : BM023030.D
 Acq On : 07 Oct 2019 11:21
 Operator : JU
 Sample : DFTPP05
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP05

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM092719MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Oct 08 13:35:36 2019



AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2479

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.2	64403	PASS
68	69	0.00	2	1.7	1220	PASS
69	198	0.00	100	35.0	69930	PASS
70	69	0.00	2	0.5	378	PASS
127	198	10	80	46.6	93216	PASS
197	198	0.00	2	0.4	878	PASS
198	198	100	100	100.0	199872	PASS
199	198	5	9	6.8	13602	PASS
275	198	10	60	28.7	57448	PASS
365	198	1	100	4.0	7998	PASS
441	198	0.01	100	13.8	27640	PASS
442	198	50	100	91.0	181981	PASS
443	442	15	24	19.3	35106	PASS

m/z	Abundance
37.00	457.0
38.00	928.0
39.10	4980.0
40.00	1422.0
41.00	476.0
44.00	2335.0
47.00	1284.0
47.90	612.0
49.00	5667.0
50.00	17096.0
51.10	68496.0
52.00	3607.0
55.00	496.0
56.00	1908.0
57.00	4633.0
61.00	766.0
62.00	1112.0
63.10	2727.0
64.00	357.0
65.00	1249.0
68.00	1223.0
69.00	69640.0
70.10	332.0
73.00	1107.0
74.00	7138.0
75.00	11622.0
76.00	3952.0
77.10	83856.0
78.10	5563.0
79.00	4960.0
80.00	4064.0
81.00	5812.0
81.90	1283.0
83.00	1420.0
84.00	4827.0
85.00	1052.0
86.00	4407.0
87.00	842.0
88.00	754.0
91.00	1194.0
92.00	1628.0
93.00	9074.0
94.00	703.0
96.00	525.0
98.00	6614.0
99.00	7167.0
100.00	561.0
101.00	3332.0
103.00	1021.0
104.00	2022.0
105.00	1750.0
106.00	548.0
107.00	24224.0
108.00	3699.0
109.00	633.0
110.00	42984.0
111.00	6852.0
111.90	926.0
116.00	1399.0
117.00	20112.0
118.00	1676.0
122.00	1465.0
123.00	2460.0
124.00	1229.0
125.00	1019.0
127.00	89896.0
128.00	6478.0
129.00	32128.0
130.00	2639.0
131.00	518.0
132.90	304.0
134.00	944.0
135.00	2786.0
136.00	1144.0
137.00	1993.0
138.00	351.0
140.00	464.0
141.00	4156.0
142.00	1288.0
142.90	1030.0
146.00	710.0
147.00	2373.0
148.00	5361.0
149.00	1234.0
150.00	415.0
151.00	456.0
151.90	427.0
153.00	1528.0
154.00	1010.0
155.00	2441.0
156.00	3831.0

Instrument :
BNA_M
ClientSampleId :
DFTPP05

157.00	731.0
157.90	810.0
158.90	532.0
160.00	1244.0
161.00	2156.0
162.00	625.0
165.00	1463.0
165.90	1382.0
167.00	9234.0
168.00	7913.0
169.00	987.0
170.90	379.0
171.90	746.0
173.00	981.0
174.00	1794.0
175.00	3412.0
176.00	850.0
177.00	1616.0
178.00	531.0
179.00	6333.0
180.00	5183.0
181.00	2255.0
182.00	368.0
183.90	443.0
185.00	3193.0
186.00	24472.0
187.00	7271.0
188.00	943.0
189.00	1480.0
189.90	345.0
190.90	1021.0
192.00	2480.0
193.00	2553.0
194.00	612.0
196.00	4935.0
197.00	816.0
198.00	176064.0
199.00	12004.0
199.90	1076.0
201.40	829.0
203.00	1179.0
204.00	6853.0
205.00	10890.0
206.00	47080.0
207.00	7880.0
208.00	1880.0
209.00	841.0
210.20	875.0
211.00	1784.0
215.00	506.0
216.00	1091.0
217.00	12234.0
218.00	1661.0
221.00	7284.0
221.80	1885.0
223.00	2788.0
224.00	26368.0
225.00	6790.0
226.00	1001.0
227.00	9981.0
228.00	1384.0
229.00	2216.0
229.90	370.0
231.00	961.0
233.90	747.0
235.00	933.0
236.00	406.0
237.00	1519.0
237.90	344.0
239.00	563.0
239.90	326.0
241.00	662.0
242.00	1536.0
243.00	1490.0
244.00	20144.0
245.00	2855.0
246.00	3932.0
247.00	814.0
248.90	821.0
252.90	455.0
253.90	602.0
255.00	105704.0
256.00	15791.0
257.00	1230.0
258.00	5777.0
259.00	1173.0
265.00	2237.0
266.00	404.0
270.90	344.0
272.00	323.0
273.00	3021.0
274.00	7862.0
275.00	45816.0
276.00	6164.0
277.00	4115.0

Instrument :
BNA_M
ClientSampleId :
DFTPP05

277.90	726.0
281.00	525.0
283.00	537.0
285.00	689.0
293.00	964.0
294.00	357.0
296.00	13642.0
297.00	2012.0
303.00	1528.0
304.00	498.0
314.00	674.0
315.00	1524.0
316.00	858.0
321.00	423.0
323.00	4186.0
324.00	905.0
327.00	983.0
328.10	331.0
332.00	300.0
333.00	396.0
334.00	2645.0
335.00	681.0
340.90	459.0
346.00	872.0
352.00	1114.0
353.10	776.0
354.10	1400.0
355.00	481.0
365.00	6147.0
366.00	989.0
372.10	1914.0
373.00	463.0
383.00	559.0
402.00	922.0
403.00	1283.0
404.00	461.0
421.00	831.0
422.00	979.0
423.10	6741.0
424.10	2470.0
424.90	333.0
441.10	18808.0
442.10	123800.0
443.10	24944.0
444.10	2426.0

Instrument :
BNA_M
ClientSampleId :
DFTPP05

m/z	Abundance
37.00	506.0
38.00	1033.0
39.00	5971.0
40.00	1471.0
41.00	560.0
44.00	2340.0
47.00	1123.0
48.00	599.0
49.00	6077.0
50.00	20480.0
51.10	79424.0
52.10	4023.0
55.00	480.0
56.00	2389.0
57.00	5335.0
61.00	1049.0
62.00	1166.0
63.00	3130.0
64.00	552.0
65.00	1702.0
68.00	1547.0
69.00	83992.0
70.00	501.0
73.00	1130.0
74.00	8207.0
75.00	13604.0
76.00	4603.0
77.10	98184.0
78.10	6600.0
79.00	6258.0
80.00	4710.0
81.00	6835.0
82.00	1628.0
83.00	1760.0
84.00	4565.0
85.00	1238.0
86.00	4953.0
87.00	1007.0
88.00	775.0
91.00	1344.0
92.00	1806.0
93.00	10583.0
94.00	802.0
96.00	730.0
98.00	8088.0
99.00	8208.0
99.90	677.0
101.00	4150.0
103.00	1189.0
104.00	2343.0
105.00	2305.0
106.00	742.0
107.00	29664.0
108.00	4824.0
109.00	697.0
110.00	55768.0
111.00	9074.0
112.00	1005.0
113.00	355.0
116.00	1693.0
117.00	25472.0
118.00	1921.0
118.90	353.0
120.00	493.0
122.00	2065.0
123.00	3307.0
124.00	1313.0
125.00	1271.0
126.00	390.0
127.00	111424.0
128.00	8119.0
129.00	41408.0
130.00	3578.0
130.90	815.0
132.90	348.0
134.00	1081.0
135.00	3611.0
136.00	1337.0
137.00	2096.0
138.00	498.0
140.00	572.0
141.00	5284.0
142.00	1652.0
143.00	1342.0
145.00	335.0
146.00	991.0
147.00	3128.0
148.00	6980.0
149.00	1774.0
150.00	558.0
151.10	676.0

Instrument :
BNA_M
ClientSampleId :
DFTPP05

152.00	422.0
153.00	1757.0
154.00	1344.0
155.00	3246.0
156.00	4757.0
157.10	896.0
157.90	1037.0
159.00	904.0
160.00	1650.0
161.00	2817.0
161.90	797.0
164.90	2180.0
166.00	1591.0
167.00	12174.0
168.00	8728.0
169.00	1041.0
170.00	334.0
170.90	507.0
172.00	961.0
173.00	1284.0
174.00	2494.0
175.00	5071.0
176.00	1259.0
177.00	2468.0
178.00	721.0
179.00	8231.0
180.00	6815.0
181.00	2963.0
182.10	363.0
183.90	715.0
185.00	4085.0
186.00	32696.0
187.00	10012.0
188.00	1198.0
189.00	1727.0
189.90	342.0
191.00	1161.0
192.00	2836.0
193.00	3431.0
194.00	685.0
194.90	349.0
196.00	6919.0
196.90	977.0
198.00	243072.0
198.90	16952.0
200.00	1413.0
201.50	1144.0
203.00	1711.0
204.00	9020.0
205.00	14897.0
206.00	63848.0
207.00	10329.0
208.00	2406.0
209.00	1073.0
210.10	992.0
211.00	2726.0
215.00	647.0
216.00	1310.0
217.00	16736.0
218.00	2453.0
220.20	412.0
221.00	10398.0
221.90	2499.0
223.00	3625.0
224.00	36312.0
225.00	9277.0
226.00	1120.0
227.00	14964.0
228.00	1978.0
229.00	3675.0
229.90	349.0
231.00	1365.0
233.90	1036.0
235.00	1014.0
236.00	826.0
237.00	1547.0
239.00	624.0
240.00	554.0
241.00	884.0
242.00	2025.0
243.00	2021.0
244.00	28120.0
245.00	4272.0
246.00	5733.0
246.90	1136.0
248.00	334.0
249.00	988.0
252.90	606.0
254.10	874.0
255.00	148608.0
256.00	22792.0
257.00	1642.0
258.00	8912.0
259.00	1362.0
265.00	3567.0

Instrument :
BNA_M
ClientSampleId :
DFTPP05

266.00	556.0
271.00	477.0
271.90	447.0
273.00	4766.0
274.00	11588.0
275.00	69624.0
276.00	8828.0
277.00	5955.0
278.00	1183.0
281.00	658.0
282.00	392.0
283.10	943.0
284.00	484.0
285.00	1160.0
293.00	1390.0
294.00	434.0
295.00	511.0
296.00	18728.0
297.00	2995.0
302.00	382.0
303.00	2359.0
304.00	551.0
310.00	343.0
314.10	1020.0
315.00	2174.0
316.00	1109.0
321.00	606.0
323.10	6709.0
324.00	1220.0
326.90	1430.0
328.00	814.0
332.00	418.0
333.00	517.0
334.00	4389.0
335.00	1294.0
341.00	789.0
346.00	1414.0
346.90	304.0
352.00	2045.0
353.10	1230.0
354.00	1861.0
355.10	540.0
365.00	9168.0
366.00	1349.0
371.00	480.0
372.00	3074.0
373.10	733.0
383.10	940.0
384.00	366.0
390.10	520.0
402.00	1363.0
403.00	1943.0
404.00	651.0
421.00	1718.0
422.00	1493.0
423.10	11464.0
424.10	3849.0
425.00	510.0
441.10	32280.0
442.10	214784.0
443.10	42168.0
444.10	3668.0

Instrument :
BNA_M
ClientSampleId :
DFTPP05

m/z	Abundance
37.00	472.0
38.00	569.0
39.10	3875.0
40.00	1348.0
41.00	573.0
44.00	2250.0
47.00	1135.0
48.00	628.0
49.00	5779.0
50.10	13247.0
51.10	51624.0
52.10	2629.0
55.00	495.0
56.00	1464.0
57.00	3762.0
61.00	711.0
62.00	650.0
63.10	2061.0
65.00	998.0
68.10	891.0
69.00	56160.0
70.00	301.0
73.10	986.0
74.00	6005.0
75.00	9328.0
76.00	3289.0
77.10	67224.0
78.10	4383.0
79.00	4095.0
80.00	3284.0
81.00	4562.0
82.00	1223.0
83.00	1337.0
84.00	4407.0
85.00	917.0
86.00	4047.0
86.90	689.0
88.00	800.0
91.00	1260.0
92.00	1366.0
93.00	7322.0
94.00	571.0
96.00	486.0
98.00	5663.0
99.00	5218.0
99.90	325.0
101.00	2811.0
103.00	748.0
104.00	1835.0
105.00	1521.0
106.00	503.0
107.00	21744.0
108.00	3093.0
109.10	564.0
110.00	38000.0
111.00	5574.0
112.00	766.0
116.00	1235.0
117.00	17464.0
118.00	1447.0
122.00	1281.0
123.00	2213.0
124.00	997.0
124.90	827.0
127.00	78328.0
128.00	5706.0
129.00	29664.0
130.00	2480.0
131.00	503.0
134.00	795.0
135.00	2526.0
135.90	939.0
137.00	1566.0
137.90	315.0
139.90	394.0
141.00	3747.0
142.00	1294.0
143.00	923.0
145.90	735.0
147.00	2400.0
148.00	4793.0
149.00	1272.0
149.90	380.0
151.10	584.0
151.90	321.0
153.00	1254.0
154.00	1075.0
155.00	2346.0
156.00	3791.0
157.00	655.0
157.90	640.0

Instrument :
BNA_M
ClientSampleId :
DFTPP05

159.00	533.0
160.00	1101.0
161.00	2014.0
162.00	529.0
164.90	1503.0
166.00	1200.0
167.00	8476.0
168.00	6134.0
168.90	733.0
170.90	311.0
171.90	761.0
173.00	1084.0
174.00	1728.0
175.00	3643.0
175.90	822.0
177.00	1600.0
178.00	430.0
179.00	6545.0
180.00	5006.0
180.90	2186.0
181.90	365.0
184.00	532.0
185.00	3167.0
186.00	25304.0
187.00	7112.0
188.00	823.0
189.00	1628.0
191.00	986.0
192.00	2044.0
193.00	2906.0
193.90	560.0
195.00	327.0
196.00	5232.0
197.00	843.0
198.00	180480.0
198.90	11852.0
200.00	913.0
201.40	711.0
202.90	1393.0
204.00	6739.0
205.00	11682.0
206.00	49192.0
207.00	7456.0
208.00	1877.0
209.00	751.0
210.20	728.0
211.00	1939.0
211.80	344.0
215.00	478.0
216.00	986.0
217.00	13021.0
218.00	1823.0
220.30	424.0
221.00	8147.0
221.90	1651.0
223.00	3122.0
224.00	28416.0
225.00	6805.0
226.00	803.0
227.00	11713.0
228.00	1443.0
229.00	2386.0
229.90	300.0
231.00	1155.0
234.00	651.0
235.00	947.0
236.00	656.0
237.00	1279.0
239.00	356.0
240.00	368.0
241.00	783.0
242.00	1598.0
243.00	1627.0
244.00	22944.0
245.00	3100.0
246.00	4269.0
247.00	814.0
249.00	963.0
253.00	397.0
254.00	686.0
255.00	123744.0
256.00	17080.0
257.00	1263.0
258.00	6707.0
259.00	979.0
265.00	2974.0
266.00	384.0
271.00	369.0
272.00	379.0
273.00	3758.0
274.00	9764.0
275.00	56904.0
276.00	7236.0
277.00	4641.0
277.90	878.0

Instrument :
BNA_M
ClientSampleId :
DFTPP05

281.00	658.0
282.10	312.0
283.00	684.0
285.00	814.0
292.10	318.0
293.00	1183.0
293.90	384.0
295.00	454.0
296.00	17344.0
297.00	2252.0
303.00	1774.0
304.00	492.0
309.90	303.0
314.00	824.0
315.00	1909.0
316.00	986.0
321.10	554.0
323.00	5727.0
324.00	1007.0
327.00	1358.0
327.90	514.0
332.00	373.0
333.00	470.0
334.00	3509.0
335.00	877.0
341.10	728.0
346.00	1384.0
352.00	1661.0
353.00	1118.0
354.10	1725.0
355.10	381.0
365.00	8679.0
366.00	1217.0
371.00	439.0
372.10	3234.0
373.10	750.0
383.00	875.0
390.00	430.0
402.00	1274.0
403.10	1888.0
404.00	620.0
421.00	1569.0
422.00	1521.0
423.10	10038.0
424.10	2852.0
425.10	388.0
441.10	31832.0
442.10	207360.0
443.10	38208.0
444.10	3825.0

Instrument :

BNA_M

ClientSampleId :

DFTPP05

m/z	Abundance
40.00	1349.0
41.10	301.0
44.00	2481.0
47.00	1267.0
48.00	607.0
49.00	5845.0
51.00	2111.0
73.00	633.0
84.00	4366.0
86.00	2856.0
87.90	507.0
207.00	1646.0
208.00	372.0
281.00	518.0

Instrument :
BNA_M
ClientSampleId :
DFTPP05