

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100721\
 Data File : BM032350.D
 Acq On : 07 Oct 2021 10:34
 Operator : CG/JU
 Sample : PB139036BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK036

Quant Time: Oct 07 11:11:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 07 09:34:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.619	152	284712	20.00	ng/ul	0.00
20) Naphthalene-d8	10.413	136	1217759	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.266	164	782497	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.995	188	1620430	20.00	ng/ul	0.00
79) Chrysene-d12	21.153	240	1551545	20.00	ng/ul	0.00
88) Perylene-d12	23.300	264	1559160	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.960	96	40590	5.72	ng/uL	0.00
4) Pyridine-d5	3.384	84	575973	28.71	ng/ul	0.00
7) Phenol-d5	6.807	99	689277	28.47	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.948	67	413567	29.75	ng/ul	0.00
11) 2-Chlorophenol-d4	7.160	132	549744	29.74	ng/ul	0.00
15) 4-Methylphenol-d8	8.348	113	545604	27.90	ng/ul	0.00
21) Nitrobenzene-d5	8.772	128	257756	30.00	ng/ul	0.00
24) 2-Nitrophenol-d4	9.501	143	276506	29.86	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.042	165	509280	27.62	ng/ul	0.00
31) 4-Chloroaniline-d4	10.542	131	777954	28.61	ng/ul	0.00
46) Dimethylphthalate-d6	13.672	166	1703725	31.46	ng/ul	0.00
49) Acenaphthylene-d8	13.960	160	2010568	30.56	ng/ul	0.00
54) 4-Nitrophenol-d4	14.471	143	268810	25.28	ng/ul	0.00
60) Fluorene-d10	15.254	176	1435467	31.26	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.371	200	217145	23.51	ng/ul	0.00
73) Anthracene-d10	17.089	188	2289010	32.45	ng/ul	0.00
81) Pyrene-d10	19.377	212	2606519	33.39	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.165	264	2423680	31.04	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100721\
Data File : BM032350.D
Acq On : 07 Oct 2021 10:34
Operator : CG/JU
Sample : PB139036BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK036

Quant Time: Oct 07 11:11:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 07 09:34:26 2021
Response via : Initial Calibration

