

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100821\
 Data File : BM032376.D
 Acq On : 08 Oct 2021 03:35
 Operator : CG/JU
 Sample : M4003-01
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C00Q0

Quant Time: Oct 08 04:55:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 07 15:38:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.622	152	6208	0.40	ng/ul	0.00
4) Naphthalene-d8	10.411	136	21627	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.266	164	12424	0.40	ng/ul	0.00
13) Phenanthrene-d10	16.995	188	25211	0.40	ng/ul	0.00
17) Chrysene-d12	21.154	240	19382	0.40	ng/ul	0.00
23) Perylene-d12	23.295	264	19286	0.40	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.965	96	34660	5.19	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.003	152	9688	0.34	ng/ul	0.00
18) Fluoranthene-d10	19.011	212	28027	0.53	ng/ul	0.00
Target Compounds						
15) Phenanthrene	17.036	178	2201	0.03	ng/ul#	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100821\
Data File : BM032376.D
Acq On : 08 Oct 2021 03:35
Operator : CG/JU
Sample : M4003-01
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C00Q0

Quant Time: Oct 08 04:55:36 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 07 15:38:40 2021
Response via : Initial Calibration

