

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100923\
 Data File : BM042223.D
 Acq On : 09 Oct 2023 13:41
 Operator : MA/JU
 Sample : 04607-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BBJF9

Quant Time: Oct 09 15:31:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM100223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 09 13:33:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.055	152	3576	0.400	ng/ul	0.00
4) Naphthalene-d8	10.878	136	9257	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.694	164	3387	0.400	ng/ul	#-0.02
13) Phenanthrene-d10	17.447	188	6965	0.400	ng/ul	#-0.03
17) Chrysene-d12	21.638	240	3238	0.400	ng/ul	#-0.02
23) Perylene-d12	24.135	264	3522	0.400	ng/ul	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.410	96	19498	4.094	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.462	152	2518	0.198	ng/ul	0.00
18) Fluoranthene-d10	19.469	212	4082	0.336	ng/ul	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.452	88	1319	0.265	ng/ul	91
20) Pyrene	19.840	202	1463	0.056	ng/ul#	1
26) Benzo(a)pyrene	23.971	252	2695	0.151	ng/ul#	51

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100923\
Data File : BM042223.D
Acq On : 09 Oct 2023 13:41
Operator : MA/JU
Sample : 04607-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BBJF9

Quant Time: Oct 09 15:31:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM100223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 09 13:33:32 2023
Response via : Initial Calibration

