

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100923\
 Data File : BM042240.D
 Acq On : 10 Oct 2023 00:59
 Operator : MA/JU
 Sample : 04654-12
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BBBT7

Quant Time: Oct 10 01:40:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM100223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 09 23:59:28 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.055	152	4876	0.400	ng/ul	0.00
4) Naphthalene-d8	10.873	136	12912	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.689	164	5460	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.443	188	11109	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.629	240	4373	0.400	ng/ul	# 0.00
23) Perylene-d12	24.126	264	4501	0.400	ng/ul	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.410	96	29081	4.479	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.456	152	5990	0.338	ng/ul	0.00
18) Fluoranthene-d10	19.464	212	7749	0.472	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.448	88	11021	1.625	ng/ul#	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 10 01:40:48 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM100223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 09 23:59:28 2023
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 1.625 ng/ul
 RT: 3.448 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BM042240.D
 Acq: 10 Oct 2023 00:59

Instrument :
 BNA_M
 ClientSampleId :
 BBBT7

Tgt Ion: 88 Resp: 11021

Ion	Ratio	Lower	Upper
88	100		
43	40.0	52.1	78.1#
58	76.7	53.8	80.6

