


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101419\
Data File : BM023147.D
Acq On : 14 Oct 2019 16:29
Operator : JU
Sample : DFTPP09
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP09

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM101419MA.M
Title : SVOA CALIBRATION
Last Update : Tue Oct 15 02:28:10 2019



AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.9	99470	PASS
68	69	0.00	2	1.7	1944	PASS
69	198	0.00	100	35.4	113653	PASS
70	69	0.00	2	0.2	193	PASS
127	198	10	80	43.5	139697	PASS
197	198	0.00	2	0.1	450	PASS
198	198	100	100	100.0	321408	PASS
199	198	5	9	6.7	21580	PASS
275	198	10	60	31.2	100152	PASS
365	198	1	100	4.2	13658	PASS
441	198	0.01	100	12.9	41458	PASS
442	198	50	100	82.6	265536	PASS
443	442	15	24	19.1	50714	PASS

m/z	Abundance
36.00	2710.0
36.90	493.0
38.00	1437.0
39.00	4987.0
40.00	2997.0
41.10	14938.0
42.10	4356.0
43.10	22576.0
44.00	20944.0
45.00	1420.0
46.00	397.0
46.90	360.0
49.00	986.0
50.10	717.0
51.10	1428.0
51.90	1079.0
53.00	2236.0
54.00	2803.0
55.10	18840.0
56.10	6216.0
57.10	23288.0
58.10	5768.0
59.10	3154.0
60.00	2012.0
61.00	567.0
63.00	527.0
63.90	651.0
65.10	1395.0
66.00	968.0
67.10	6183.0
68.10	3746.0
69.10	9056.0
70.10	5376.0
71.10	15371.0
72.10	1261.0
73.10	3060.0
74.10	587.0
75.00	720.0
75.90	401.0
77.10	1951.0
78.10	1277.0
79.10	3553.0
80.10	1595.0
81.10	5400.0
82.10	5071.0
83.10	8529.0
84.10	5070.0
85.10	10031.0
86.00	1296.0
87.10	874.0
89.00	303.0
91.00	3183.0
92.00	947.0
93.00	1905.0
94.10	1437.0
95.10	5627.0
96.00	4155.0
97.10	6841.0
98.10	8381.0
99.10	3298.0
100.00	660.0
100.80	336.0
103.10	604.0
104.00	523.0
105.10	2418.0
106.00	572.0
107.00	1857.0
108.10	561.0
109.10	2259.0
110.10	1722.0
111.10	3241.0
112.10	1920.0
113.10	1754.0
114.00	665.0
115.10	1588.0
116.00	576.0
117.10	1179.0
118.00	405.0
119.10	1364.0
120.00	738.0
121.10	1198.0
121.90	740.0
123.10	1562.0
124.00	741.0
125.00	1511.0
126.00	752.0
127.10	1764.0
128.00	1223.0
129.10	1381.0
130.10	482.0
131.00	1040.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

132.00	309.0
133.00	1386.0
133.90	533.0
135.10	885.0
136.00	423.0
137.10	1042.0
138.00	824.0
139.10	438.0
140.10	340.0
141.00	1168.0
142.00	680.0
143.00	760.0
144.00	379.0
145.00	1200.0
146.10	506.0
147.00	758.0
148.10	369.0
149.00	711.0
151.00	661.0
152.10	991.0
153.00	559.0
154.20	405.0
155.10	885.0
156.00	443.0
157.00	835.0
158.10	512.0
159.00	429.0
160.90	457.0
163.00	359.0
165.10	896.0
166.10	565.0
167.10	500.0
169.20	727.0
170.00	339.0
171.10	582.0
177.10	370.0
178.00	446.0
179.00	424.0
180.10	446.0
181.10	488.0
183.00	576.0
185.10	499.0
189.00	532.0
191.00	760.0
192.10	383.0
193.00	846.0
194.20	673.0
195.20	527.0
196.00	311.0
197.20	427.0
198.90	300.0
203.10	339.0
205.10	322.0
207.00	3486.0
208.00	1006.0
209.00	694.0
210.10	322.0
213.10	432.0
221.10	496.0
222.10	601.0
223.20	340.0
236.10	404.0
239.20	3319.0
240.20	637.0
241.10	376.0
253.20	478.0
254.20	532.0
255.20	2006.0
256.20	447.0
265.10	342.0
267.30	4290.0
268.20	1026.0
281.10	984.0
282.30	623.0
283.30	2235.0
284.30	707.0
295.40	427.0
309.30	366.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

m/z	Abundance
35.10	546.0
36.00	2149.0
37.00	1264.0
38.00	2791.0
39.10	12760.0
40.00	3274.0
41.10	14737.0
42.10	4368.0
43.10	22344.0
44.00	18736.0
45.00	1512.0
45.90	554.0
49.10	1461.0
50.00	30760.0
51.10	119112.0
52.00	6572.0
53.10	2496.0
54.10	2422.0
55.10	16904.0
56.10	9913.0
57.10	30824.0
58.10	6023.0
59.00	2602.0
60.10	2235.0
61.00	2034.0
62.00	2102.0
63.10	5087.0
64.10	960.0
65.10	3766.0
66.00	1015.0
67.00	5480.0
68.10	6764.0
69.00	143168.0
70.10	5572.0
71.10	16488.0
72.00	1312.0
73.10	3326.0
74.00	14299.0
75.00	20920.0
76.10	7603.0
77.10	148224.0
78.10	10804.0
79.00	12840.0
80.00	8692.0
81.10	16100.0
82.00	6687.0
83.10	10274.0
84.10	4630.0
85.10	11355.0
86.00	4295.0
87.00	2041.0
88.00	840.0
89.00	565.0
91.00	5936.0
92.00	3558.0
93.00	18112.0
94.00	2007.0
95.10	5145.0
96.10	5023.0
97.10	6873.0
98.10	19136.0
99.00	12386.0
100.10	1859.0
101.00	6335.0
101.90	650.0
103.00	2351.0
104.00	4673.0
105.00	4796.0
106.10	2161.0
107.00	47480.0
108.00	7659.0
109.10	4335.0
110.00	87608.0
111.00	15153.0
112.00	3096.0
113.10	2301.0
114.10	943.0
115.10	1414.0
116.00	3002.0
117.00	36152.0
118.00	2971.0
119.00	1924.0
120.00	1103.0
121.00	1570.0
122.00	3403.0
123.00	5531.0
124.00	2541.0
125.00	2675.0
127.00	158976.0
128.00	12312.0
129.00	63400.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

130.00	5513.0
131.00	1937.0
132.00	901.0
133.00	1215.0
134.00	2519.0
135.00	5762.0
136.00	2613.0
137.00	3322.0
138.00	1182.0
139.00	911.0
140.10	1536.0
141.00	9452.0
142.00	3162.0
143.00	2310.0
144.10	831.0
145.00	1248.0
146.10	1712.0
147.00	4290.0
148.00	10103.0
149.00	2249.0
150.00	631.0
151.10	1925.0
152.10	1270.0
153.00	3411.0
154.00	2302.0
155.00	5353.0
156.00	8108.0
157.10	2152.0
158.00	1417.0
159.10	1538.0
160.00	3106.0
161.00	4227.0
162.00	892.0
163.10	722.0
164.00	802.0
165.00	4158.0
166.00	3056.0
167.00	16688.0
168.00	7613.0
169.00	2476.0
170.10	792.0
171.00	1498.0
172.00	1535.0
172.90	2275.0
174.00	3503.0
175.10	6742.0
175.90	1894.0
177.00	2924.0
178.00	1693.0
178.90	13302.0
180.00	8434.0
181.00	4570.0
182.10	803.0
183.10	1048.0
184.00	1123.0
185.00	6706.0
186.00	47288.0
187.00	13499.0
188.00	1911.0
189.00	3531.0
190.00	668.0
191.00	2071.0
192.00	4081.0
193.10	5951.0
194.10	1214.0
195.10	1103.0
196.00	10586.0
197.90	341440.0
199.00	22848.0
200.00	1950.0
201.30	1588.0
202.20	503.0
203.00	2235.0
204.00	13690.0
205.00	20432.0
206.00	90208.0
207.00	15747.0
208.10	4438.0
209.00	1424.0
210.10	1648.0
211.00	4026.0
212.10	644.0
213.00	815.0
215.00	1412.0
216.00	2312.0
217.00	24112.0
218.00	3277.0
219.00	577.0
221.00	15136.0
222.00	2977.0
223.00	5591.0
224.00	48472.0
225.00	13722.0
226.10	1508.0
227.00	21472.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

228.00	3456.0
229.00	5591.0
229.90	1185.0
231.00	2000.0
232.10	409.0
233.00	662.0
234.00	1589.0
235.00	1772.0
236.00	1017.0
236.90	1857.0
238.20	414.0
239.20	5022.0
240.20	1417.0
241.00	1521.0
242.00	2864.0
243.10	3196.0
244.00	41736.0
245.00	5988.0
246.00	8074.0
247.00	1809.0
248.00	309.0
249.00	1585.0
250.10	319.0
250.90	456.0
252.00	359.0
253.00	1200.0
255.00	220096.0
256.00	32984.0
257.00	2833.0
258.00	12405.0
259.00	1976.0
260.00	467.0
261.00	393.0
264.20	666.0
265.00	5415.0
266.00	1082.0
267.30	4797.0
268.20	996.0
269.00	337.0
270.10	429.0
271.10	496.0
272.00	703.0
273.00	6496.0
274.00	16520.0
275.00	101504.0
276.00	14086.0
277.00	8520.0
277.90	1323.0
279.00	305.0
281.20	950.0
282.20	1175.0
283.20	2637.0
284.10	1201.0
285.00	1408.0
285.90	365.0
289.00	322.0
289.90	394.0
292.10	406.0
293.00	1932.0
294.10	546.0
295.10	1373.0
296.00	27848.0
297.00	4096.0
301.00	430.0
301.90	529.0
303.10	3164.0
304.10	865.0
309.10	680.0
310.00	493.0
314.00	1335.0
315.00	2966.0
316.10	2037.0
316.90	503.0
321.00	893.0
322.10	618.0
323.00	7533.0
324.00	1984.0
325.10	377.0
327.00	1704.0
327.90	1146.0
332.00	641.0
333.00	1034.0
334.00	5852.0
335.10	1634.0
336.00	336.0
341.10	1128.0
342.00	427.0
346.00	1983.0
351.20	301.0
352.10	2118.0
353.00	2133.0
354.00	3142.0
355.00	911.0
365.00	11299.0
366.00	1941.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

370.00	388.0
371.00	797.0
372.00	4603.0
373.00	1143.0
383.10	1395.0
390.00	664.0
391.10	524.0
392.10	489.0
401.00	389.0
402.10	1828.0
403.00	2866.0
404.10	729.0
421.10	1993.0
422.00	1750.0
423.10	13686.0
424.00	2530.0
441.10	33496.0
442.10	223104.0
443.10	44952.0
444.10	4308.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

m/z	Abundance
36.00	2081.0
37.00	865.0
38.00	3369.0
39.10	12907.0
40.00	3091.0
41.10	14103.0
42.10	3896.0
43.10	21224.0
44.00	18488.0
45.00	1884.0
46.00	301.0
49.10	1482.0
50.00	28880.0
51.10	118544.0
52.00	6513.0
53.10	2924.0
54.10	2537.0
55.10	19744.0
56.10	9671.0
57.10	31320.0
58.10	5661.0
59.00	2908.0
60.00	2480.0
61.00	1863.0
62.00	2305.0
63.00	5712.0
64.00	870.0
65.10	3565.0
66.10	1124.0
67.10	5832.0
68.10	5585.0
69.00	139136.0
70.10	5373.0
71.10	15823.0
72.00	1198.0
73.10	3791.0
74.10	13488.0
75.00	21792.0
76.00	7899.0
77.10	146176.0
78.10	10849.0
79.00	12597.0
80.00	8755.0
81.00	16656.0
82.00	7139.0
83.10	10593.0
84.10	4491.0
85.10	10586.0
86.10	4226.0
87.10	2146.0
87.90	797.0
89.00	567.0
90.00	311.0
91.00	4934.0
92.10	4482.0
93.00	18872.0
94.00	2890.0
95.10	4870.0
96.10	5605.0
97.10	7255.0
98.00	19872.0
99.00	11640.0
100.00	1624.0
101.00	6176.0
102.00	371.0
102.90	2466.0
104.00	4056.0
105.00	5604.0
105.90	1584.0
107.00	48616.0
108.00	7677.0
109.10	4173.0
110.00	85432.0
111.00	15158.0
112.00	2889.0
113.00	2330.0
114.10	420.0
115.00	1796.0
116.10	3116.0
117.00	38080.0
118.00	3376.0
119.00	1440.0
120.10	1077.0
121.00	1664.0
122.00	3893.0
123.00	6179.0
124.00	2951.0
125.10	3138.0
127.00	164288.0
128.00	13345.0
129.00	68136.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

130.00	5958.0
131.00	1763.0
132.10	797.0
133.10	1251.0
134.10	2276.0
135.00	5602.0
136.00	2296.0
137.10	3370.0
138.00	1354.0
139.10	1154.0
140.00	1591.0
141.00	9804.0
142.00	3209.0
143.00	3243.0
143.90	1144.0
145.00	1139.0
146.00	2503.0
147.00	5725.0
148.00	9558.0
149.00	2611.0
150.00	711.0
151.00	1834.0
152.00	1566.0
153.00	3549.0
154.00	2407.0
155.00	6053.0
156.00	8010.0
157.10	2112.0
158.00	1622.0
159.10	1645.0
160.00	3602.0
161.00	4379.0
162.00	1151.0
162.90	740.0
164.00	538.0
165.00	4350.0
166.10	3863.0
167.00	19664.0
168.00	8516.0
169.00	2086.0
170.00	937.0
171.00	1177.0
172.00	2022.0
173.00	2609.0
174.00	4049.0
175.00	7599.0
176.00	2581.0
176.90	4011.0
178.00	1746.0
179.00	14626.0
180.00	9977.0
181.00	5325.0
182.10	1158.0
183.10	783.0
184.10	1547.0
185.00	7436.0
186.00	51712.0
187.00	16968.0
188.00	1618.0
189.00	3546.0
190.10	554.0
190.90	2465.0
192.00	4836.0
193.00	5658.0
194.10	2060.0
195.20	1411.0
196.00	12465.0
198.00	381184.0
199.00	26568.0
200.00	1912.0
201.40	1970.0
202.20	846.0
203.00	2966.0
204.00	14868.0
205.00	24544.0
206.00	102752.0
207.00	17640.0
208.00	4211.0
209.00	1893.0
210.10	2091.0
211.00	4415.0
212.10	776.0
213.10	780.0
215.10	1502.0
216.00	2638.0
217.00	26472.0
218.00	3623.0
219.10	408.0
221.00	16440.0
222.10	4175.0
223.00	6899.0
224.00	60440.0
225.00	14813.0
226.10	1807.0
227.00	23984.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

228.00	3867.0
229.00	5123.0
230.10	580.0
231.00	2196.0
232.10	759.0
233.10	410.0
233.90	1733.0
235.10	1959.0
236.00	1111.0
237.00	1883.0
237.90	346.0
239.20	4351.0
240.20	1413.0
241.10	1631.0
242.00	3377.0
243.00	4331.0
244.00	45440.0
245.00	5882.0
246.00	9393.0
247.00	1808.0
247.90	735.0
248.90	1757.0
249.90	425.0
251.00	576.0
252.10	606.0
253.00	968.0
255.00	259008.0
256.00	37824.0
257.00	2675.0
258.00	14637.0
258.90	2132.0
259.90	431.0
261.00	420.0
264.10	725.0
265.00	5795.0
266.00	881.0
267.20	4864.0
268.20	917.0
270.80	471.0
272.10	854.0
273.00	7207.0
274.00	21416.0
275.00	118824.0
276.00	16160.0
277.00	11285.0
278.00	1741.0
279.00	385.0
281.10	1047.0
282.10	1009.0
283.20	3253.0
284.20	1360.0
285.00	1939.0
286.00	545.0
288.00	333.0
288.90	695.0
290.10	427.0
292.00	558.0
293.00	2099.0
294.00	535.0
295.20	981.0
296.00	35928.0
297.00	4454.0
299.00	390.0
301.00	364.0
301.80	470.0
303.00	3367.0
304.00	1148.0
305.10	330.0
308.10	632.0
309.10	497.0
309.90	519.0
312.90	446.0
314.00	1226.0
315.00	4310.0
316.00	2188.0
317.10	477.0
321.00	1213.0
322.00	711.0
323.00	11047.0
324.00	2356.0
326.00	350.0
327.00	3046.0
327.90	1164.0
332.00	984.0
333.00	793.0
334.00	7844.0
335.00	1902.0
336.10	389.0
341.00	1224.0
342.10	444.0
346.00	2862.0
347.00	504.0
352.00	2935.0
353.00	1945.0
354.10	3862.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

355.10	684.0
364.10	303.0
365.00	16367.0
366.00	2188.0
366.90	303.0
370.10	316.0
371.00	749.0
372.10	5477.0
373.00	1412.0
383.00	1542.0
384.10	500.0
384.90	301.0
390.00	677.0
391.10	613.0
392.00	433.0
401.00	422.0
402.00	2530.0
403.00	3239.0
404.00	1339.0
421.00	2304.0
422.10	2844.0
423.00	18072.0
424.10	3393.0
425.10	347.0
440.20	351.0
441.10	52016.0
442.10	316864.0
443.10	60248.0
444.00	5390.0
445.10	365.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

m/z	Abundance
36.00	2055.0
36.80	812.0
38.00	2215.0
39.00	9449.0
40.00	3193.0
41.10	15834.0
42.10	4913.0
43.10	21696.0
44.00	20240.0
45.10	1464.0
46.10	337.0
49.00	1093.0
50.10	17528.0
51.10	65040.0
52.10	3958.0
53.00	2162.0
54.10	2655.0
55.10	18136.0
56.10	7592.0
57.10	26232.0
58.00	5410.0
59.10	2623.0
60.00	2388.0
60.90	1442.0
62.00	1084.0
63.00	3469.0
64.00	1054.0
65.00	2386.0
66.10	1066.0
67.10	6816.0
68.10	4723.0
69.00	85824.0
70.10	5763.0
71.10	16760.0
72.10	1523.0
73.10	3737.0
74.00	7666.0
75.00	12753.0
76.10	4190.0
77.10	88368.0
78.10	6909.0
79.00	9282.0
80.00	5825.0
81.00	11288.0
82.10	6519.0
83.10	10410.0
84.10	4341.0
85.10	10892.0
86.00	2690.0
87.00	1552.0
88.10	491.0
89.00	611.0
91.10	5116.0
92.10	2544.0
93.00	12461.0
94.00	1769.0
95.10	5340.0
96.10	5426.0
97.10	7301.0
98.00	16148.0
99.00	8005.0
100.00	1422.0
101.00	4015.0
102.10	387.0
103.00	1848.0
104.00	2749.0
105.00	3894.0
106.10	1386.0
107.00	27976.0
108.00	4670.0
109.10	3778.0
110.00	52304.0
111.00	10573.0
112.10	3046.0
113.10	1678.0
114.10	800.0
115.10	1515.0
116.00	2231.0
117.00	23640.0
118.00	1801.0
119.00	1346.0
119.90	918.0
121.00	1537.0
122.00	2061.0
123.00	4403.0
124.10	1624.0
125.10	2220.0
126.20	1407.0
127.00	101120.0
128.00	7923.0
129.00	39624.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

130.00	3634.0
131.00	1630.0
132.00	471.0
133.00	1632.0
134.00	1431.0
135.00	4280.0
136.00	1964.0
137.10	2332.0
138.20	672.0
139.00	927.0
140.20	895.0
141.10	6505.0
142.00	2272.0
143.00	1808.0
143.90	640.0
145.00	904.0
146.00	1121.0
147.00	2970.0
148.00	6436.0
149.00	1773.0
150.00	562.0
151.10	1149.0
152.00	981.0
153.00	2417.0
154.00	1382.0
155.00	3629.0
156.00	4471.0
157.00	1434.0
158.00	1849.0
159.00	1096.0
160.00	2189.0
161.00	3180.0
162.00	880.0
163.10	796.0
164.10	655.0
165.00	2844.0
166.00	2216.0
167.00	11845.0
168.00	5874.0
169.00	1792.0
170.10	775.0
171.00	1277.0
172.00	1293.0
173.00	2053.0
174.00	2439.0
175.00	5141.0
176.00	1342.0
176.90	2859.0
178.00	1058.0
179.00	9101.0
180.00	6412.0
181.00	2748.0
182.00	649.0
183.00	848.0
184.00	1144.0
185.00	4384.0
186.00	32832.0
187.10	9510.0
188.00	1191.0
189.00	2457.0
190.00	458.0
191.00	1808.0
192.00	3040.0
193.00	4154.0
194.00	1482.0
194.90	710.0
196.00	6866.0
197.10	2631.0
197.90	241600.0
199.00	16225.0
200.10	1342.0
201.50	1406.0
202.90	1454.0
204.00	9616.0
205.00	14568.0
206.00	64984.0
207.00	11602.0
208.00	3050.0
209.00	1407.0
210.00	1422.0
211.20	2542.0
211.90	640.0
213.10	656.0
215.10	1293.0
216.00	1636.0
217.00	17904.0
218.00	2877.0
219.00	542.0
221.00	10929.0
221.90	2494.0
223.00	4298.0
224.00	37464.0
225.00	10004.0
226.00	1153.0
227.00	16318.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

228.00	2630.0
229.00	3467.0
230.00	676.0
231.10	1435.0
232.10	379.0
232.90	572.0
233.90	938.0
235.00	1224.0
236.00	1088.0
236.90	1300.0
238.10	343.0
239.20	3634.0
240.20	1390.0
241.10	1246.0
242.00	1873.0
243.00	2278.0
244.00	29984.0
245.10	4096.0
246.00	6208.0
246.90	1033.0
249.00	1419.0
250.90	332.0
251.90	546.0
253.10	909.0
254.20	2172.0
255.00	174528.0
256.00	24768.0
257.00	1967.0
258.00	9162.0
259.00	1510.0
260.00	473.0
263.90	454.0
265.00	4036.0
266.00	639.0
267.30	4833.0
268.30	858.0
269.20	431.0
271.00	485.0
271.90	469.0
272.20	455.0
273.00	5900.0
274.00	13796.0
275.00	80128.0
276.00	10922.0
276.90	6317.0
278.00	1195.0
281.10	836.0
282.10	1034.0
283.20	2866.0
284.20	855.0
285.00	1133.0
292.10	393.0
293.00	1510.0
293.90	332.0
295.00	910.0
296.00	23112.0
297.00	3381.0
297.90	339.0
300.90	384.0
302.00	396.0
303.00	3078.0
304.00	889.0
308.00	410.0
309.10	327.0
310.10	429.0
313.10	378.0
314.00	1245.0
315.00	2704.0
316.00	1662.0
317.00	369.0
320.90	1018.0
322.20	530.0
323.10	7416.0
323.90	1437.0
327.00	1886.0
327.90	694.0
332.00	622.0
333.00	783.0
334.10	4890.0
335.10	1434.0
341.00	1294.0
342.00	463.0
346.00	1651.0
352.10	2809.0
353.00	1895.0
354.00	2654.0
355.00	757.0
365.00	13310.0
365.90	1483.0
369.90	439.0
371.00	725.0
372.00	4550.0
373.00	843.0
383.00	1205.0
384.00	402.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

390.00	714.0
391.10	551.0
400.90	480.0
402.00	2023.0
403.00	2676.0
404.00	948.0
421.10	1860.0
422.10	2006.0
423.10	14064.0
424.00	2914.0
441.10	38864.0
442.10	256640.0
443.10	46944.0
444.10	4253.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09