

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM101617\
 Data File : BM012217.D
 Acq On : 16 Oct 2017 16:05
 Operator : SJ/JU
 Sample : PB103282BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK82

Quant Time: Oct 16 23:42:53 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM101217.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 16 23:37:16 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	132205	20.00	ng/ul	0.00
18) Naphthalene-d8	10.60	136	562659	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.45	164	398204	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.20	188	1011008	20.00	ng/ul	0.00
75) Chrysene-d12	21.38	240	998103	20.00	ng/ul	0.00
83) Perylene-d12	23.69	264	823671	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.22	96	8531	3.07	ng/uL	0.00
5) Phenol-d5	6.97	99	161699	15.07	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.13	67	95920	14.91	ng/ul	0.00
9) 2-Chlorophenol-d4	7.33	132	148123	16.28	ng/ul	0.00
13) 4-Methylphenol-d8	8.52	113	148918	16.12	ng/ul	0.00
19) Nitrobenzene-d5	8.97	128	75437	17.57	ng/ul	0.00
22) 2-Nitrophenol-d4	9.69	143	90498	18.04	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.23	165	164909	17.22	ng/ul	0.00
29) 4-Chloroaniline-d4	10.75	131	213413	23.81	ng/ul	0.00
43) Dimethylphthalate-d6	13.86	166	664313	18.88	ng/ul	0.00
46) Acenaphthylene-d8	14.15	160	734877	17.25	ng/ul	0.00
51) 4-Nitrophenol-d4	14.69	143	69132	13.07	ng/ul	0.01
57) Fluorene-d10	15.45	176	538298	18.64	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.59	200	72600	10.46	ng/ul	0.00
70) Anthracene-d10	17.30	188	877633	17.56	ng/ul	0.00
76) Pyrene-d10	19.59	212	1026166	20.25	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.54	264	699462	17.61	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM101617\
Data File : BM012217.D
Acq On : 16 Oct 2017 16:05
Operator : SJ/JU
Sample : PB103282BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK82

Quant Time: Oct 16 23:42:53 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM101217.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 16 23:37:16 2017
Response via : Initial Calibration

