

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/9/2023	BNA_F	<u>BF135642.D</u>
Compound Name	Response	Retention Time
DDT	563530	13.445
DDD	13882	13.169
DDE	386	12.804
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
14268	577798	2.47