

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/9/2023	BNA_F	<u>BF135654.D</u>
Compound Name	Response	Retention Time
DDT	537033	13.445
DDD	15341	13.145
DDE	786	12.804
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
16127	553160	2.92