

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101623\
 Data File : BM042304.D
 Acq On : 16 Oct 2023 18:22
 Operator : MA/JU
 Sample : PB156402BL
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK402

Quant Time: Oct 17 01:29:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 12 15:38:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.016	152	163126	20.000	ng/ul	0.00
20) Naphthalene-d8	10.839	136	737880	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.657	164	449330	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.410	188	960851	20.000	ng/ul	0.00
79) Chrysene-d12	21.592	240	869812	20.000	ng/ul	0.00
88) Perylene-d12	24.068	264	941664	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	25767	4.936	ng/uL	0.00
4) Pyridine-d5	3.810	84	173534	11.223	ng/ul	0.00
7) Phenol-d5	7.163	99	413638	22.880	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.357	67	301502	24.682	ng/ul	0.00
11) 2-Chlorophenol-d4	7.534	132	302402	24.316	ng/ul	0.00
15) 4-Methylphenol-d8	8.722	113	323569	22.600	ng/ul	0.00
21) Nitrobenzene-d5	9.216	128	151902	24.240	ng/ul	0.00
24) 2-Nitrophenol-d4	9.928	143	159323	23.783	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.451	165	269375	22.035	ng/ul	0.00
31) 4-Chloroaniline-d4	10.998	131	119865	6.227	ng/ul	0.00
46) Dimethylphthalate-d6	14.074	166	973076	24.982	ng/ul	0.00
49) Acenaphthylene-d8	14.357	160	1067745	25.046	ng/ul	0.00
54) 4-Nitrophenol-d4	14.868	143	168854	21.553	ng/ul	0.00
60) Fluorene-d10	15.651	176	809886	25.294	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.774	200	126114	18.737	ng/ul	0.00
73) Anthracene-d10	17.509	188	1291799	26.245	ng/ul	0.00
81) Pyrene-d10	19.798	212	1487825	28.065	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.903	264	1393669	26.964	ng/ul	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101623\
Data File : BM042304.D
Acq On : 16 Oct 2023 18:22
Operator : MA/JU
Sample : PB156402BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK402

Quant Time: Oct 17 01:29:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 12 15:38:21 2023
Response via : Initial Calibration

