

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101623\
 Data File : BM042301.D
 Acq On : 16 Oct 2023 16:31
 Operator : MA/JU
 Sample : PB156298BL
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK298

Quant Time: Oct 17 01:26:10 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 12 15:38:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.016	152	132068	20.000	ng/u1	0.00
20) Naphthalene-d8	10.839	136	597926	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.657	164	359796	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.410	188	761198	20.000	ng/u1	0.00
79) Chrysene-d12	21.592	240	658941	20.000	ng/u1	0.00
88) Perylene-d12	24.068	264	727774	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	30851	7.299	ng/uL	0.00
4) Pyridine-d5	3.805	84	451991	36.107	ng/u1	0.00
7) Phenol-d5	7.163	99	474583	32.425	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.357	67	346342	35.021	ng/u1	0.00
11) 2-Chlorophenol-d4	7.534	132	349978	34.759	ng/u1	0.00
15) 4-Methylphenol-d8	8.722	113	373089	32.187	ng/u1	0.00
21) Nitrobenzene-d5	9.216	128	174813	34.425	ng/u1	0.00
24) 2-Nitrophenol-d4	9.928	143	183602	33.823	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.451	165	309199	31.212	ng/u1	0.00
31) 4-Chloroaniline-d4	10.998	131	523138	33.540	ng/u1	0.00
46) Dimethylphthalate-d6	14.075	166	1088443	34.898	ng/u1	0.00
49) Acenaphthylene-d8	14.357	160	1210638	35.465	ng/u1	0.00
54) 4-Nitrophenol-d4	14.869	143	194735	31.042	ng/u1	0.00
60) Fluorene-d10	15.651	176	897885	35.020	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.774	200	141860	26.604	ng/u1	0.00
73) Anthracene-d10	17.510	188	1414650	36.279	ng/u1	0.00
81) Pyrene-d10	19.804	212	1646713	41.003	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.909	264	1577557	39.491	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101623\
Data File : BM042301.D
Acq On : 16 Oct 2023 16:31
Operator : MA/JU
Sample : PB156298BL
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK298

Quant Time: Oct 17 01:26:10 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 12 15:38:21 2023
Response via : Initial Calibration

