

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101623\
 Data File : BM042296.D
 Acq On : 16 Oct 2023 13:27
 Operator : MA/JU
 Sample : PB156265BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK265

Quant Time: Oct 17 01:19:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 12 15:38:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.016	152	112400	20.000	ng/u1	0.00
20) Naphthalene-d8	10.839	136	509768	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.657	164	307849	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.409	188	649598	20.000	ng/u1	0.00
79) Chrysene-d12	21.591	240	555454	20.000	ng/u1	0.00
88) Perylene-d12	24.068	264	603242	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	27922	7.762	ng/uL	0.00
4) Pyridine-d5	3.804	84	395809	37.151	ng/u1	0.00
7) Phenol-d5	7.163	99	449706	36.101	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.357	67	322775	38.349	ng/u1	0.00
11) 2-Chlorophenol-d4	7.534	132	316431	36.927	ng/u1	0.00
15) 4-Methylphenol-d8	8.722	113	345631	35.036	ng/u1	0.00
21) Nitrobenzene-d5	9.216	128	158960	36.717	ng/u1	0.00
24) 2-Nitrophenol-d4	9.927	143	172026	37.171	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.451	165	289870	34.321	ng/u1	0.00
31) 4-Chloroaniline-d4	10.998	131	472831	35.557	ng/u1	0.00
46) Dimethylphthalate-d6	14.074	166	1018374	38.161	ng/u1	0.00
49) Acenaphthylene-d8	14.357	160	1115381	38.188	ng/u1	0.00
54) 4-Nitrophenol-d4	14.868	143	170600	31.784	ng/u1	0.00
60) Fluorene-d10	15.651	176	857293	39.079	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.774	200	129824	28.530	ng/u1	0.00
73) Anthracene-d10	17.509	188	1324138	39.792	ng/u1	0.00
81) Pyrene-d10	19.803	212	1499296	44.288	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.909	264	1371337	41.416	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101623\
Data File : BM042296.D
Acq On : 16 Oct 2023 13:27
Operator : MA/JU
Sample : PB156265BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK265

Quant Time: Oct 17 01:19:38 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 12 15:38:21 2023
Response via : Initial Calibration

