

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101623\
 Data File : BM042308.D
 Acq On : 16 Oct 2023 20:50
 Operator : MA/JU
 Sample : 04900-03
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AT5

Quant Time: Oct 17 01:30:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 12 15:38:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.016	152	138396	20.000	ng/u1	0.00
20) Naphthalene-d8	10.839	136	624921	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.657	164	374747	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.410	188	798051	20.000	ng/u1	0.00
79) Chrysene-d12	21.586	240	694679	20.000	ng/u1	-0.01
88) Perylene-d12	24.062	264	741158	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.375	96	20107	4.540	ng/uL	0.00
4) Pyridine-d5	3.810	84	114633	8.739	ng/u1	0.00
7) Phenol-d5	7.163	99	103198	6.728	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.357	67	445346	42.972	ng/u1	0.00
11) 2-Chlorophenol-d4	7.534	132	325424	30.843	ng/u1	0.00
15) 4-Methylphenol-d8	8.722	113	203731	16.773	ng/u1	0.00
21) Nitrobenzene-d5	9.210	128	226741	42.722	ng/u1	0.00
24) 2-Nitrophenol-d4	9.928	143	233765	41.203	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.445	165	360214	34.791	ng/u1	0.00
31) 4-Chloroaniline-d4	10.998	131	485937	29.809	ng/u1	0.00
46) Dimethylphthalate-d6	14.075	166	1390851	42.815	ng/u1	0.00
49) Acenaphthylene-d8	14.357	160	1526438	42.932	ng/u1	0.00
54) 4-Nitrophenol-d4	14.869	143	32759	5.014	ng/u1	0.00
60) Fluorene-d10	15.651	176	1189325	44.536	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.774	200	189491	33.896	ng/u1	0.00
73) Anthracene-d10	17.510	188	1960809	47.964	ng/u1	0.00
81) Pyrene-d10	19.798	212	2290603	54.101	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.903	264	2060362	50.646	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101623\
Data File : BM042308.D
Acq On : 16 Oct 2023 20:50
Operator : MA/JU
Sample : 04900-03
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AT5

Quant Time: Oct 17 01:30:18 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 12 15:38:21 2023
Response via : Initial Calibration

