

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM101718\
 Data File : BM017178.D
 Acq On : 18 Oct 2018 04:22
 Operator : MJ/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 18 13:02:37 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM101718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 17 14:23:17 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.604	0.545	9.8	94	0.00
3	Pyridine	1.389	1.348	3.0	95	0.00
4	n-Nitrosodimethylamine	0.558	0.560	-0.4	104	0.00
5 S	2-Fluorophenol	1.182	1.171	0.9	99	0.00
6	Aniline	2.014	1.965	2.4	96	0.00
7 S	Phenol-d6	1.610	1.574	2.2	96	0.00
8	2-Chlorophenol	1.368	1.363	0.4	98	0.00
9	Benzaldehyde	1.027	0.917	10.7	92	0.00
10 C	Phenol	1.733	1.656	4.4	94	0.00
11	bis(2-Chloroethyl)ether	1.382	1.310	5.2	95	0.00
12	1,3-Dichlorobenzene	1.595	1.574	1.3	101	0.00
13 C	1,4-Dichlorobenzene	1.629	1.585	2.7	99	0.00
14	1,2-Dichlorobenzene	1.570	1.530	2.5	98	0.00
15	Benzyl Alcohol	1.177	1.255	-6.6	102	0.00
16	2,2'-oxybis(1-Chloropropane	1.930	1.761	8.8	92	0.00
17	2-Methylphenol	1.113	1.113	0.0	98	0.00
18	Hexachloroethane	0.558	0.557	0.2	100	0.00
19 P	n-Nitroso-di-n-propylamine	1.060	1.088	-2.6	98	0.00
20	3+4-Methylphenols	1.522	1.522	0.0	97	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
22	Acetophenone	0.507	0.498	1.8	96	0.00
23 S	Nitrobenzene-d5	0.342	0.370	-8.2	100	0.00
24	Nitrobenzene	0.362	0.378	-4.4	97	0.00
25	Isophorone	0.565	0.577	-2.1	96	0.00
26 C	2-Nitrophenol	0.156	0.170	-9.0	105	0.00
27	2,4-Dimethylphenol	0.250	0.257	-2.8	98	0.00
28	bis(2-Chloroethoxy)methane	0.397	0.394	0.8	95	0.00
29 C	2,4-Dichlorophenol	0.290	0.307	-5.9	100	0.00
30	1,2,4-Trichlorobenzene	0.348	0.354	-1.7	101	0.00
31	Naphthalene	0.980	0.964	1.6	98	0.00
32	Benzoic acid	0.186	0.205	-10.2	111	0.00
33	4-Chloroaniline	0.423	0.437	-3.3	98	0.00
34 C	Hexachlorobutadiene	0.220	0.225	-2.3	102	0.00
35	Caprolactam	0.106	0.106	0.0	97	0.00
36 C	4-Chloro-3-methylphenol	0.327	0.347	-6.1	100	0.00
37	2-Methylnaphthalene	0.671	0.690	-2.8	99	0.00
38	1-Methylnaphthalene	0.653	0.667	-2.1	98	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	0.704	0.707	-0.4	101	0.00
41 P	Hexachlorocyclopentadiene	0.340	0.393	-15.6	111	0.00
42 S	2,4,6-Tribromophenol	0.270	0.288	-6.7	107	0.00
43 C	2,4,6-Trichlorophenol	0.409	0.446	-9.0	102	0.00
44	2,4,5-Trichlorophenol	0.438	0.470	-7.3	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM101718\
 Data File : BM017178.D
 Acq On : 18 Oct 2018 04:22
 Operator : MJ/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 18 13:02:37 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM101718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 17 14:23:17 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.503	1.515	-0.8	101	0.00
46	1,1'-Biphenyl	1.800	1.793	0.4	100	0.00
47	2-Chloronaphthalene	1.324	1.321	0.2	99	0.00
48	2-Nitroaniline	0.342	0.382	-11.7	106	0.00
49	Acenaphthylene	1.919	1.983	-3.3	99	0.00
50	Dimethylphthalate	1.543	1.595	-3.4	101	0.00
51	2,6-Dinitrotoluene	0.340	0.358	-5.3	103	0.00
52 C	Acenaphthene	1.205	1.205	0.0	99	0.00
53	3-Nitroaniline	0.368	0.381	-3.5	101	0.00
54 P	2,4-Dinitrophenol	0.162	0.193	-19.1	116	0.00
55	Dibenzofuran	2.004	1.984	1.0	100	0.00
56 P	4-Nitrophenol	0.299	0.318	-6.4	103	0.00
57	2,4-Dinitrotoluene	0.456	0.493	-8.1	104	0.00
58	Fluorene	1.483	1.502	-1.3	101	0.00
59	2,3,4,6-Tetrachlorophenol	0.400	0.432	-8.0	103	-0.01
60	Diethylphthalate	1.530	1.639	-7.1	103	0.00
61	4-Chlorophenyl-phenylether	0.846	0.851	-0.6	99	0.00
62	4-Nitroaniline	0.381	0.411	-7.9	104	0.00
63	Azobenzene	1.487	1.557	-4.7	99	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	0.116	0.133	-14.7	116	0.00
66 c	n-Nitrosodiphenylamine	0.605	0.623	-3.0	103	0.00
67	4-Bromophenyl-phenylether	0.220	0.229	-4.1	104	0.00
68	Hexachlorobenzene	0.255	0.254	0.4	103	0.00
69	Atrazine	0.217	0.227	-4.6	105	0.00
70 C	Pentachlorophenol	0.175	0.184	-5.1	106	0.00
71	Phenanthrene	1.079	1.059	1.9	102	0.00
72	Anthracene	1.025	1.049	-2.3	102	0.00
73	Carbazole	1.047	1.068	-2.0	103	0.00
74	Di-n-butylphthalate	1.105	1.183	-7.1	108	0.00
75 C	Fluoranthene	1.214	1.240	-2.1	102	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
77	Benzidine	0.507	0.587	-15.8	113	0.00
78	Pyrene	1.118	1.135	-1.5	103	0.00
79 S	Terphenyl-d14	0.887	0.891	-0.5	104	0.00
80	Butylbenzylphthalate	0.384	0.464	-20.8	126	0.00
81	Benzo(a)anthracene	1.125	1.132	-0.6	105	0.00
82	3,3'-Dichlorobenzidine	0.419	0.443	-5.7	109	0.00
83	Chrysene	1.114	1.102	1.1	105	0.00
84	Bis(2-ethylhexyl)phthalate	0.601	0.683	-13.6	117	0.00
85 c	Di-n-octyl phthalate	1.018	1.138	-11.8	115	0.00
86	Indeno(1,2,3-cd)pyrene	1.246	1.252	-0.5	107	0.00
87 I	Perylene-d12	1.000	1.000	0.0	109	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM101718\
Data File : BM017178.D
Acq On : 18 Oct 2018 04:22
Operator : MJ/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Oct 18 13:02:37 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM101718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 17 14:23:17 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.093	1.121	-2.6	110	0.00
89	Benzo(k)fluoranthene	1.097	1.049	4.4	102	0.00
90 C	Benzo(a)pyrene	1.038	1.043	-0.5	107	0.00
91	Dibenzo(a,h)anthracene	1.031	1.030	0.1	108	0.00
92	Benzo(g,h,i)perylene	1.022	1.014	0.8	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0