

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM101719\
 Data File : BM023239.D
 Acq On : 18 Oct 2019 02:53
 Operator : JU
 Sample : PB123996BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK96

Quant Time: Oct 18 06:59:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM101419MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 18 06:52:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	344220	20.00	ng/ul	0.00
18) Naphthalene-d8	10.44	136	1337355	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.31	164	825307	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.05	188	1696612	20.00	ng/ul	0.00
78) Chrysene-d12	21.24	240	1744637	20.00	ng/ul	-0.01
86) Perylene-d12	23.45	264	2087270	20.00	ng/ul	-0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.17	96	42845	5.48	ng/uL	0.00
5) Phenol-d5	6.83	99	767024	25.59	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.00	67	438838	25.87	ng/ul	0.00
9) 2-Chlorophenol-d4	7.19	132	637981	27.80	ng/ul	0.00
13) 4-Methylphenol-d8	8.37	113	610669	25.36	ng/ul	0.00
19) Nitrobenzene-d5	8.80	128	294541	32.55	ng/ul	0.00
22) 2-Nitrophenol-d4	9.53	143	326634	37.91	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.06	165	629657	28.74	ng/ul	0.00
29) 4-Chloroaniline-d4	10.57	131	819404	30.74	ng/ul	0.00
44) Dimethylphthalate-d6	13.72	166	1861585	29.43	ng/ul	0.00
47) Acenaphthylene-d8	14.00	160	2173630	29.30	ng/ul	0.00
52) 4-Nitrophenol-d4	14.50	143	294655	28.78	ng/ul	0.00
58) Fluorene-d10	15.30	176	1577842	29.28	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.42	200	194218	28.87	ng/ul	0.00
71) Anthracene-d10	17.15	188	2575450	31.48	ng/ul	0.00
79) Pyrene-d10	19.44	212	2867950	33.94	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.31	264	3371321	31.86	ng/ul	-0.01

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM101719\
Data File : BM023239.D
Acq On : 18 Oct 2019 02:53
Operator : JU
Sample : PB123996BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK96

Quant Time: Oct 18 06:59:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM101419MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Oct 18 06:52:01 2019
Response via : Initial Calibration

