

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101723\
 Data File : BM042331.D
 Acq On : 17 Oct 2023 21:27
 Operator : MA/JU
 Sample : PB156384TB 10X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLEB384

Quant Time: Oct 18 02:23:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 18 02:04:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	124587	20.000	ng/u1	0.00
20) Naphthalene-d8	10.834	136	571103	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	348713	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.404	188	746219	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	645601	20.000	ng/u1	0.00
88) Perylene-d12	24.050	264	682161	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	30040	7.534	ng/uL	0.00
4) Pyridine-d5	3.805	84	425799	36.057	ng/u1	0.00
7) Phenol-d5	7.157	99	505211	36.590	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.351	67	349701	37.483	ng/u1	0.00
11) 2-Chlorophenol-d4	7.534	132	358794	37.775	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	387932	35.477	ng/u1	0.00
21) Nitrobenzene-d5	9.210	128	180080	37.128	ng/u1	0.00
24) 2-Nitrophenol-d4	9.922	143	193159	37.255	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.445	165	331970	35.085	ng/u1	0.00
31) 4-Chloroaniline-d4	10.992	131	524299	35.193	ng/u1	0.00
46) Dimethylphthalate-d6	14.069	166	1173687	38.827	ng/u1	0.00
49) Acenaphthylene-d8	14.357	160	1257605	38.012	ng/u1	0.00
54) 4-Nitrophenol-d4	14.863	143	203044	33.395	ng/u1	0.00
60) Fluorene-d10	15.645	176	972044	39.117	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.769	200	155053	29.662	ng/u1	0.00
73) Anthracene-d10	17.504	188	1571737	41.117	ng/u1	0.00
81) Pyrene-d10	19.792	212	1834494	46.622	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.892	264	1623455	43.358	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101723\
 Data File : BM042331.D
 Acq On : 17 Oct 2023 21:27
 Operator : MA/JU
 Sample : PB156384TB 10X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLEB384

Quant Time: Oct 18 02:23:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 18 02:04:27 2023
 Response via : Initial Calibration

