

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101723\
 Data File : BM042336.D
 Acq On : 18 Oct 2023 00:30
 Operator : MA/JU
 Sample : 04885-07 10X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 COM29

Quant Time: Oct 18 02:29:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 18 02:04:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	145129	20.000	ng/u1	0.00
20) Naphthalene-d8	10.833	136	653543	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	399465	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.404	188	845405	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	728387	20.000	ng/u1	0.00
88) Perylene-d12	24.050	264	740128	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	13240	2.851	ng/uL	0.00
4) Pyridine-d5	3.810	84	113989	8.286	ng/u1	0.00
7) Phenol-d5	7.157	99	134052	8.335	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.351	67	429937	39.561	ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	345140	31.194	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	247339	19.418	ng/u1	0.00
21) Nitrobenzene-d5	9.210	128	221522	39.911	ng/u1	0.00
24) 2-Nitrophenol-d4	9.922	143	231795	39.067	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.445	165	393270	36.320	ng/u1	0.00
31) 4-Chloroaniline-d4	10.992	131	444491	26.072	ng/u1	0.00
46) Dimethylphthalate-d6	14.069	166	1548420	44.716	ng/u1	0.00
49) Acenaphthylene-d8	14.357	160	1637553	43.208	ng/u1	0.00
54) 4-Nitrophenol-d4	14.863	143	49539	7.113	ng/u1	0.00
60) Fluorene-d10	15.645	176	1320186	46.378	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.768	200	214336	36.192	ng/u1	0.00
73) Anthracene-d10	17.504	188	2120074	48.955	ng/u1	0.00
81) Pyrene-d10	19.792	212	2506906	56.470	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.891	264	2125270	52.314	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101723\
Data File : BM042336.D
Acq On : 18 Oct 2023 00:30
Operator : MA/JU
Sample : 04885-07 10X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
COM29

Quant Time: Oct 18 02:29:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 18 02:04:27 2023
Response via : Initial Calibration

