

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101723\
 Data File : BM042337.D
 Acq On : 18 Oct 2023 01:06
 Operator : MA/JU
 Sample : 04885-08 10X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 COM30

Quant Time: Oct 18 02:31:11 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 18 02:04:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	141084	20.000	ng/u1	0.00
20) Naphthalene-d8	10.834	136	642368	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	386833	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.404	188	815306	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	678815	20.000	ng/u1	0.00
88) Perylene-d12	24.050	264	686295	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	11940	2.644	ng/uL	0.00
4) Pyridine-d5	3.805	84	140140	10.479	ng/u1	0.00
7) Phenol-d5	7.157	99	109849	7.026	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.351	67	432349	40.923	ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	329479	30.632	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	214047	17.286	ng/u1	0.00
21) Nitrobenzene-d5	9.210	128	225677	41.367	ng/u1	0.00
24) 2-Nitrophenol-d4	9.922	143	230832	39.581	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.439	165	370250	34.789	ng/u1	0.00
31) 4-Chloroaniline-d4	10.992	131	494090	29.486	ng/u1	0.00
46) Dimethylphthalate-d6	14.069	166	1413979	42.167	ng/u1	0.00
49) Acenaphthylene-d8	14.351	160	1561492	42.546	ng/u1	0.00
54) 4-Nitrophenol-d4	14.863	143	36634	5.432	ng/u1	0.00
60) Fluorene-d10	15.645	176	1219049	44.223	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.769	200	202530	35.461	ng/u1	0.00
73) Anthracene-d10	17.504	188	2035437	48.735	ng/u1	0.00
81) Pyrene-d10	19.792	212	2391484	57.804	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.886	264	2005862	53.248	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101723\
Data File : BM042337.D
Acq On : 18 Oct 2023 01:06
Operator : MA/JU
Sample : 04885-08 10X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
COM30

Quant Time: Oct 18 02:31:11 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 18 02:04:27 2023
Response via : Initial Calibration

